

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450, BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-300: BS-300, BS-320;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)

17.BS-2800M:BS-2600M, BS-2800M.

(Serial Number starts with "W9-" and software version starts with "35.00");

18.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

9.BS-380: BS-380, BS-390;

19.Applicable models of the chemistry shall be subject to the

10.BS-400: BS-400, BS-420;

parameter list and instructions.



LOT : 145323009



EXP : 2024-12-14

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

Fe	
English	Iron
Русский	Железо
Português	Ferro
Español	Hierro
Italiano	Ferro
Türkçe	Demir

Fe Calibrator



Abbreviated name	Unit	Calibration Rule			Two-point linear		
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
Fe	μmol/L	BS-120 ¹	41.0	0.5	BS-400 ¹⁰	42.4	0.5
		BS-200 ²	41.3	0.5	BS-430 ¹¹	43.1	0.5
		BS-200E ³	43.1	0.5	BS-480 ¹²	43.1	0.5
		BS-240 ⁴	41.0	0.5	BS-600 ¹³	43.1	0.5
		BS-240E ⁵	43.1	0.5	BS-600M ¹⁴	43.1	0.5
		BS-300 ⁶	41.8	0.5	BS-800 ¹⁵	43.1	0.5
		BS-330E ⁷	43.1	0.5	BS-2000 ¹⁶	43.2	0.5
		BS-360E ⁸	43.1	0.5	BS-2800M ¹⁷	43.7	0.5
		BS-380 ⁹	43.1	0.5			
	mg/L	BS-120 ¹	2.29	0.025	BS-400 ¹⁰	2.37	0.026
		BS-200 ²	2.31	0.026	BS-430 ¹¹	2.41	0.027
		BS-200E ³	2.41	0.027	BS-480 ¹²	2.41	0.027
		BS-240 ⁴	2.29	0.025	BS-600 ¹³	2.41	0.027
		BS-240E ⁵	2.41	0.027	BS-600M ¹⁴	2.41	0.027
		BS-300 ⁶	2.34	0.026	BS-800 ¹⁵	2.41	0.027
		BS-330E ⁷	2.41	0.027	BS-2000 ¹⁶	2.41	0.029
		BS-360E ⁸	2.41	0.027	BS-2800M ¹⁷	2.44	0.027
		BS-380 ⁹	2.41	0.027			

mindray

Fe Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 145323009

2024-12-14

