

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450, BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-300: BS-300, BS-320;

16.BS-1000M: BS-1000M, BS-1100M;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)

18.BS-2800M:BS-2600M, BS-2800M.

(Serial Number starts with "W9-" and software version starts with "35.00");

19.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

9.BS-380: BS-380, BS-390;

20.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

10.BS-400: BS-400, BS-420;



: 145324004



: 2025-07-07

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

Fe	
English	Iron
Русский	Железо
Português	Ferro
Español	Hierro
Italiano	Ferro
Türkçe	Demir


Abbreviated name	Unit	Calibration Rule				Two-point linear	
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
Fe	μmol/L	BS-120 ¹	44.5	0.5	BS-400 ¹⁰	46.8	0.5
		BS-200 ²	44.8	0.5	BS-430 ¹¹	46.8	0.5
		BS-200E ³	46.8	0.5	BS-480 ¹²	46.8	0.5
		BS-240 ⁴	44.5	0.5	BS-600 ¹³	46.8	0.5
		BS-240E ⁵	46.8	0.5	BS-600M ¹⁴	46.8	0.5
		BS-300 ⁶	45.8	0.5	BS-800 ¹⁵	46.8	0.5
		BS-330E ⁷	46.8	0.5	BS-1000M ¹⁶	47.6	0.5
		BS-360E ⁸	46.8	0.5	BS-2000 ¹⁷	47.5	0.6
		BS-380 ⁹	47.0	0.5	BS-2800M ¹⁸	47.6	0.5
	mg/L	BS-120 ¹	2.49	0.028	BS-400 ¹⁰	2.61	0.029
		BS-200 ²	2.50	0.028	BS-430 ¹¹	2.61	0.029
		BS-200E ³	2.61	0.029	BS-480 ¹²	2.61	0.029
		BS-240 ⁴	2.49	0.028	BS-600 ¹³	2.61	0.029
		BS-240E ⁵	2.61	0.029	BS-600M ¹⁴	2.61	0.029
		BS-300 ⁶	2.56	0.028	BS-800 ¹⁵	2.61	0.029
		BS-330E ⁷	2.61	0.029	BS-1000M ¹⁶	2.66	0.03
		BS-360E ⁸	2.61	0.029	BS-2000 ¹⁷	2.65	0.03
		BS-380 ⁹	2.63	0.029	BS-2800M ¹⁸	2.66	0.03

mindray

Fe Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 145324004

 2025-07-07

