

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-240: BS-230, BS-240;

7.BS-600: BS-600, BS-620;

2.BS-240E: BS240E, BS240Pro;

8.BS-600M: BS-600M, BS-620M;

3.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

9.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

4.BS-380: BS-380, BS-390;

10.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

5.BS-400: BS-400, BS-420;

11.BS-2800M:BS-2600M, BS-2800M;

6.BS-480: BS-480, BS-490;

12.Applicable models of the chemistry shall be subject to the parameter list and instructions.

LOT : 145924002

: 2025-04-30

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

IgE	
English	Immunoglobulin E
Русский	Иммуноглобулин E
Português	Imunoglobulina E
Español	Inmunoglobulina E
Italiano	Immunoglobulina E
Türkçe	İmmünglobulin E

Abbreviated name		IgE		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
a		0.0	/	0.0	/	0.0	/	0.0	/
b		91.8	2.3	91.8	2.3	91.8	2.3	91.8	2.3
c	IU/mL	222	4	222	4	222	4	226	4
d		681	16	681	16	681	16	683	16
e		1309	32	1309	32	1309	32	1309	32
f		1952	22	1952	22	1952	22	1952	22

Model		BS-400 ⁵		BS-480 ⁶		BS-600 ⁷		BS-600M ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
a		0.0	/	0.0	/	0.0	/	0.0	/
b		91.8	2.3	91.8	2.3	91.8	2.3	92.2	2.3
c	IU/mL	226	4	222	4	222	4	221	4
d		683	16	681	16	681	16	680	16
e		1309	32	1309	32	1309	32	1299	31
f		1952	22	1952	22	1952	22	1944	22

Model		BS-800 ⁹		BS-2000 ¹⁰		BS-2800M ¹¹	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
a		0.0	/	0.0	/	0.0	/
b		91.8	2.3	90.9	2.4	92.2	2.3
c	IU/mL	222	4	221	4	221	4
d		681	16	682	16	680	16
e		1309	32	1300	31	1299	31
f		1952	22	1987	23	1944	22

mindray

IgE Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 145924002

2025-04-30

