

**Value sheet of Mindray BS Measurement System**

**Русский:** Таблица результатов для системы BS компании Mindray

**Português :** Planilha de valores do Sistema de Medição BS da Mindray

**Español :** Hoja de valores del sistema de medición Mindray BS

**Italiano :** Scheda dei valori del sistema di misurazione BS di Mindray

**Türkçe :** Mindray BS Ölçüm Sistemi'nin değer sayfası



**The data of each group is same.**

**Русский:** Данные совпадают во всех группах.

**Português :** A dados de cada grupo é a mesma.

**Español :** la datos de cada grupo es la misma.

**Italiano :** la dati di ogni gruppo è la stessa.

**Türkçe :** her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

10.BS-480: BS-480, BS-490;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

11.BS-600: BS-600, BS-620;

3.BS-200E: BS-200E, BS-220E,;

12.BS-600M: BS-600M, BS-620M;

4.BS-240: BS-230, BS-240;

13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

5.BS-240E: BS240E, BS240Pro;

14.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)

15.BS-2800M:BS-2600M, BS-2800M.

(Serial Number starts with "W9-" and software version starts with "35.00");

16.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

7.BS-380: BS-380, BS-390;

17.Applicable models of the chemistry shall be subject to the

8.BS-400: BS-400, BS-420;

parameter list and instructions.

9.BS-430: BS-430, BS-450, BS-460;



**LOT : 146523011**



**: 2025-05-05**

English	Abbreviated name	Unit	Calibration Rule	Model
<b>Русский</b>	сокращенное наименование	Прибор	Принцип калибровки	модель
<b>Português</b>	Nome abreviado	Unidade	Regra de calibração	Modelo
<b>Español</b>	nombre abreviado	Unidad	Regla de calibración	modelo
<b>Italiano</b>	abbreviazione	Unità	Regola di calibrazione	modelli
<b>Türkçe</b>	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model

English	Expanded Uncertainty
<b>Русский</b>	Расширенная неопределенность
<b>Português</b>	Incerteza Expandida
<b>Español</b>	Incertidumbre expandida
<b>Italiano</b>	Incertezza estesa
<b>Türkçe</b>	Daha Uzun Süreli Belirsizlik

<b>β-HB</b>	
English	
<b>Русский</b>	β-гидроксибутирата
<b>Português</b>	β-Hidroxibutirato
<b>Español</b>	β-hidroxibutirato
<b>Italiano</b>	β-idrossibutirato
<b>Türkçe</b>	β-Hidroksibötrat


Abbreviated name	Unit	Calibration Rule				Two-point linear	
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
<b>β-HB</b>	mmol/ L	<b>BS-120<sup>1</sup></b>	0.920	0.004	<b>BS-430<sup>9</sup></b>	0.970	0.004
		<b>BS-200<sup>2</sup></b>	0.960	0.004	<b>BS-480<sup>10</sup></b>	0.970	0.004
		<b>BS-200E<sup>3</sup></b>	0.960	0.004	<b>BS-600<sup>11</sup></b>	0.970	0.004
		<b>BS-240<sup>4</sup></b>	0.930	0.004	<b>BS-600M<sup>12</sup></b>	0.980	0.004
		<b>BS-240E<sup>5</sup></b>	0.920	0.004	<b>BS-800<sup>13</sup></b>	0.970	0.004
		<b>BS-360E<sup>6</sup></b>	0.970	0.004	<b>BS-2000<sup>14</sup></b>	0.980	0.004
		<b>BS-380<sup>7</sup></b>	0.970	0.004	<b>BS-2800M<sup>15</sup></b>	0.980	0.003
		<b>BS-400<sup>8</sup></b>	0.970	0.004			

# mindray

## $\beta$ -HB Calibrator

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

**LOT** 146523011

 2025-05-05

