

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;	9.BS-600: BS-600, BS-620;
2.BS-240: BS-230, BS-240;	10.BS-600M: BS-600M, BS-620M;
3.BS-240E: BS240E, BS240Pro;	11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)(Serial Number starts with "W9-" and software version starts with "35.00");	12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
5.BS-380: BS-380, BS-390;	13.BS-2800M:BS-2600M, BS-2800M.
6.BS-400: BS-400, BS-420;	14.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;
7.BS-430: BS-430, BS-450, BS-460;	15.Applicable models of the chemistry shall be subject to the parameter list and instructions.
8.BS-480: BS-480, BS-490;	



: 146924001



: 2025-05-20

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model
English	Expanded Uncertainty			
Русский	Расширенная неопределенность			
Português	Incerteza Expandida			
Español	Incertidumbre expandida			
Italiano	Incertezza estesa			
Türkçe	Daha Uzun Süreli Belirsizlik			
ASO				
English	Antistreptolysin "O"			
Русский	антистрептолизина O			
Português	Antiestreptolisina "O"			
Español	antiestreptolisina "O"			
Italiano	Anti-Streptolisina "O"			
Türkçe	Antistreptolisin "O"			

# Antistreptolysin "O" Calibrator



Abbreviated name	Unit	Calibration Rule				Two-point linear	
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
<b>ASO II</b>	IU/mL	<b>BS-200E<sup>1</sup></b>	428	7	<b>BS-480<sup>8</sup></b>	428	7
		<b>BS-240<sup>2</sup></b>	419	6	<b>BS-600<sup>9</sup></b>	428	7
		<b>BS-240E<sup>3</sup></b>	428	7	<b>BS-600M<sup>10</sup></b>	427	7
		<b>BS-360E<sup>4</sup></b>	428	7	<b>BS-800<sup>11</sup></b>	428	7
		<b>BS-380<sup>5</sup></b>	428	7	<b>BS-2000<sup>12</sup></b>	428	7
		<b>BS-400<sup>6</sup></b>	428	7	<b>BS-2800M<sup>13</sup></b>	427	7
		<b>BS-430<sup>7</sup></b>	428	7			

# mindray

## Antistreptolysin "O" Calibrator

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

LOT 146924001

2025-05-20

