

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

- | | |
|---|---|
| <p>1.BS-200E: BS-200E, BS-220E,; 2.BS-240: BS-230, BS-240; 3.BS-240E: BS240E, BS240Pro; 4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00"); 5.BS-380: BS-380, BS-390; 6.BS-400: BS-400, BS-420;</p> | <p>7.BS-430: BS-430, BS-450,BS-460; 8.BS-480: BS-480, BS-490; 9.BS-600: BS-600, BS-620; 10.BS-600M: BS-600M, BS-620M; 11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus; 12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M; 13.Applicable models of the chemistry shall be subject to the parameter list and instructions.</p> |
|---|---|

LOT : 147023007 : 2025-02-27

| English | Abbreviated name | Calibration Value | Level | Model |
|-----------|--------------------------|------------------------|---------|---------|
| Русский | сокращенное наименование | Эталонное значение | Уровень | модель |
| Português | Nome abreviado | Valor de calibração | Nível | Modelo |
| Español | nombre abreviado | Valor de calibración | Nivel | modelo |
| Italiano | abbreviazione | Valore di calibrazione | Livello | modelli |
| Türkçe | kısaltılmış ad | Kalibrasyon Değeri | Düzey | model |

| English | Expanded Uncertainty |
|-----------|------------------------------|
| Русский | Расширенная неопределенность |
| Português | Incerteza Expandida |
| Español | Incertidumbre expandida |
| Italiano | Incertezza estesa |
| Türkçe | Daha Uzun Süreli Belirsizlik |

RF

| English | Rheumatoid Factor |
|-----------|----------------------|
| Русский | ревматоидные факторы |
| Português | Fator Reumatoide |
| Español | factor reumatoide |
| Italiano | fattore reumatoide |
| Türkçe | Romatoid Faktör |

| Abbreviated name | | RF II | | Calibration Rule | | spline | | | |
|------------------|-------|----------------------|----------------------|---------------------|----------------------|----------------------|----------------------|----------------------|----------------------|
| Model | | BS-200E ¹ | | BS-240 ² | | BS-240E ³ | | BS-360E ⁴ | |
| Level | Unit | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty |
| a | | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 |
| b | | 25.0 | 1.0 | 25.0 | 1.0 | 25.0 | 1.0 | 25.0 | 1.0 |
| c | IU/mL | 94.4 | 1.8 | 94.4 | 1.8 | 94.4 | 1.8 | 94.4 | 1.8 |
| d | | 235 | 8 | 235 | 8 | 235 | 8 | 235 | 8 |
| e | | 518 | 19 | 518 | 19 | 518 | 19 | 518 | 19 |
| Model | | BS-380 ⁵ | | BS-400 ⁶ | | BS-430 ⁷ | | BS-480 ⁸ | |
| Level | Unit | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty |

Rheumatoid Factor Calibrator



| | | | | | | | | | |
|----------|-------|------|-----|------|-----|------|-----|------|-----|
| a | | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 |
| b | | 25.0 | 1.0 | 25.0 | 1.0 | 25.0 | 1.0 | 25.0 | 1.0 |
| c | IU/mL | 94.4 | 1.8 | 94.4 | 1.8 | 94.4 | 1.8 | 94.4 | 1.8 |
| d | | 235 | 8 | 235 | 8 | 235 | 8 | 235 | 8 |
| e | | 518 | 19 | 518 | 19 | 518 | 19 | 518 | 19 |

| Model | | BS-600 ⁹ | | BS-600M ¹⁰ | | BS-800 ¹¹ | | BS-2000 ¹² | |
|----------|-------|---------------------|----------------------|-----------------------|----------------------|----------------------|----------------------|-----------------------|----------------------|
| Level | Unit | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty |
| a | | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 | 0.0 | 0 |
| b | | 25.0 | 1.0 | 25.0 | 1.0 | 25.0 | 1.0 | 25.5 | 1.0 |
| c | IU/mL | 94.4 | 1.8 | 94.4 | 1.8 | 94.4 | 1.8 | 94.7 | 1.8 |
| d | | 235 | 8 | 235 | 8 | 235 | 8 | 240 | 8 |
| e | | 518 | 19 | 518 | 19 | 518 | 19 | 514 | 18 |

mindray

Rheumatoid Factor Calibrator

For use on: BS-2000, BS-600M, BS-800

LOT 147023007

 2025-02-27

