

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-230**: BS-230, BS-240;

3. **BS-240E**: BS240E, BS240Pro;

4. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

BS-330E(V35.00)(Serial Number starts with "W8-"

and software version starts with "35.00"),

BS-350E(V35.00) (Serial Number starts with "W9-"

and software version starts with "35.00");

14. **S0: 0.9% NaCl, Conc. Of S0=0;**

Português: S0: 0,9% NaCl, Conc. de S0=0;

Italiano: S0: 0,9% NaCl, conc. di S0=0;

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

5. **BS-380**: BS-380, BS-390;

6. **BS-400**: BS-400, BS-420;

7. **BS-430**: BS-430, BS-450, BS-460;

8. **BS-480**: BS-480, BS-490;

9. **BS-600**: BS-600, BS-620;

10. **BS-600M**: BS-600M, BS-620M;

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

Русский: S0: 0, 9% NaCl, конц. S0=0;

Español: S0: 0,9% NaCl, Conc. de S0=0;

Türkçe: S0: %0,9 NaCl, S0 Kons.=0.

LOT : 147223005

EXP : 2024-09-29

Abbreviated name		CysC II	Calibration Rule		Spline
Model	Level	Calibration Value ¹⁴ mg/L	Model	Level	Calibration Value ¹⁴ mg/L
BS-200E ¹	a	0.590	BS-480 ⁸	a	0.590
	b	1.16		b	1.16
	c	2.27		c	2.27
	d	4.64		d	4.64
	e	8.62		e	8.62
BS-230 ²	a	0.590	BS-600 ⁹	a	0.590
	b	1.16		b	1.16
	c	2.27		c	2.27
	d	4.64		d	4.64
	e	8.62		e	8.62
BS-240E ³	a	0.590	BS-600M ¹⁰	a	0.600
	b	1.16		b	1.17
	c	2.27		c	2.29
	d	4.64		d	4.70
	e	8.62		e	8.82
BS-360E ⁴	a	0.590	BS-800 ¹¹	a	0.590
	b	1.16		b	1.16
	c	2.27		c	2.27
	d	4.64		d	4.64
	e	8.62		e	8.62

Cystatin C Calibrator

Abbreviated name		CysC II	Calibration Rule		Spline		
Model	Level	Calibration Value ¹⁴		Model	Level	Calibration Value ¹⁴	
		mg/L				mg/L	
BS-380 ⁵	a	0.590		BS-2000 ¹²	a	0.600	
	b	1.16			b	1.20	
	c	2.27			c	2.34	
	d	4.64			d	4.75	
	e	8.62			e	8.78	
BS-400 ⁶	a	0.590		BS-2800M ¹³	a	0.600	
	b	1.16			b	1.17	
	c	2.27			c	2.29	
	d	4.64			d	4.70	
	e	8.62			e	8.82	
BS-430 ⁷	a	0.590					
	b	1.16					
	c	2.27					
	d	4.64					
	e	8.62					

English	Abbreviated name	Calibration Rule	Model	Level	Calibration Value
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень	Эталонное значение
Português	Nome abreviado	Regra de calibração	Modelo	Nível	Valor de calibração
Español	nombre abreviado	Regla de calibración	modelo	Nivel	Valor de calibración
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello	Valore di calibrazione
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey	Kalibrasyon Değeri

CysC	
English	Cystatin C
Русский	цистатин C
Português	Cistatina C
Español	cistatina C
Italiano	cistatina C
Türkçe	Cystatin C