

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

1. **BS-120**: BS-120, BS-130;

2. **BS-200**: BS-200, BS-220;

3. **BS-230**: BS-230, BS-240;

4. **BS-240E**: BS240E, BS240Pro;

5. **BS-330**: BS-330, BS-350;

6. **BS-330E**: BS-330E (Serial Number starts with "XQ-"),

BS-350E (Serial Number starts with "XS-");

16. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Português: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Italiano: Cal0: 0,9% NaCl, conc. di Cal0=0;

7. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00) (Serial Number starts with "W8-"
and software version starts with "35.00"),

BS-350E(V35.00) (Serial Number starts with "W9-"
and software version starts with "35.00");

8. **BS-380**: BS-380, BS-390;

9. **BS-400**: BS-400, BS-420;

10. **BS-430**: BS-430, BS-450, BS-460;

11. **BS-480**: BS-480, BS-490;

12. **BS-600**: BS-600, BS-620;

13. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,
BS-1800, BS-1800plus;

14. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15. **BS-2800M**: BS-2600M, BS-2800M.

Русский: Cal0: 0,9% NaCl, конц. Cal0=0;

Español: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Türkçe: Cal0: %0,9 NaCl, Cal0 Kons.=0

LOT: 148022003

EXP: 2023-02-11

Abbreviated name		HCY	Calibration Rule		Spline
Model	Level	Calibrator Value ¹⁶ ($\mu\text{mol/L}$)	Model	Level	Calibrator Value ¹⁶ ($\mu\text{mol/L}$)
BS-120 ¹	a	2.68	BS-400 ⁹	a	2.71
	b	6.80		b	6.87
	c	14.5		c	14.9
	d	28.8		d	29.2
	e	45.8		e	45.8
BS-200 ²	a	2.60	BS-430 ¹⁰	a	2.71
	b	6.60		b	6.87
	c	14.6		c	14.9
	d	29.0		d	29.2
	e	45.8		e	45.8
BS-230 ³	a	2.71	BS-480 ¹¹	a	2.65
	b	6.87		b	6.80
	c	14.9		c	14.5
	d	29.2		d	28.5
	e	45.8		e	45.2
BS-240E ⁴	a	2.71	BS-600 ¹²	a	2.71
	b	6.87		b	6.87
	c	14.9		c	14.9
	d	29.2		d	29.2
	e	45.8		e	45.8
BS-330 ⁵	a	2.60	BS-800 ¹³	a	2.71
	b	6.60		b	6.87
	c	14.6		c	14.9
	d	29.0		d	29.2
	e	45.8		e	45.8

Abbreviated name		HCY	Calibration Rule		Spline
Model	Level	Calibrator Value ¹⁶ ($\mu\text{mol/L}$)	Model	Level	Calibrator Value ¹⁶ ($\mu\text{mol/L}$)
BS-330E ⁶	a	2.63	BS-2000 ¹⁴	a	2.78
	b	6.83		b	7.00
	c	16.0		c	15.2
	d	28.3		d	29.6
	e	45.7		e	46.4
BS-360E ⁷	a	2.71	BS-2800M ¹⁵	a	2.72
	b	6.87		b	6.98
	c	14.9		c	15.2
	d	29.2		d	29.6
	e	45.8		e	46.1
BS-380 ⁸	a	2.71			
	b	6.87			
	c	14.9			
	d	29.2			
	e	45.8			

English	Abbreviated name	Calibration Rule	Model	Level	Calibration Value
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень	Эталонное значение
Português	Nome abreviado	Regra de calibração	Modelo	Nível	Valor de calibração
Español	nombre breviado	Regla de calibración	modelo	Nivel	Valor de calibración
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello	Valore di calibrazione
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey	Kalibrasyon Değeri

HCY	
English	Homocysteine (HCY)
Русский	Гомоцистеин (HCY)
Português	Homocisteína (HCY)
Español	Homocisteína (HCY)
Italiano	Omocisteina (HCY)
Türkçe	Homosistain (HCY)