

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
 Português : Planilha de valores do Sistema de Medição BS da Mindray
 Español : Hoja de valores del sistema de medición Mindray BS
 Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
 Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



This Lot is the new version ApoB reagent, the updated calibrator value sheet is inside the box. Please select the corresponding value and update.

The data of each group is same.

Русский : Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

- | | |
|---|--|
| 1. BS-120 : BS-120, BS-130; | 11. BS-380 : BS-380, BS-390; |
| 2. BS-180 : BS-180, BS-190; | 12. BS-400 : BS-400, BS-420; |
| 3. BS-200 : BS-200, BS-220; | 13. BS-430 : BS-430, BS-450, BS-460; |
| 4. BS-200E : BS-200E, BS-220E; | 14. BS-480 : BS-480, BS-490; |
| 5. BS-230 : BS-230, BS-240; | 15. BS-600 : BS-600, BS-620; |
| 6. BS-240E : BS-240E, BS-240Pro; | 16. BS-600M : BS-600M, BS-620M; |
| 7. BS-300 : BS-300, BS-320; | 17. BS-800 : BS-800, BS-820, BS-800M,
BS-820M, BS-1800, BS-1800plus; |
| 8. BS-330 : BS-330, BS-350; | 18. BS-2000 : BS-2000, BS-2200,
BS-2000M, BS-2200M; |
| 9. BS-330E : BS-330E (Serial Number starts with "XQ-"),
BS-350E (Serial Number starts with "XS-"); | 19. BS-2800M : BS-2600M. |
| 10. BS-360E : BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)
(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00"); | |

20. **S1:0.9% NaCl, Conc. Of S1=0;**

Русский : S1: 0,9% NaCl, конц. S1=0;

Português : S1:0,9% NaCl, Conc. de S1=0;

Español : S1:0,9% NaCl, Conc. de S1=0;

Italiano : S1:0,9% NaCl, conc. di S1=0;

Türkçe : S1:%0,9 NaCl, S1 Kons.=0.

LOT: 150422004 : 2023-08-28

Abbreviated name	Unit	Calibration Value ²⁰					
		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
		1.70	1.70	1.65	1.72	1.70	1.70
	mmol/L	BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		1.72	1.65	1.72	1.70	1.72	1.72
	HDL-C	BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		1.70	1.70	1.70	/	1.70	1.67
	BS-2800M ¹⁹						
	1.70						
		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
		65.7	65.7	63.8	66.5	65.7	65.7
	mg/dL	BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		66.5	63.8	66.5	65.7	66.5	66.5
	LDL-C	BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		65.7	65.7	65.7	/	65.7	64.6
	BS-2800M ¹⁹						
	65.7						
		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
		3.52	3.52	3.52	3.74	3.52	3.74
	mmol/L	BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		3.74	3.52	3.74	3.74	3.74	3.74
	LDL-C	BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		3.74	3.74	3.74	3.74	3.74	3.69
	BS-2800M ¹⁹						
	3.74						
		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
		136	136	136	145	136	145
	mg/dL	BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		145	136	145	145	145	145
	LDL-C	BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		145	145	145	145	145	143
	BS-2800M ¹⁹						
	145						

Abbreviated name		ApoA1	Calibration Rule		Logit-Log(5P)	
Model	Level	Calibrator Value ²⁰		Sample Vol for Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		g/L	μmol/L			
BS-120 ¹ R1: R2: S= 300: 100: 3	S2	0.170	6.07	20	180	3
	S3	0.440	15.7	5	194	32
	S4	0.920	32.8	9	187	35
	S5	1.55	55.3	45	180	12
	S6	2.90	104	/	/	3.5
BS-180 ² R1: R2: S= 300: 100: 3	S2	0.170	6.07	20	180	3
	S3	0.440	15.7	5	194	32
	S4	0.920	32.8	9	187	35
	S5	1.55	55.3	45	180	12
	S6	2.90	104	/	/	3.5
BS-200 ³ R1: R2: S= 300: 100: 3	S2	0.160	5.71	20	180	3
	S3	0.490	17.5	5	194	32
	S4	1.00	35.7	9	187	35
	S5	1.69	60.3	45	180	12
	S6	2.89	103	/	/	3.5
BS-200E ⁴ R1: R2: S= 200: 67: 2	S2	0.130	4.64	15	135	2
	S3	0.550	19.6	45	135	2
	S4	1.16	41.4	45	135	4
	S5	1.80	64.3	35	140	8
	S6	2.90	104	/	/	2.3
BS-300 ⁷ R1: R2: S= 300: 100: 3	S2	0.150	5.36	20	180	3
	S3	0.480	17.1	5	194	32
	S4	1.05	37.5	9	187	35
	S5	1.82	65.0	45	180	12
	S6	2.98	106	/	/	3.5
BS-330 ⁸ R1: R2: S= 300: 100: 3	S2	0.160	5.71	20	180	3
	S3	0.490	17.5	5	194	32
	S4	1.00	35.7	9	187	35
	S5	1.69	60.3	45	180	12
	S6	2.89	103	/	/	3.5
BS-330E ⁹ R1: R2: S= 200: 67: 2	S2	0.130	4.64	15	135	2
	S3	0.550	19.6	45	135	2
	S4	1.16	41.4	45	135	4
	S5	1.80	64.3	35	140	8
	S6	2.90	104	/	/	2.3
BS-380 ¹¹ R1: R2: S= 200: 67: 2	S2	0.130	4.64	14	126	2
	S3	0.540	19.3	40	120	2
	S4	1.10	39.3	40	120	4
	S5	1.71	61.0	30	120	8
	S6	2.90	104	/	/	2.3
BS-400 ¹² R1: R2: S= 200: 67: 2	S2	0.130	4.64	14	126	2
	S3	0.530	18.9	40	120	2
	S4	1.12	40.0	40	120	4
	S5	1.77	63.2	30	120	8
	S6	2.92	104	/	/	2.3

★ Note: The following value is applicable to 141922001 and before ApoB reagents. ★

Abbreviated name	ApoB	Calibration Rule		Logit-Log(5P)		
Model	Level	Calibrator Value ²⁰		Sample Vol for Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		g/L	μmol/L			
BS-120 ¹ R1: R2: S= 300: 100: 3	S2	0.230	0.449	45	180	3
	S3	0.430	0.839	45	180	6
	S4	0.900	1.76	45	180	12
	S5	1.27	2.48	/	/	3
	S6	2.30	4.49	/	/	6
	BS-180 ² R1: R2: S= 300: 100: 3	S2	0.230	0.449	45	180
S3	0.430	0.839	45	180	6	
S4	0.900	1.76	45	180	12	
S5	1.27	2.48	/	/	3	
S6	2.30	4.49	/	/	6	
BS-200 ³ R1: R2: S= 300: 100: 3	S2	0.250	0.488	45	180	3
S3	0.460	0.897	45	180	6	
S4	0.940	1.83	45	180	12	
S5	1.32	2.57	/	/	3	
S6	2.40	4.68	/	/	6	
BS-200E ⁴ R1: R2: S= 300: 100: 3	S2	0.220	0.429	35	140	3
S3	0.500	0.975	35	140	6	
S4	0.970	1.89	35	140	12	
S5	1.37	2.67	/	/	3	
S6	2.52	4.91	/	/	6	
BS-300 ⁷ R1: R2: S= 300: 100: 3	S2	0.240	0.468	45	180	3
S3	0.500	0.975	45	180	6	
S4	0.990	1.93	45	180	12	
S5	1.38	2.69	/	/	3	
S6	2.48	4.84	/	/	6	
BS-330 ⁸ R1: R2: S= 300: 100: 3	S2	0.250	0.488	45	180	3
S3	0.460	0.897	45	180	6	
S4	0.940	1.83	45	180	12	
S5	1.32	2.57	/	/	3	
S6	2.40	4.68	/	/	6	
BS-330E ⁹ R1: R2: S= 300: 100: 3	S2	0.220	0.429	35	140	3
S3	0.500	0.975	35	140	6	
S4	0.970	1.89	35	140	12	
S5	1.37	2.67	/	/	3	
S6	2.52	4.91	/	/	6	
BS-380 ¹¹ R1: R2: S= 200: 67: 2	S2	0.220	0.429	30	120	2
S3	0.500	0.975	30	120	4	
S4	0.970	1.89	30	120	8	
S5	1.37	2.67	/	/	2	
S6	2.52	4.91	/	/	4	
BS-400 ¹² R1: R2: S= 200: 67: 2	S2	0.220	0.429	30	120	2
S3	0.490	0.956	30	120	4	
S4	0.970	1.89	30	120	8	
S5	1.35	2.63	/	/	2	
S6	2.51	4.89	/	/	4	

★ Note: The following value is applicable to 141922002 and subsequent ApoB reagents. ★

Abbreviated name	ApoB	Calibration Rule		Logit-Log(5P)		
Model	Level	Calibrator Value ²⁰		Sample Vol for Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		g/L	μmol/L			
BS-120 ¹ R1: R2: S= 300: 100: 3	S2	0.210	0.410	45	180	3
	S3	0.450	0.878	45	180	6
	S4	0.920	1.79	45	180	12
	S5	1.39	2.71	/	/	3
	S6	2.70	5.27	/	/	6
BS-180 ² R1: R2: S= 300: 100: 3	S2	0.210	0.410	45	180	3
	S3	0.450	0.878	45	180	6
	S4	0.920	1.79	45	180	12
	S5	1.39	2.71	/	/	3
	S6	2.70	5.27	/	/	6
BS-200 ³ R1: R2: S= 300: 100: 3	S2	0.240	0.468	45	180	3
	S3	0.550	1.07	45	180	6
	S4	0.980	1.91	45	180	12
	S5	1.36	2.65	/	/	3
	S6	2.70	5.27	/	/	6
BS-200E ⁴ R1: R2: S= 300: 100: 3	S2	0.240	0.468	35	140	3
	S3	0.510	0.995	35	140	6
	S4	0.980	1.91	35	140	12
	S5	1.42	2.77	/	/	3
	S6	2.70	5.27	/	/	6
BS-300 ⁷ R1: R2: S= 300: 100: 3	S2	0.240	0.468	45	180	3
	S3	0.520	1.01	45	180	6
	S4	1.00	1.95	45	180	12
	S5	1.37	2.67	/	/	3
	S6	2.70	5.27	/	/	6
BS-330 ⁸ R1: R2: S= 300: 100: 3	S2	0.240	0.468	45	180	3
	S3	0.550	1.07	45	180	6
	S4	0.980	1.91	45	180	12
	S5	1.36	2.65	/	/	3
	S6	2.70	5.27	/	/	6
BS-330E ⁹ R1: R2: S= 300: 100: 3	S2	0.240	0.468	35	140	3
	S3	0.510	0.995	35	140	6
	S4	0.980	1.91	35	140	12
	S5	1.42	2.77	/	/	3
	S6	2.70	5.27	/	/	6
BS-380 ¹¹ R1: R2: S= 200: 67: 2	S2	0.250	0.488	30	120	2
	S3	0.540	1.05	30	120	4
	S4	1.04	2.03	30	120	8
	S5	1.43	2.79	/	/	2
	S6	2.70	5.27	/	/	4
BS-400 ¹² R1: R2: S= 200: 67: 2	S2	0.240	0.468	30	120	2
	S3	0.510	0.995	30	120	4
	S4	1.02	1.99	30	120	8
	S5	1.39	2.71	/	/	2
	S6	2.70	5.27	/	/	4

Abbreviated name		ApoA1	Calibration Rule		Logit-Log(5P)		
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)	
		g/L	μmol/L				
BS-230 ⁵ R1: R2: S= 200: 67: 2	S2	0.150	5.36	2	13	117	
	S3	0.460	16.4	23.8	3	129	
	S4	0.980	35.0	23.4	6	125	
	S5	1.64	58.5	8	30	120	
	S6	2.90	104	2.3	/	/	
	S2	0.130	4.64	2	13	117	
BS-240E ⁶ R1: R2: S= 200: 67: 2	S3	0.560	20.0	2	34	102	
	S4	1.11	39.6	4	34	102	
	S5	1.65	58.9	8	25	100	
	S6	2.90	104	2.3	/	/	
	S2	0.130	4.64	2	13	117	
	S3	0.530	18.9	2	34	102	
BS-360E ¹⁰ R1: R2: S= 200: 67: 2	S4	1.10	39.3	4	34	102	
	S5	1.70	60.7	8	25	100	
	S6	2.90	104	2.3	/	/	
	S2	0.130	4.64	2	11	99	
	BS-430 ¹³ R1: R2: S= 200: 67: 2	S3	0.530	18.9	2	30	90
		S4	1.10	39.3	4	30	90
S5		1.72	61.4	8	25	100	
S6		2.92	104	2.3	/	/	
BS-480 ¹⁴ R1: R2: S= 200: 67: 2		S2	0.140	5.00	2	14	126
		S3	0.530	18.9	2	40	120
	S4	1.13	40.3	4	40	120	
	S5	1.79	63.9	8	30	120	
	S6	2.90	104	2.3	/	/	
	S2	0.130	4.64	2	11	99	
BS-600 ¹⁵ R1: R2: S= 200: 67: 2	S3	0.560	20.0	2	30	90	
	S4	1.12	40.0	4	30	90	
	S5	1.76	62.8	8	25	100	
	S6	2.92	104	2.3	/	/	
	BS-600M ¹⁶ R1: R2: S= 200: 67: 2	S2	0.150	5.36	2	10	90
		S3	0.560	20.0	2	25	75
S4		1.19	42.5	4	25	75	
S5		1.88	67.1	8	25	100	
S6		2.82	101	2.3	/	/	
S2		0.130	4.64	2	10	90	
BS-800 ¹⁷ R1: R2: S= 200: 67: 2	S3	0.530	18.9	2	30	90	
	S4	1.08	38.6	4	30	90	
	S5	1.65	58.9	8	25	100	
	S6	2.80	100	2.3	/	/	
	BS-2000 ¹⁸ R1: R2: S= 200: 67: 2	S2	0.150	5.36	2	10	90
		S3	0.560	20.0	2	25	75
S4		1.16	41.4	4	25	75	
S5		1.80	64.3	8	25	100	
S6		2.78	99.2	2.3	/	/	
S2		0.150	5.36	2	10	90	
BS-2800M ¹⁹ R1: R2: S= 200: 67: 2	S3	0.560	20.0	2	25	75	
	S4	1.19	42.5	4	25	75	
	S5	1.88	67.1	8	25	100	
	S6	2.82	101	2.3	/	/	

★ Note: The following value is applicable to 141922001 and before ApoB reagents. ★

Abbreviated name	ApoB	Calibration Rule		Logit-Log(5P)		
		Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)
Model	Level	g/L	μmol/L			
BS-230 ⁵ R1: R2: S= 200: 67: 2	S2	0.240	0.468	2	30	120
	S3	0.470	0.917	4	30	120
	S4	0.980	1.91	8	30	120
	S5	1.36	2.65	2	/	/
	S6	2.48	4.84	4	/	/
BS-240E ⁶ R1: R2: S= 300: 100: 3	S2	0.270	0.527	2	25	100
	S3	0.400	0.780	4	25	100
	S4	0.900	1.76	8	25	100
	S5	1.35	2.63	2	/	/
	S6	2.35	4.58	4	/	/
BS-360E ¹⁰ R1: R2: S= 300: 100: 3	S2	0.250	0.488	2	25	100
	S3	0.420	0.819	4	25	100
	S4	0.920	1.79	8	25	100
	S5	1.35	2.63	2	/	/
	S6	2.44	4.76	4	/	/
BS-430 ¹³ R1: R2: S= 200: 67: 2	S2	0.220	0.429	2	25	100
	S3	0.480	0.936	4	25	100
	S4	0.960	1.87	8	25	100
	S5	1.36	2.65	2	/	/
	S6	2.54	4.95	4	/	/
BS-480 ¹⁴ R1: R2: S= 200: 67: 2	S2	0.230	0.449	2	30	120
	S3	0.460	0.897	4	30	120
	S4	0.940	1.83	8	30	120
	S5	1.35	2.63	2	/	/
	S6	2.51	4.89	4	/	/
BS-600 ¹⁵ R1: R2: S= 200: 67: 2	S2	0.230	0.449	2	25	100
	S3	0.460	0.897	4	25	100
	S4	0.940	1.83	8	25	100
	S5	1.35	2.63	2	/	/
	S6	2.51	4.89	4	/	/
BS-600M ¹⁶ R1: R2: S= 200: 67: 2	S2	0.250	0.488	2	25	100
	S3	0.520	1.01	4	25	100
	S4	1.07	2.09	8	25	100
	S5	1.37	2.67	2	/	/
	S6	2.56	4.99	4	/	/
BS-800 ¹⁷ R1: R2: S= 200: 67: 2	S2	0.230	0.449	2	25	100
	S3	0.460	0.897	4	25	100
	S4	0.940	1.83	8	25	100
	S5	1.35	2.63	2	/	/
	S6	2.51	4.89	4	/	/
BS-2000 ¹⁸ R1: R2: S= 200: 67: 2	S2	0.240	0.468	2	25	100
	S3	0.500	0.975	4	25	100
	S4	1.01	1.97	8	25	100
	S5	1.33	2.59	2	/	/
	S6	2.53	4.93	4	/	/
BS-2800M ¹⁹ R1: R2: S= 200: 67: 2	S2	0.250	0.488	2	25	100
	S3	0.520	1.01	4	25	100
	S4	1.07	2.09	8	25	100
	S5	1.37	2.67	2	/	/
	S6	2.56	4.99	4	/	/

★ Note: The following value is applicable to 141922002 and subsequent ApoB reagents. ★

Abbreviated name	ApoB	Calibration Rule		Logit-Log(5P)		
		Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)
Model	Level	g/L	μmol/L			
BS-230 ⁵ R1: R2: S= 200: 67: 2	S2	0.240	0.468	2	30	120
	S3	0.520	1.01	4	30	120
	S4	1.00	1.95	8	30	120
	S5	1.41	2.75	2	/	/
	S6	2.70	5.27	4	/	/
	S2	0.220	0.429	2	25	100
BS-240E ⁶ R1: R2: S= 300: 100: 3	S3	0.490	0.956	4	25	100
	S4	0.920	1.79	8	25	100
	S5	1.43	2.79	2	/	/
	S6	2.70	5.27	4	/	/
	S2	0.220	0.429	2	25	100
	BS-360E ¹⁰ R1: R2: S= 300: 100: 3	S3	0.500	0.975	4	25
S4		0.970	1.89	8	25	100
S5		1.42	2.77	2	/	/
S6		2.70	5.27	4	/	/
S2		0.240	0.468	2	25	100
BS-430 ¹³ R1: R2: S= 200: 67: 2		S3	0.510	0.995	4	25
	S4	1.02	1.99	8	25	100
	S5	1.43	2.79	2	/	/
	S6	2.70	5.27	4	/	/
	S2	0.240	0.468	2	30	120
	BS-480 ¹⁴ R1: R2: S= 200: 67: 2	S3	0.520	1.01	4	30
S4		1.03	2.01	8	30	120
S5		1.43	2.79	2	/	/
S6		2.70	5.27	4	/	/
S2		0.240	0.468	2	25	100
BS-600 ¹⁵ R1: R2: S= 200: 67: 2		S3	0.530	1.03	4	25
	S4	1.02	1.99	8	25	100
	S5	1.43	2.79	2	/	/
	S6	2.70	5.27	4	/	/
	S2	0.250	0.488	2	25	100
	BS-600M ¹⁶ R1: R2: S= 200: 67: 2	S3	0.540	1.05	4	25
S4		1.09	2.13	8	25	100
S5		1.40	2.73	2	/	/
S6		2.70	5.27	4	/	/
S2		0.230	0.449	2	25	100
BS-800 ¹⁷ R1: R2: S= 200: 67: 2		S3	0.480	0.936	4	25
	S4	0.970	1.89	8	25	100
	S5	1.38	2.69	2	/	/
	S6	2.70	5.27	4	/	/
	S2	0.240	0.468	2	25	100
	BS-2000 ¹⁸ R1: R2: S= 200: 67: 2	S3	0.520	1.01	4	25
S4		1.04	2.03	8	25	100
S5		1.37	2.67	2	/	/
S6		2.70	5.27	4	/	/
S2		0.250	0.488	2	25	100
BS-2800M ¹⁹ R1: R2: S= 200: 67: 2		S3	0.540	1.05	4	25
	S4	1.09	2.13	8	25	100
	S5	1.40	2.73	2	/	/
	S6	2.70	5.27	4	/	/

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis
Русский	Эталонное значение	Объем пробы для	Объем	Объем пробы для анализа
Português	Valor de calibração	Volume da amostra	Volume de	Volume da amostra
Español	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
Italiano	Valore di calibrazione	Vol. campione per	Vol. diluente	Vol. campione per analisi
Türkçe	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

	HDL-C	LDL-C	ApoA1	ApoB
English	HDL-Cholesterol	LDL-Cholesterol	Apolipoprotein A1	Apolipoprotein B
Русский	Холестерин ЛПВП	Холестерин ЛПНП	Аполипопротеин А1	Аполипопротеин В
Português	Colesterol HDL	Colesterol LDL	Apolipoproteína A1	Apolipoproteína B
Español	Colesterol HDL	Colesterol LDL	Apolipoproteína A1	Apolipoproteína B
Italiano	Colesterolo HDL	Colesterolo LDL	Apolipoproteina A1	Apolipoproteina B
Türkçe	HDL-Kolesterol	LDL-Kolesterol	Apolipoprotein A1	Apolipoprotein B