

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português : Planilha de valores do Sistema de Medição BS da Mindray

Español : Hoja de valores del sistema de medición Mindray BS

Italiano : Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450,BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-300: BS-300, BS-320;

16.BS-1000M: BS-1000M, BS-1100M;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

18.BS-2800M:BS-2600M, BS-2800M;

9.BS-380: BS-380, BS-390;

19.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

10.BS-400: BS-400, BS-420;

LOT : 150424007

: 2025-10-07

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey
English	Calibration Value	Sample Vol for Dilution	Diluent Vol	
Русский	Эталонное значение	Объем пробы для разбавления	Объем разбавителя	
Português	Valor de calibração	Volume da amostra para diluição	Volume de diluente	
Español	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	
Italiano	Valore di calibrazione	Vol. campione per la diluizione	Vol. diluente	
Türkçe	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	
English	Sample Vol For Analysis	Expanded Uncertainty		
Русский	Объем пробы для анализа	Расширенная неопределенность		
Português	Volume da amostra para análise	Incerteza Expandida		
Español	Vol. muestra para análisis	Incertidumbre expandida		
Italiano	Vol. campione per analisi	Incertezza estesa		
Türkçe	Analiz için Numune Hacmi	Daha Uzun Süreli Belirsizlik		
English	HDL-C	LDL-C	ApoA1	ApoB
English	HDL-Cholesterol	LDL-Cholesterol	Apolipoprotein A1	Apolipoprotein B
Русский	Холестерин ЛПВП	Холестерин ЛПНП	Аполипопротеин A1	Аполипопротеин B
Português	Colesterol HDL	Colesterol LDL	Apolipoproteína A1	Apolipoproteína B
Español	Colesterol HDL	Colesterol LDL	Apolipoproteína A1	Apolipoproteína B
Italiano	Colesterolo HDL	Colesterolo LDL	Apolipoproteina A1	Apolipoproteina B

Türkçe		HDL-Kolesterol	LDL-Kolesterol	Apolipoprotein A1		Apolipoprotein B	
Abbreviated name	Unit	Calibration Rule			Two-point linear		
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
HDL-C	mmol/L	BS-120 <sup>1</sup>	1.65	0.06	BS-400 <sup>10</sup>	1.62	0.05
		BS-200 <sup>2</sup>	1.65	0.10	BS-430 <sup>11</sup>	1.63	0.07
		BS-200E <sup>3</sup>	1.63	0.06	BS-480 <sup>12</sup>	1.62	0.06
		BS-240 <sup>4</sup>	1.65	0.06	BS-600 <sup>13</sup>	1.62	0.06
		BS-240E <sup>5</sup>	1.63	0.07	BS-600M <sup>14</sup>	1.67	0.07
		BS-300 <sup>6</sup>	1.62	0.08	BS-800 <sup>15</sup>	1.63	0.07
		BS-330E <sup>7</sup>	1.63	0.06	BS-1000M <sup>16</sup>	1.62	0.07
		BS-360E <sup>8</sup>	1.63	0.07	BS-2000 <sup>17</sup>	1.59	0.07
		BS-380 <sup>9</sup>	1.62	0.06	BS-2800M <sup>18</sup>	1.62	0.07
	mg/dL	BS-120 <sup>1</sup>	63.8	2.2	BS-400 <sup>10</sup>	62.6	2.1
		BS-200 <sup>2</sup>	63.8	4	BS-430 <sup>11</sup>	63.0	2.7
		BS-200E <sup>3</sup>	63.0	2.3	BS-480 <sup>12</sup>	62.6	2.4
		BS-240 <sup>4</sup>	63.8	2.2	BS-600 <sup>13</sup>	62.6	2.4
		BS-240E <sup>5</sup>	63.0	2.7	BS-600M <sup>14</sup>	64.6	2.8
		BS-300 <sup>6</sup>	62.6	2.9	BS-800 <sup>15</sup>	63.0	2.7
		BS-330E <sup>7</sup>	63.0	2.3	BS-1000M <sup>16</sup>	62.6	2.7
		BS-360E <sup>8</sup>	63.0	2.7	BS-2000 <sup>17</sup>	61.5	2.7
		BS-380 <sup>9</sup>	62.6	2.3	BS-2800M <sup>18</sup>	62.6	2.7

Abbreviated name	Unit	Calibration Rule			Two-point linear		
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
LDL-C	mmol/L	BS-120 <sup>1</sup>	5.07	0.16	BS-400 <sup>10</sup>	5.43	0.24
		BS-200 <sup>2</sup>	5.07	0.14	BS-430 <sup>11</sup>	5.51	0.15
		BS-200E <sup>3</sup>	5.51	0.4	BS-480 <sup>12</sup>	5.43	0.3
		BS-240 <sup>4</sup>	5.07	0.16	BS-600 <sup>13</sup>	5.43	0.3
		BS-240E <sup>5</sup>	5.51	0.15	BS-600M <sup>14</sup>	5.43	0.15
		BS-300 <sup>6</sup>	5.43	0.3	BS-800 <sup>15</sup>	5.51	0.15
		BS-330E <sup>7</sup>	5.51	0.4	BS-1000M <sup>16</sup>	5.43	0.15
		BS-360E <sup>8</sup>	5.51	0.15	BS-2000 <sup>17</sup>	5.42	0.15
		BS-380 <sup>9</sup>	5.43	0.3	BS-2800M <sup>18</sup>	5.43	0.15
	mg/dL	BS-120 <sup>1</sup>	196	6	BS-400 <sup>10</sup>	210	9
		BS-200 <sup>2</sup>	196	5	BS-430 <sup>11</sup>	213	6
		BS-200E <sup>3</sup>	213	14	BS-480 <sup>12</sup>	210	12
		BS-240 <sup>4</sup>	196	6	BS-600 <sup>13</sup>	210	12
		BS-240E <sup>5</sup>	213	6	BS-600M <sup>14</sup>	210	6
		BS-300 <sup>6</sup>	210	12	BS-800 <sup>15</sup>	213	6
		BS-330E <sup>7</sup>	213	14	BS-1000M <sup>16</sup>	210	6
		BS-360E <sup>8</sup>	213	6	BS-2000 <sup>17</sup>	210	6
		BS-380 <sup>9</sup>	210	13	BS-2800M <sup>18</sup>	210	6

Abbreviated name		ApoA1		Calibration Rule		Logit-Log(5P)	
Model	Level	Calibrator Value		Expanded Uncertainty	Sample Vol For Dilution(μL)	Dilution Vol (μL)	Dilution Vol For Analysis(μL)
		g/L	μmol/L	g/L			

BS-120 <sup>1</sup>	0.9%NaCl	0.0000	0.00	/	/	/	/	/
	S2	0.0400	1.43	0.0016	0.06	20	180	3
	S3	0.330	11.8	0.013	0.5	5	194	32
	S4	0.810	28.9	0.03	1.1	9	187	35
	S5	1.44	51.4	0.06	2.0	45	180	12
	S6	2.50	89.3	0.10	3	/	/	3.5
BS-200 <sup>2</sup>	0.9%NaCl	0.0000	0.00	/	/	/	/	/
	S2	0.0400	1.43	0.0014	0.05	20	180	3
	S3	0.380	13.6	0.014	0.5	5	194	32
	S4	0.860	30.7	0.03	1.1	9	187	35
	S5	1.51	53.9	0.05	2.0	45	180	12
	S6	2.50	89.3	0.09	3	/	/	3.5
BS-200E <sup>3</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.0004	0.013	15	135	2
	S3	0.370	13.2	0.014	0.5	45	135	2
	S4	0.980	35.0	0.04	1.3	45	135	4
	S5	1.52	54.3	0.06	2.0	35	140	8
	S6	2.58	92.1	0.09	3	/	/	2.3
BS-300 <sup>6</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0200	0.714	0.0007	0.026	20	180	3
	S3	0.350	12.5	0.013	0.5	5	194	32
	S4	0.910	32.5	0.03	1.2	9	187	35
	S5	1.58	56.4	0.06	2.1	45	180	12
	S6	2.72	97.1	0.10	4	/	/	3.5
BS-330E <sup>7</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.0004	0.013	15	135	2
	S3	0.370	13.2	0.014	0.5	45	135	2
	S4	0.980	35.0	0.04	1.3	45	135	4
	S5	1.52	54.3	0.06	2.0	35	140	8
	S6	2.58	92.1	0.09	3	/	/	2.3
BS-360E <sup>8</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.00024	0.008	13	117	2
	S3	0.360	12.9	0.008	0.3	34	102	2
	S4	0.960	34.3	0.023	0.8	34	102	4
	S5	1.48	52.8	0.03	1.2	25	100	8
	S6	2.59	92.5	0.06	2.2	/	/	2.3
BS-380 <sup>9</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.0004	0.013	14	126	2
	S3	0.360	12.9	0.013	0.5	40	120	2
	S4	0.980	35.0	0.04	1.3	40	120	4
	S5	1.55	55.3	0.06	2.0	30	120	8
	S6	2.61	93.2	0.10	3	/	/	2.3
BS-400 <sup>10</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.0004	0.013	14	126	2
	S3	0.350	12.5	0.012	0.4	40	120	2
	S4	0.990	35.3	0.04	1.3	40	120	4
	S5	1.60	57.1	0.06	2.0	30	120	8

S6		2.62	93.5	0.09	3	/	/	2.3
Abbreviated name		ApoA1		Calibration Rule			Logit-Log(5P)	
Model	Level	Calibrator Value		Expanded Uncertainty		Dilution Vol For Analysis( $\mu$ L)	Sample Vol For Dilution( $\mu$ L)	Dilution Vol ( $\mu$ L)
		g/L	$\mu$ mol/L	g/L	$\mu$ mol/L			
BS-240 <sup>4</sup>	0.9%NaCl	0.0000	0.00	/	/	/	/	/
	S2	0.0400	1.43	0.0016	0.06	2	13	117
	S3	0.370	13.2	0.015	0.5	23.8	3	129
	S4	0.880	31.4	0.04	1.3	23.4	6	125
	S5	1.54	55.0	0.06	2.2	8	30	120
	S6	2.53	90.3	0.10	4	2.3	/	/
BS-240E <sup>5</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.00024	0.008	2	13	117
	S3	0.370	13.2	0.009	0.3	2	34	102
	S4	0.920	32.8	0.022	0.8	4	34	102
	S5	1.40	50.0	0.03	1.2	8	25	100
	S6	2.49	88.9	0.06	2.1	2.3	/	/
BS-430 <sup>11</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0200	0.714	0.0005	0.017	2	11	99
	S3	0.390	13.9	0.009	0.3	2	30	90
	S4	0.980	35.0	0.023	0.8	4	30	90
	S5	1.54	55.0	0.04	1.3	8	25	100
	S6	2.61	93.2	0.06	2.2	2.3	/	/
BS-480 <sup>12</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.0004	0.014	2	14	126
	S3	0.350	12.5	0.013	0.5	2	40	120
	S4	0.950	33.9	0.04	1.3	4	40	120
	S5	1.49	53.2	0.06	2.0	8	30	120
	S6	2.58	92.1	0.10	4	2.3	/	/
BS-600 <sup>13</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.0004	0.014	2	11	99
	S3	0.350	12.5	0.013	0.5	2	30	90
	S4	0.950	33.9	0.04	1.3	4	30	90
	S5	1.49	53.2	0.06	2.0	8	25	100
	S6	2.58	92.1	0.10	4	2.3	/	/
BS-600M <sup>14</sup>	0.9%NaCl	0.0000	0.00	/	/	/	/	/
	S2	0.0300	1.07	0.0007	0.026	2	10	90
	S3	0.420	15.0	0.010	0.4	2	25	75
	S4	1.03	36.8	0.025	0.9	4	25	75
	S5	1.66	59.3	0.04	1.4	8	25	100
	S6	2.58	92.1	0.06	2.2	2.3	/	/
BS-800 <sup>15</sup>	0.9%NaCl	0.0000	0.000	/	/	/	/	/
	S2	0.0100	0.357	0.00024	0.008	2	10	90
	S3	0.350	12.5	0.008	0.29	2	30	90
	S4	0.950	33.9	0.022	0.8	4	30	90
	S5	1.49	53.2	0.04	1.3	8	25	100
	S6	2.58	92.1	0.06	2.2	2.3	/	/

BS-1000M <sup>16</sup>	0.9%NaCl	0.0000	0.00	/	/	/	/	/
	S2	0.0300	1.07	0.0007	0.026	2	10	90
	S3	0.420	15.0	0.010	0.4	2	25	75
	S4	1.03	36.8	0.025	0.9	4	25	75
	S5	1.66	59.3	0.04	1.4	8	25	100
	S6	2.58	92.1	0.06	2.2	2.3	/	/
BS-2000 <sup>17</sup>	0.9%NaCl	0.0000	0.00	/	/	/	/	/
	S2	0.0300	1.07	0.0008	0.029	2	10	90
	S3	0.400	14.3	0.011	0.4	2	25	75
	S4	0.940	33.6	0.025	0.9	4	25	75
	S5	1.48	52.8	0.04	1.4	8	25	100
	S6	2.51	89.6	0.07	2.4	2.3	/	/
BS-2800M <sup>18</sup>	0.9%NaCl	0.0000	0.00	/	/	/	/	/
	S2	0.0300	1.07	0.0007	0.026	2	10	90
	S3	0.420	15.0	0.010	0.4	2	25	75
	S4	1.03	36.8	0.025	0.9	4	25	75
	S5	1.66	59.3	0.04	1.4	8	25	100
	S6	2.58	92.1	0.06	2.2	2.3	/	/
Abbreviated name		ApoB		Calibration Rule			Logit-Log(5P)	
Model	Level	Calibrator Value		Expanded Uncertainty		Sample Vol For Dilution(μL)	Dilution Vol (μL)	Dilution Vol For Analysis(μL)
		g/L	μmol/L	g/L	μmol/L			
BS-120 <sup>1</sup>	0.9%NaCl	0.000	0.000	/	/	/	/	/
	S2	0.290	0.566	0.010	0.019	45	180	3
	S3	0.610	1.19	0.021	0.04	45	180	6
	S4	1.18	2.30	0.04	0.08	45	180	12
	S5	1.71	3.33	0.06	0.11	/	/	3
	S6	3.23	6.30	0.11	0.22	/	/	6
BS-200 <sup>2</sup>	0.9%NaCl	0.000	0.000	/	/	/	/	/
	S2	0.300	0.585	0.012	0.023	45	180	3
	S3	0.700	1.37	0.027	0.05	45	180	6
	S4	1.27	2.48	0.05	0.10	45	180	12
	S5	1.75	3.41	0.07	0.13	/	/	3
	S6	3.27	6.38	0.13	0.25	/	/	6
BS-200E <sup>3</sup>	0.9%NaCl	0.000	0.000	/	/	/	/	/
	S2	0.350	0.683	0.013	0.025	35	140	3
	S3	0.750	1.46	0.027	0.05	35	140	6
	S4	1.27	2.48	0.05	0.09	35	140	12
	S5	1.78	3.47	0.06	0.13	/	/	3
	S6	3.24	6.32	0.12	0.23	/	/	6
BS-300 <sup>6</sup>	0.9%NaCl	0.000	0.000	/	/	/	/	/
	S2	0.300	0.585	0.021	0.04	45	180	3
	S3	0.690	1.35	0.05	0.09	45	180	6
	S4	1.36	2.65	0.09	0.18	45	180	12
	S5	1.89	3.69	0.13	0.25	/	/	3
	S6	3.29	6.42	0.23	0.4	/	/	6
	0.9%NaCl	0.000	0.000	/	/	/	/	/

BS-330E <sup>7</sup>	S2	0.350	0.683	0.013	0.025	35	140	3	
	S3	0.750	1.46	0.027	0.05	35	140	6	
	S4	1.27	2.48	0.05	0.09	35	140	12	
	S5	1.78	3.47	0.06	0.13	/	/	3	
	S6	3.24	6.32	0.12	0.23	/	/	6	
BS-360E <sup>8</sup>	0.9%NaCl	0.000	0.000	/	/	/	/	/	
	S2	0.350	0.683	0.010	0.020	25	100	2	
	S3	0.730	1.42	0.021	0.04	25	100	4	
	S4	1.18	2.30	0.03	0.07	25	100	8	
	S5	1.78	3.47	0.05	0.10	/	/	2	
BS-380 <sup>9</sup>	S6	3.21	6.26	0.09	0.18	/	/	4	
	0.9%NaCl	0.000	0.000	/	/	/	/	/	
	S2	0.300	0.585	0.011	0.021	30	120	2	
	S3	0.690	1.35	0.025	0.05	30	120	4	
	S4	1.35	2.63	0.05	0.09	30	120	8	
BS-400 <sup>10</sup>	S5	1.88	3.67	0.07	0.13	/	/	2	
	S6	3.34	6.51	0.12	0.24	/	/	4	
	0.9%NaCl	0.000	0.000	/	/	/	/	/	
	S2	0.310	0.605	0.015	0.029	30	120	2	
	S3	0.690	1.35	0.03	0.06	30	120	4	
BS-240 <sup>4</sup>	S4	1.33	2.59	0.06	0.12	30	120	8	
	S5	1.84	3.59	0.09	0.17	/	/	2	
	S6	3.34	6.51	0.16	0.3	/	/	4	
	Abbreviated name		ApoB		Calibration Rule			Logit-Log(5P)	
	Model	Level	Calibrator Value		Expanded Uncertainty		Dilution Vol For Analysis(μL)	Sample Vol For Dilution(μL)	Dilution Vol (μL)
g/L			μmol/L	g/L	μmol/L				
BS-240E <sup>5</sup>	0.9%NaCl	0.000	0.000	/	/	/	/	/	
	S2	0.320	0.624	0.011	0.022	2	30	120	
	S3	0.680	1.33	0.024	0.05	4	30	120	
	S4	1.28	2.50	0.05	0.09	8	30	120	
	S5	1.75	3.41	0.06	0.12	2	/	/	
BS-430 <sup>11</sup>	S6	3.27	6.38	0.12	0.23	4	/	/	
	0.9%NaCl	0.000	0.000	/	/	/	/	/	
	S2	0.310	0.605	0.009	0.017	2	25	100	
	S3	0.660	1.29	0.019	0.04	4	25	100	
	S4	1.19	2.32	0.03	0.07	8	25	100	
BS-430 <sup>11</sup>	S5	1.74	3.39	0.05	0.10	2	/	/	
	S6	3.20	6.24	0.09	0.18	4	/	/	
	0.9%NaCl	0.000	0.000	/	/	/	/	/	
	S2	0.300	0.585	0.009	0.017	2	25	100	
	S3	0.690	1.35	0.020	0.04	4	25	100	
BS-430 <sup>11</sup>	S4	1.30	2.54	0.04	0.07	8	25	100	
	S5	1.79	3.49	0.05	0.10	2	/	/	
	S6	3.26	6.36	0.09	0.18	4	/	/	
	0.9%NaCl	0.000	0.000	/	/	/	/	/	
	S2	0.280	0.546	0.010	0.020	2	30	120	

<b>BS-480<sup>12</sup></b>	<b>S3</b>	0.630	1.23	0.023	0.05	4	30	120
	<b>S4</b>	1.27	2.48	0.05	0.09	8	30	120
	<b>S5</b>	1.79	3.49	0.07	0.13	2	/	/
	<b>S6</b>	3.30	6.44	0.12	0.24	4	/	/
<b>BS-600<sup>13</sup></b>	<b>0.9%NaCl</b>	0.000	0.000	/	/	/	/	/
	<b>S2</b>	0.280	0.546	0.010	0.020	2	25	100
	<b>S3</b>	0.630	1.23	0.023	0.05	4	25	100
	<b>S4</b>	1.27	2.48	0.05	0.09	8	25	100
	<b>S5</b>	1.79	3.49	0.07	0.13	2	/	/
	<b>S6</b>	3.30	6.44	0.12	0.24	4	/	/
<b>BS-600M<sup>14</sup></b>	<b>0.9%NaCl</b>	0.000	0.000	/	/	/	/	/
	<b>S2</b>	0.320	0.624	0.009	0.018	2	25	100
	<b>S3</b>	0.750	1.46	0.021	0.04	4	25	100
	<b>S4</b>	1.43	2.79	0.04	0.08	8	25	100
	<b>S5</b>	1.75	3.41	0.05	0.10	2	/	/
	<b>S6</b>	3.24	6.32	0.09	0.18	4	/	/
<b>BS-800<sup>15</sup></b>	<b>0.9%NaCl</b>	0.000	0.000	/	/	/	/	/
	<b>S2</b>	0.280	0.546	0.008	0.016	2	25	100
	<b>S3</b>	0.630	1.23	0.018	0.04	4	25	100
	<b>S4</b>	1.27	2.48	0.04	0.07	8	25	100
	<b>S5</b>	1.79	3.49	0.05	0.10	2	/	/
	<b>S6</b>	3.30	6.44	0.10	0.19	4	/	/
<b>BS-1000M<sup>16</sup></b>	<b>0.9%NaCl</b>	0.000	0.000	/	/	/	/	/
	<b>S2</b>	0.320	0.624	0.009	0.018	2	25	100
	<b>S3</b>	0.730	1.42	0.021	0.04	4	25	100
	<b>S4</b>	1.45	2.83	0.04	0.08	8	25	100
	<b>S5</b>	1.84	3.59	0.05	0.10	2	/	/
	<b>S6</b>	3.34	6.51	0.10	0.19	4	/	/
<b>BS-2000<sup>17</sup></b>	<b>0.9%NaCl</b>	0.000	0.000	/	/	/	/	/
	<b>S2</b>	0.340	0.663	0.010	0.019	2	25	100
	<b>S3</b>	0.740	1.44	0.021	0.04	4	25	100
	<b>S4</b>	1.36	2.65	0.04	0.08	8	25	100
	<b>S5</b>	1.80	3.51	0.05	0.10	2	/	/
	<b>S6</b>	3.29	6.42	0.09	0.18	4	/	/
<b>BS-2800M<sup>18</sup></b>	<b>0.9%NaCl</b>	0.000	0.000	/	/	/	/	/
	<b>S2</b>	0.320	0.624	0.009	0.018	2	25	100
	<b>S3</b>	0.730	1.42	0.021	0.04	4	25	100
	<b>S4</b>	1.45	2.83	0.04	0.08	8	25	100
	<b>S5</b>	1.84	3.59	0.05	0.10	2	/	/
	<b>S6</b>	3.34	6.51	0.10	0.19	4	/	/

# **mindray**

## **Lipids Calibrator**

For use on: BS-1000M, BS-2000,  
BS-2800M, BS-600M,  
BS-800

**LOT** 150424007

**EXP** 2025-10-07

