

Prealbumin Calibrator



Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120:** BS-120, BS-130;

11. **BS-380:** BS-380, BS-390;

2. **BS-180:** BS-180, BS-190;

12. **BS-400:** BS-400, BS-420;

3. **BS-200:** BS-200, BS-220;

13. **BS-430:** BS-430, BS-450, BS-460;

4. **BS-200E:** BS-200E, BS-220E;

14. **BS-480:** BS-480, BS-490;

5. **BS-230:** BS-230, BS-240;

15. **BS-600:** BS-600, BS-620;

6. **BS-240E:** BS-240E, BS-240Pro;

16. **BS-600M:** BS-600M, BS-620M;

7. **BS-300:** BS-300, BS-320;

17. **BS-800:** BS-800, BS-820, BS-800M,

8. **BS-330:** BS-330, BS-350;

BS-820M, BS-1800, BS-1800plus;

9. **BS-330E: BS-330E(Serial Number starts with "XQ-"),**

BS-350E(Serial Number starts with "XS-");

18. **BS-2000:** BS-2000, BS-2200,

10. **BS-360E:** BS-360E, BS-370E, BS-350S, BS-360S,

BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

BS-2000M, BS-2200M;

19. **BS-2800M:** BS-2800M, BS-2600M.

20. **S1:0.9% NaCl, Conc. Of S1=0;**

Русский: S1: 0,9% NaCl, конц. S1=0;

Português: S1:0,9% NaCl, Conc. de S1=0;

Türkçe: S1:%0,9 NaCl, S1 Kons.=0;

Italiano: S1:0,9% NaCl, conc. di S1=0;

Español: S1:0,9% NaCl, Conc. de S1=0.

LOT: 151022002

: 2023-10-14

Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol for Dilution (µL)	Diluent Vol (µL)	Sample Vol For Analysis (µL)
		mg/L	µmol/L			
BS-120 ¹	S2	45.1	0.821	33	187	7
	S3	99.6	1.81	45	180	10.5
	S4	195	3.55	45	180	21
	S5	358	6.52	/	/	7
	S6	772	14.1	/	/	20
BS-180 ²	S2	45.1	0.821	33	187	7
	S3	99.6	1.81	45	180	10.5
	S4	195	3.55	45	180	21
	S5	358	6.52	/	/	7
	S6	772	14.1	/	/	20
BS-200 ³	S2	46.5	0.846	33	187	7
	S3	99.6	1.81	45	180	10.5
	S4	183	3.33	45	180	21
	S5	345	6.28	/	/	7
	S6	772	14.1	/	/	20
BS-200E ⁴	S2	47.8	0.870	30	170	8.4
	S3	108	1.97	35	140	12.8
	S4	207	3.77	35	140	25.9
	S5	367	6.68	/	/	8.4
	S6	811	14.8	/	/	24
BS-300 ⁷	S2	47.8	0.870	33	187	7
	S3	108	1.97	15	195	32
	S4	207	3.77	35	230	34
	S5	368	6.70	/	/	7
	S6	779	14.2	/	/	20

Prealbumin Calibrator



Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		mg/L	μmol/L			
BS-330 ⁸	S2	46.5	0.846	33	187	7
	S3	99.6	1.81	45	180	10.5
	S4	183	3.33	45	180	21
	S5	345	6.28	/	/	7
	S6	772	14.1	/	/	20
BS-330E ⁹	S2	47.8	0.870	30	170	8.4
	S3	108	1.97	35	140	12.8
	S4	207	3.77	35	140	25.9
	S5	367	6.68	/	/	8.4
	S6	811	14.8	/	/	24
BS-380 ¹¹	S2	45.1	0.821	24	136	7
	S3	99.6	1.81	10	130	32
	S4	195	3.55	21	138	34
	S5	358	6.52	/	/	7
	S6	772	14.1	/	/	20
BS-400 ¹²	S2	45.1	0.821	24	136	7
	S3	99.6	1.81	10	130	32
	S4	195	3.55	21	138	34
	S5	358	6.52	/	/	7
	S6	772	14.1	/	/	20
Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)
		mg/L	μmol/L			
BS-230 ⁵	S2	45.1	0.821	5.6	27	153
	S3	99.6	1.81	8.4	36	144
	S4	195	3.55	16.8	36	144
	S5	358	6.52	5.6	/	/
	S6	772	14.1	16	/	/
BS-240E ⁶	S2	45.1	0.821	5.6	21	119
	S3	99.6	1.81	25.6	10	130
	S4	195	3.55	27.2	21	138
	S5	358	6.52	5.6	/	/
	S6	772	14.1	16	/	/
BS-360E ¹⁰	S2	45.1	0.821	7	21	119
	S3	99.6	1.81	32	10	130
	S4	195	3.55	34	21	138
	S5	358	6.52	7	/	/
	S6	772	14.1	20	/	/
BS-430 ¹³	S2	45.1	0.821	7	18	102
	S3	99.6	1.81	32	10	130
	S4	195	3.55	34	21	138
	S5	358	6.52	7	/	/
	S6	772	14.1	20	/	/
BS-480 ¹⁴	S2	45.1	0.821	7	24	136
	S3	99.6	1.81	32	10	130
	S4	195	3.55	34	21	138
	S5	358	6.52	7	/	/
	S6	772	14.1	20	/	/

Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (µL)	Sample Vol for Dilution (µL)	Diluent Vol (µL)
		mg/L	µmol/L			
BS-600 ¹⁵	S2	45.1	0.821	7	18	102
	S3	99.6	1.81	32	10	130
	S4	195	3.55	34	21	138
	S5	358	6.52	7	/	/
	S6	772	14.1	20	/	/
	BS-600M ¹⁶	S2	/	/	5.6	18
S3		/	/	8.4	25	100
S4		/	/	16.8	25	100
S5		/	/	5.6	/	/
S6		/	/	16	/	/
BS-800 ¹⁷ R1: R2: S =120:45:4.2		S2	45.1	0.821	4.2	18
	S3	99.6	1.81	19.2	10	130
	S4	195	3.55	20.4	21	138
	S5	358	6.52	4.2	/	/
	S6	772	14.1	12	/	/
	BS-800 ¹⁷ R1: R2: S =200:75:7	S2	45.1	0.821	7	18
S3		99.6	1.81	32	10	130
S4		195	3.55	34	21	138
S5		358	6.52	7	/	/
S6		772	14.1	20	/	/
BS-2000 ¹⁸		S2	47.9	0.871	4.2	18
	S3	101	1.83	19.2	10	130
	S4	198	3.60	20.4	21	138
	S5	357	6.50	4.2	/	/
	S6	769	14.0	12	/	/
	BS-2800M ¹⁹	S2	50.7	0.923	4.2	18
S3		105	1.90	19.2	10	130
S4		205	3.72	20.4	21	138
S5		358	6.52	4.2	/	/
S6		849	15.5	12	/	/

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis
Русский	Эталонное значение	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
Português	Valor de calibração	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
Español	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
Italiano	Valore di calibrazione	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
Türkçe	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

PA	
English	Prealbumin
Русский	преальбумина
Português	pré-albumina
Español	Prealbúmina
Italiano	prealbumina
Türkçe	Prealbümin