

Prealbumin Calibrator

mindray

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120:** BS-120, BS-130;

11. **BS-380:** BS-380, BS-390;

2. **BS-180:** BS-180, BS-190;

12. **BS-400:** BS-400, BS-420;

3. **BS-200:** BS-200, BS-220;

13. **BS-430:** BS-430, BS-450, BS-460;

4. **BS-200E:** BS-200E, BS-220E;

14. **BS-480:** BS-480, BS-490;

5. **BS-230:** BS-230, BS-240;

15. **BS-600:** BS-600, BS-620;

6. **BS-240E:** BS-240E, BS-240Pro;

16. **BS-600M:** BS-600M, BS-620M;

7. **BS-300:** BS-300, BS-320;

17. **BS-800:** BS-800, BS-820, BS-800M,

8. **BS-330:** BS-330, BS-350;

BS-820M, BS-1800, BS-1800plus;

9. **BS-330E: BS-330E(Serial Number starts with "XQ-"),**

18. **BS-2000:** BS-2000, BS-2200,

BS-350E(Serial Number starts with "XS-");

BS-2000M, BS-2200M;

10. **BS-360E:** BS-360E, BS-370E, BS-350S, BS-360S,

19. **BS-2800M:** BS-2800M, BS-2600M.

BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

20. **S1:0.9% NaCl, Conc. Of S1=0;**

Русский: S1: 0,9% NaCl, конц. S1=0;

Português: S1:0,9% NaCl, Conc. de S1=0;

Türkçe: S1:%0,9 NaCl, S1 Kons.=0;

Italiano: S1:0,9% NaCl, conc. di S1=0;

Español: S1:0,9% NaCl, Conc. de S1=0.

LOT : 151023008

: 2025-03-23

Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol for Dilution (µL)	Diluent Vol (µL)	Sample Vol For Analysis (µL)
		mg/L	µmol/L			
BS-120 ¹	S2	59.0	1.07	33	187	7
	S3	113	2.06	45	180	10.5
	S4	224	4.08	45	180	21
	S5	414	7.53	/	/	7
	S6	977	17.8	/	/	20
BS-180 ²	S2	59.0	1.07	33	187	7
	S3	113	2.06	45	180	10.5
	S4	224	4.08	45	180	21
	S5	414	7.53	/	/	7
	S6	977	17.8	/	/	20
BS-200 ³	S2	62.5	1.14	33	187	7
	S3	117	2.13	45	180	10.5
	S4	215	3.91	45	180	21
	S5	414	7.53	/	/	7
	S6	956	17.4	/	/	20
BS-200E ⁴	S2	61.0	1.11	30	170	8.4
	S3	124	2.26	35	140	12.8
	S4	242	4.40	35	140	25.9
	S5	440	8.01	/	/	8.4
	S6	999	18.2	/	/	24
BS-300 ⁷	S2	62.0	1.13	33	187	7
	S3	118	2.15	15	195	32
	S4	238	4.33	35	230	34
	S5	425	7.74	/	/	7
	S6	999	18.2	/	/	20

Prealbumin Calibrator



Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		mg/L	μmol/L			
BS-330 ⁸	S2	62.5	1.14	33	187	7
	S3	117	2.13	45	180	10.5
	S4	215	3.91	45	180	21
	S5	414	7.53	/	/	7
	S6	956	17.4	/	/	20
BS-330E ⁹	S2	61.0	1.11	30	170	8.4
	S3	124	2.26	35	140	12.8
	S4	242	4.40	35	140	25.9
	S5	440	8.01	/	/	8.4
	S6	999	18.2	/	/	24
BS-380 ¹¹	S2	59.0	1.07	24	136	7
	S3	113	2.06	10	130	32
	S4	224	4.08	21	138	34
	S5	414	7.53	/	/	7
	S6	977	17.8	/	/	20
BS-400 ¹²	S2	59.0	1.07	24	136	7
	S3	113	2.06	10	130	32
	S4	224	4.08	21	138	34
	S5	414	7.53	/	/	7
	S6	977	17.8	/	/	20
Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)
		mg/L	μmol/L			
BS-230 ⁵	S2	59.0	1.07	5.6	27	153
	S3	113	2.06	8.4	36	144
	S4	224	4.08	16.8	36	144
	S5	414	7.53	5.6	/	/
	S6	977	17.8	16	/	/
BS-240E ⁶	S2	59.0	1.07	5.6	21	119
	S3	113	2.06	25.6	10	130
	S4	224	4.08	27.2	21	138
	S5	414	7.53	5.6	/	/
	S6	977	17.8	16	/	/
BS-360E ¹⁰	S2	62.0	1.13	7	21	119
	S3	117	2.13	32	10	130
	S4	230	4.19	34	21	138
	S5	424	7.72	7	/	/
	S6	999	18.2	20	/	/
BS-430 ¹³	S2	59.0	1.07	7	18	102
	S3	113	2.06	32	10	130
	S4	224	4.08	34	21	138
	S5	414	7.53	7	/	/
	S6	977	17.8	20	/	/
BS-480 ¹⁴	S2	59.0	1.07	7	24	136
	S3	113	2.06	32	10	130
	S4	224	4.08	34	21	138
	S5	414	7.53	7	/	/
	S6	977	17.8	20	/	/

Prealbumin Calibrator

Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (µL)	Sample Vol for Dilution (µL)	Diluent Vol (µL)
		mg/L	µmol/L			
BS-600 ¹⁵	S2	59.0	1.07	7	18	102
	S3	113	2.06	32	10	130
	S4	224	4.08	34	21	138
	S5	414	7.53	7	/	/
	S6	977	17.8	20	/	/
	BS-600M ¹⁶	S2	61.0	1.11	5.6	18
S3		130	2.37	8.4	25	100
S4		233	4.24	16.8	25	100
S5		414	7.53	5.6	/	/
S6		977	17.8	16	/	/
BS-800 ¹⁷ R1: R2: S =120:45:4.2		S2	59.0	1.07	4.2	18
	S3	113	2.06	19.2	10	130
	S4	224	4.08	20.4	21	138
	S5	414	7.53	4.2	/	/
	S6	977	17.8	12	/	/
	BS-800 ¹⁷ R1: R2: S =200:75:7	S2	59.0	1.07	7	18
S3		113	2.06	32	10	130
S4		224	4.08	34	21	138
S5		414	7.53	7	/	/
S6		977	17.8	20	/	/
BS-2000 ¹⁸		S2	61.0	1.11	4.2	18
	S3	113	2.06	19.2	10	130
	S4	218	3.97	20.4	21	138
	S5	411	7.48	4.2	/	/
	S6	931	16.9	12	/	/
	BS-2800M ¹⁹	S2	62.2	1.13	4.2	18
S3		120	2.18	19.2	10	130
S4		231	4.20	20.4	21	138
S5		410	7.46	4.2	/	/
S6		983	17.9	12	/	/

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis
Русский	Эталонное значение	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
Português	Valor de calibração	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
Español	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
Italiano	Valore di calibrazione	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
Türkçe	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

PA	
English	Prealbumin
Русский	преальбумина
Português	pré-albumina
Español	Prealbúmina
Italiano	prealbumina
Türkçe	Prealbümin