

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası

The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120**: BS-120, BS-130;

2. **BS-180**: BS-180, BS-190;

3. **BS-200**: BS-200, BS-220;

4. **BS-200E**: BS-200E, BS-220E;

5. **BS-230**: BS-230, BS-240;

6. **BS-240E**: BS240E, BS240Pro;

7. **BS-300**: BS-300, BS-320;

8. **BS-330**: BS-330, BS-350;

9. **BS-330E**: BS-330E (Serial Number starts with "XQ-"),

**BS-350E** (Serial Number starts with "XS-");

10. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)** (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

11. **BS-380**: BS-380, BS-390;

12. **BS-400**: BS-400, BS-420;

13. **BS-430**: BS-430, BS-450, BS-460;

14. **BS-480**: BS-480, BS-490;

15. **BS-600**: BS-600, BS-620;

16. **BS-600M**: BS-600M, BS-620M;

17. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

18. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

19. **BS-2800M**: BS-2600M, BS-2800M.

**LOT**: 151423008

**EXP**: 2025-04-19

| Abbreviated name             | Model                       | Unit   | Calibration Value | Model                         | Unit   | Calibration Value |
|------------------------------|-----------------------------|--------|-------------------|-------------------------------|--------|-------------------|
| CK-MB                        | <b>BS-120</b> <sup>1</sup>  | U/L    | 119               | <b>BS-380</b> <sup>11</sup>   | U/L    | 123               |
|                              |                             | µkat/L | 1.99              |                               | µkat/L | 2.05              |
|                              | <b>BS-180</b> <sup>2</sup>  | U/L    | 119               | <b>BS-400</b> <sup>12</sup>   | U/L    | 123               |
|                              |                             | µkat/L | 1.99              |                               | µkat/L | 2.05              |
|                              | <b>BS-200</b> <sup>3</sup>  | U/L    | 119               | <b>BS-430</b> <sup>13</sup>   | U/L    | 123               |
|                              |                             | µkat/L | 1.99              |                               | µkat/L | 2.05              |
|                              | <b>BS-200E</b> <sup>4</sup> | U/L    | 123               | <b>BS-480</b> <sup>14</sup>   | U/L    | 123               |
|                              |                             | µkat/L | 2.05              |                               | µkat/L | 2.05              |
|                              | <b>BS-230</b> <sup>5</sup>  | U/L    | 119               | <b>BS-600</b> <sup>15</sup>   | U/L    | 123               |
|                              |                             | µkat/L | 1.99              |                               | µkat/L | 2.05              |
|                              | <b>BS-240E</b> <sup>6</sup> | U/L    | 123               | <b>BS-600M</b> <sup>16</sup>  | U/L    | 125               |
|                              |                             | µkat/L | 2.05              |                               | µkat/L | 2.09              |
|                              | <b>BS-300</b> <sup>7</sup>  | U/L    | 123               | <b>BS-800</b> <sup>17</sup>   | U/L    | 123               |
|                              |                             | µkat/L | 2.05              |                               | µkat/L | 2.05              |
|                              | <b>BS-330</b> <sup>8</sup>  | U/L    | 119               | <b>BS-2000</b> <sup>18</sup>  | U/L    | 123               |
|                              |                             | µkat/L | 1.99              |                               | µkat/L | 2.05              |
|                              | <b>BS-330E</b> <sup>9</sup> | U/L    | 123               | <b>BS-2800M</b> <sup>19</sup> | U/L    | 125               |
|                              |                             | µkat/L | 2.05              |                               | µkat/L | 2.09              |
| <b>BS-360E</b> <sup>10</sup> | U/L                         | 123    |                   |                               |        |                   |
|                              | µkat/L                      | 2.05   |                   |                               |        |                   |

| English   | Abbreviated name         | Unit    | Calibration Value      | CK-MB                    |
|-----------|--------------------------|---------|------------------------|--------------------------|
| Русский   | сокращенное наименование | Прибор  | Эталонное значение     | МВ фракцию креатинкиназы |
| Português | Nome abreviado           | Unidade | Valor de calibração    | creatina quinase-MB      |
| Español   | nombre abreviado         | Unidad  | Valor de calibración   | creatina quinasa-MB      |
| Italiano  | abbreviazione            | Unità   | Valore di calibrazione | creatina chinasi-MB      |
| Türkçe    | kısaltılmış ad           | Ünite   | Kalibrasyon Değeri     | Kreatin Kinaz-MB         |