

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

10.BS-480: BS-480, BS-490;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

11.BS-600: BS-600, BS-620;

3.BS-200E: BS-200E, BS-220E,;

12.BS-600M: BS-600M, BS-620M;

4.BS-240: BS-230, BS-240;

13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-

5.BS-240E: BS240E, BS240Pro;

1800plus;

6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

14.BS-1000M: BS-1000M, BS-1100M;

330E(V35.00)(Serial Number starts with "W8-" and

software version starts with "35.00"), BS-350E(V35.00)

15.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

(Serial Number starts with "W9-" and software version

16.BS-2800M:BS-2600M, BS-2800M.

starts with "35.00");

17.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

7.BS-380: BS-380, BS-390;

18.For applicable models of the analyte, please refer to the reagent

8.BS-400: BS-400, BS-420;

parameter sheet and instruction.

9.BS-430: BS-430, BS-450, BS-460;



: 153924002



: 2025-03-10

| English | Abbreviated name | Unit | Calibration Rule | Model |
|-----------|--------------------------|---------|------------------------|---------|
| Русский | сокращенное наименование | Прибор | Принцип калибровки | модель |
| Português | Nome abreviado | Unidade | Regra de calibração | Modelo |
| Español | nombre abreviado | Unidad | Regla de calibración | modelo |
| Italiano | abbreviazione | Unità | Regola di calibrazione | modelli |
| Türkçe | kısaltılmış ad | Ünite | Kalibrasyon Kuralı | model |

| English | Expanded Uncertainty |
|-----------|------------------------------|
| Русский | Расширенная неопределенность |
| Português | Incerteza Expandida |
| Español | Incertidumbre expandida |
| Italiano | Incertezza estesa |
| Türkçe | Daha Uzun Süreli Belirsizlik |

| ACE | |
|-----------|---------------------------------------|
| English | Angiotensin Converting Enzyme |
| Русский | Ангиотензин-превращающего фермента |
| Português | Enzima de conversão da angiotensina |
| Español | Enzima convertidora de angiotensina |
| Italiano | Enzima convertitore dell'angiotensina |
| Türkçe | Anjiotensin Dönüştürücü Enzim |

| Abbreviated name | Unit | Calibration Rule | | | | Two-point linear | |
|------------------|--------|----------------------|------------------|----------------------|------------------------|------------------|----------------------|
| | | Model | Calibrator Value | Expanded Uncertainty | Model | Calibrator Value | Expanded Uncertainty |
| ACE | U/L | BS-120 ¹ | 76.8 | 3 | BS-430 ⁹ | 65.1 | 2.9 |
| | | BS-200 ² | 77.9 | 4 | BS-480 ¹⁰ | 65.1 | 2.9 |
| | | BS-200E ³ | 71.9 | 3 | BS-600 ¹¹ | 65.1 | 2.9 |
| | | BS-240 ⁴ | 76.8 | 3 | BS-600M ¹² | 71.1 | 3 |
| | | BS-240E ⁵ | 65.1 | 2.9 | BS-800 ¹³ | 65.1 | 2.9 |
| | | BS-360E ⁶ | 65.1 | 2.9 | BS-1000M ¹⁴ | 69.5 | 3 |
| | | BS-380 ⁷ | 66.6 | 3 | BS-2000 ¹⁵ | 68.2 | 3 |
| | | BS-400 ⁸ | 67.3 | 3 | BS-2800M ¹⁶ | 69.5 | 3 |
| | μkat/L | BS-120 ¹ | 1.28 | 0.06 | BS-430 ⁹ | 1.09 | 0.05 |
| | | BS-200 ² | 1.30 | 0.06 | BS-480 ¹⁰ | 1.09 | 0.05 |
| | | BS-200E ³ | 1.20 | 0.05 | BS-600 ¹¹ | 1.09 | 0.05 |
| | | BS-240 ⁴ | 1.28 | 0.06 | BS-600M ¹² | 1.19 | 0.05 |
| | | BS-240E ⁵ | 1.09 | 0.05 | BS-800 ¹³ | 1.09 | 0.05 |
| | | BS-360E ⁶ | 1.09 | 0.05 | BS-1000M ¹⁴ | 1.16 | 0.05 |
| | | BS-380 ⁷ | 1.11 | 0.05 | BS-2000 ¹⁵ | 1.14 | 0.05 |
| | | BS-400 ⁸ | 1.12 | 0.05 | BS-2800M ¹⁶ | 1.16 | 0.05 |

mindray

ACE Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 153924002

2025-03-10

