

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-230**: BS-230,BS-240;

3. **BS-240E**: BS-240E, BS-240Pro;

4. **BS-360E**: BS-360E, BS-370E,BS-350S,BS-360S,

BS-330E(V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

5. **BS-380**: BS-380, BS-390;

14. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Português: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Italiano: **Cal0: 0,9% NaCl, conc. di Cal0=0;**

6. **BS-400**: BS-400,BS-420;

7. **BS-430**: BS-430,BS-450,BS-460;

8. **BS-480**: BS-480,BS-490;

9. **BS-600**: BS-600, BS-620;

10. **BS-600M**: BS-600M, BS-620M;

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

Русский: **Cal0: 0,9% NaCl, конц. Cal0=0;**

Español: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Türkçe: **Cal0: %0,9 NaCl, Cal0 Kons.=0**

LOT : 154023001

EXP : 2024-07-05

| Abbreviated name | | RBP | Calibration Rule | | Spline |
|---|-------|---|-----------------------------------|-----------------------------------|-----------------------------------|
| Model | Level | Calibration Value ¹⁴ (mg/dL) | Sample Vol for Dilution / μ L | Diluent Vol/ μ L | Sample Vol For Analysis / μ L |
| BS-200E ¹ R1: R2: S= 180: 60: 4 | Cal1 | 2.74 | 45 | 135 | 4 |
| | Cal2 | 5.58 | 45 | 135 | 8 |
| | Cal3 | 12.5 | / | / | 4 |
| | Cal4 | 17.2 | / | / | 6 |
| BS-380 ⁵ R1: R2: S= 180: 60: 4 | Cal1 | 2.68 | 40 | 120 | 4 |
| | Cal2 | 5.23 | 40 | 120 | 8 |
| | Cal3 | 12.2 | / | / | 4 |
| | Cal4 | 16.8 | / | / | 6 |
| BS-400 ⁶ R1: R2: S= 180: 60: 4 | Cal1 | 2.68 | 40 | 120 | 4 |
| | Cal2 | 5.38 | 40 | 120 | 8 |
| | Cal3 | 12.4 | / | / | 4 |
| | Cal4 | 17.1 | / | / | 6 |
| Abbreviated name | | RBP | Calibration Rule | | Spline |
| Model | Level | Calibration Value ¹⁴ (mg/dL) | Sample Vol For Analysis / μ L | Sample Vol for Dilution / μ L | Diluent Vol/ μ L |
| BS-230 ² R1: R2: S= 180: 60: 4 | Cal1 | 2.74 | 4 | 35 | 105 |
| | Cal2 | 5.58 | 8 | 35 | 105 |
| | Cal3 | 12.5 | 4 | / | / |
| | Cal4 | 17.2 | 6 | / | / |

| Abbreviated name | | RBP | Calibration Rule | | Spline |
|---|-------|--|--------------------------------------|--------------------------------------|----------------------|
| Model | Level | Calibration Value ¹⁴ (mg/dL) | Sample Vol For Analysis / μ L | Sample Vol for Dilution / μ L | Diluent Vol/ μ L |
| BS-240E ³ R1: R2: S= 180: 60: 4 | Cal1 | 2.68 | 4 | 34 | 102 |
| | Cal2 | 5.23 | 8 | 34 | 102 |
| | Cal3 | 12.2 | 4 | / | / |
| | Cal4 | 16.8 | 6 | / | / |
| BS-360E ⁴ R1: R2: S= 180: 60: 4 | Cal1 | 2.68 | 4 | 34 | 102 |
| | Cal2 | 5.23 | 8 | 34 | 102 |
| | Cal3 | 12.2 | 4 | / | / |
| | Cal4 | 16.8 | 6 | / | / |
| BS-430 ⁷ R1: R2: S= 180: 60: 4 | Cal1 | 2.68 | 4 | 30 | 90 |
| | Cal2 | 5.23 | 8 | 30 | 90 |
| | Cal3 | 12.2 | 4 | / | / |
| | Cal4 | 16.8 | 6 | / | / |
| BS-480 ⁸ R1: R2: S= 180: 60: 4 | Cal1 | 2.78 | 4 | 40 | 120 |
| | Cal2 | 5.44 | 8 | 40 | 120 |
| | Cal3 | 12.6 | 4 | / | / |
| | Cal4 | 16.8 | 6 | / | / |
| BS-600 ⁹ R1: R2: S= 180: 60: 4 | Cal1 | 2.68 | 4 | 30 | 90 |
| | Cal2 | 5.23 | 8 | 30 | 90 |
| | Cal3 | 12.2 | 4 | / | / |
| | Cal4 | 16.8 | 6 | / | / |
| BS-600M ¹⁰ R1: R2: S= 180: 60: 4 | Cal1 | 2.81 | 4 | 25 | 75 |
| | Cal2 | 5.75 | 8 | 25 | 75 |
| | Cal3 | 12.4 | 4 | / | / |
| | Cal4 | 17.3 | 6 | / | / |
| BS-800 ¹¹ R1: R2: S= 180: 60: 4 | Cal1 | 2.68 | 4 | 30 | 90 |
| | Cal2 | 5.23 | 8 | 30 | 90 |
| | Cal3 | 12.2 | 4 | / | / |
| | Cal4 | 16.8 | 6 | / | / |
| BS-2000 ¹² R1: R2: S= 180: 60: 4 | Cal1 | 2.81 | 4 | 25 | 75 |
| | Cal2 | 5.60 | 8 | 25 | 75 |
| | Cal3 | 12.3 | 4 | / | / |
| | Cal4 | 17.1 | 6 | / | / |
| BS-2800M ¹³ R1: R2: S= 180: 60: 4 | Cal1 | 2.81 | 4 | 25 | 75 |
| | Cal2 | 5.75 | 8 | 25 | 75 |
| | Cal3 | 12.4 | 4 | / | / |
| | Cal4 | 17.3 | 6 | / | / |

| English | Abbreviated name | Calibration Rule | Model | Level |
|------------------|--------------------------|------------------------|---------|---------|
| Русский | сокращенное наименование | Принцип калибровки | модель | Уровень |
| Português | Nome abreviado | Regra de calibração | Modelo | Nível |
| Español | nombre abreviado | Regla de calibración | modelo | Nivel |
| Italiano | abbreviazione | Regola di calibrazione | modelli | Livello |
| Türkçe | kısaltılmış ad | Kalibrasyon Kuralı | model | Düzey |

| English | Calibration Value | Sample Vol. for Dilution | Diluent Vol. | Sample Vol. For Analysis |
|------------------|------------------------|---------------------------------|--------------------|--------------------------------|
| Русский | Эталонное значение | Объем пробы для разбавления | Объем разбавителя | Объем пробы для анализа |
| Português | Valor de calibração | Volume da amostra para diluição | Volume de diluente | Volume da amostra para análise |
| Español | Valor de calibración | Vol. muestra para dilución | Vol. diluyente | Vol. muestra para análisis |
| Italiano | Valore di calibrazione | Vol. campione per la diluizione | Vol. diluente | Vol. campione per analisi |
| Türkçe | Kalibrasyon Değeri | Dilüsyon için Numune Hacmi | Seyreltici Hacmi | Analiz için Numune Hacmi |

| RBP | |
|------------------|------------------------------------|
| English | Retinol Binding Protein |
| Русский | Ретинол-связывающий белок |
| Português | Proteína transportadora do retinol |
| Español | Proteína fijadora de retinol |
| Italiano | Proteina legante il retinolo |
| Türkçe | Retinol Bağlayıcı Protein |