

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450, BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-300: BS-300, BS-320;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)

17.BS-2800M:BS-2600M, BS-2800M.

(Serial Number starts with "W9-" and software version starts with "35.00");

18.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

9.BS-380: BS-380, BS-390;

19.Applicable models of the chemistry shall be subject to the

10.BS-400: BS-400, BS-420;

parameter list and instructions.

LOT

: 155823004



: 2025-06-13

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model
English	Expanded Uncertainty			
Русский	Расширенная неопределенность			
Português	Incerteza Expandida			
Español	Incertidumbre expandida			
Italiano	Incertezza estesa			
Türkçe	Daha Uzun Süreli Belirsizlik			
ADA				
English	Adenosine Deaminase			
Русский	является аденозиндезаминаза			
Português	Adenosina deaminase			
Español	adenosina desaminasa			
Italiano	adenosina deamminasi			
Türkçe	Adenozin Deaminazdır			

Abbreviated name	Unit	Calibration Rule			Two-point linear		
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
ADA	U/L	BS-120¹	51.4	1.1	BS-400¹⁰	51.4	1.1
		BS-200²	51.4	1.1	BS-430¹¹	51.4	1.1
		BS-200E³	50.4	1.1	BS-480¹²	51.4	1.1
		BS-240⁴	51.4	1.1	BS-600¹³	51.4	1.1
		BS-240E⁵	51.4	1.1	BS-600M¹⁴	51.8	1.1
		BS-300⁶	51.4	1.1	BS-800¹⁵	51.4	1.1
		BS-330E⁷	50.4	1.1	BS-2000¹⁶	52.0	1.2
		BS-360E⁸	51.4	1.1	BS-2800M¹⁷	51.8	1.1
		BS-380⁹	50.4	1.1			
	μkat/L	BS-120¹	0.858	0.019	BS-400¹⁰	0.858	0.019
		BS-200²	0.858	0.019	BS-430¹¹	0.858	0.019
		BS-200E³	0.842	0.019	BS-480¹²	0.858	0.019
		BS-240⁴	0.858	0.019	BS-600¹³	0.858	0.019
		BS-240E⁵	0.858	0.019	BS-600M¹⁴	0.865	0.019
		BS-300⁶	0.858	0.019	BS-800¹⁵	0.858	0.019
		BS-330E⁷	0.842	0.019	BS-2000¹⁶	0.868	0.019
		BS-360E⁸	0.858	0.019	BS-2800M¹⁷	0.865	0.019
		BS-380⁹	0.842	0.019			

mindray

ADA Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 155823004

2025-06-13

