

Rheumatoid Factor Calibrator

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-230**: BS-230, BS-240;

3. **BS-240E**: BS240E, BS240Pro;

4. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"),

BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

14. **a:0.9% NaCl is not needed, Conc. Of a=0;**

Português:a: 0,9% O Nacl não é needed, Conc. de a=0;

Italiano:a: 0.9% Il NaCl non è necessario, conc. di a=0;

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

5. **BS-380**: BS-380, BS-390;

6. **BS-400**: BS-400, BS-420;

7. **BS-430**: BS-430, BS-450, BS-460;

8. **BS-480**: BS-480, BS-490;

9. **BS-600**: BS-600, BS-620;

10. **BS-600M**: BS-600M, BS-620M;

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

Русский:a: 0,9% NaCl не необходим, конц. a=0;

Español:a: 0,9% El Nacl no es necesario, Conc. de a=0;

Türkçe:a: %0,9 NaCl gerekli değildir, a Kons.=0.

LOT: 158923004

EXP: 2025-02-07

Abbreviated name		RF II	Calibration Rule		Spline		
Model	Level	Calibration Value ¹⁴		Model	Level	Calibration Value ¹⁴	
		IU/mL				IU/mL	
BS-200E ¹	a	0		BS-480 ⁸	a	0	
	b	25.0			b	25.0	
	c	94.4			c	94.4	
	d	235			d	235	
	e	518			e	518	
BS-230 ²	a	0		BS-600 ⁹	a	0	
	b	25.0			b	25.0	
	c	94.4			c	94.4	
	d	235			d	235	
	e	518			e	518	
BS-240E ³	a	0		BS-600M ¹⁰	a	0	
	b	25.0			b	25.0	
	c	94.4			c	94.4	
	d	235			d	235	
	e	518			e	518	
BS-360E ⁴	a	0		BS-800 ¹¹	a	0	
	b	25.0			b	25.0	
	c	94.4			c	94.4	
	d	235			d	235	
	e	518			e	518	
BS-380 ⁵	S1	0		BS-2000 ¹²	a	0	
	S2	25.0			b	25.5	
	S3	94.4			c	94.7	
	S4	235			d	240	
	S5	518			e	514	

Rheumatoid Factor Calibrator

Abbreviated name		RF II	Calibration Rule		Spline
Model	Level	Calibration Value ¹⁴	Model	Level	Calibration Value ¹⁴
		IU/mL			IU/mL
BS-400 ⁶	S1	0	BS-2800M ¹³	a	/
	S2	25.0		b	/
	S3	94.4		c	/
	S4	235		d	/
	S5	518		e	/
BS-430 ⁷	a	0			
	b	25.0			
	c	94.4			
	d	235			
	e	518			

English	Abbreviated name	Calibration Rule	Model	Level	Calibration Value
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень	Эталонное значение
Português	Nome abreviado	Regra de calibração	Modelo	Nível	Valor de calibração
Español	nombre abreviado	Regla de calibración	modelo	Nivel	Valor de calibración
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello	Valore di calibrazione
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey	Kalibrasyon Değeri

RF	
English	Rheumatoid Factor
Русский	ревматоидные факторы
Português	Fator Reumatoide
Español	factor reumatoide
Italiano	fattore reumatoide
Türkçe	Romatoid Faktör