

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200**: BS-200, BS-220;

8. **BS-380**: BS-380, BS-390;

2. **BS-200E**: BS-200E, BS-220E;

9. **BS-400**: BS-400, BS-420;

3. **BS-230**: BS-230, BS-240;

10. **BS-430**: BS-430, BS-450, BS-460;

4. **BS-240E**: BS-240E, BS-240Pro;

11. **BS-480**: BS-480, BS-490;

5. **BS-300**: BS-300, BS-320;

12. **BS-600**: BS-600, BS-620;

6. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-");

13. **BS-600M**: BS-600M, BS-620M;

7. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00)(Serial Number starts with "W8-" and
software version starts with "35.00"), **BS-350E(V35.00)**
(Serial Number starts with "W9-" and software version
starts with "35.00");

14. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800,
BS-1800plus;

15. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

16. **BS-2800M**: BS-2600M, BS-2800M.

17. IFCC: $HbA1c(\%) = HbA1c(\mu mol/L) / Hb(\mu mol/L) \times 100$

NGSP/DCCT: $HbA1c(\%) = 91.5 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 2.15$

JCCLS: $HbA1c(\%) = 96.3 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 1.62$

LOT: 052423007

: 2025-01-20

Abbreviated name	Model	Unit	Assay Value	Range(Assay Value \pm 3SD)	1 SD	Model	Unit	Assay Value	Range(Assay Value \pm 3SD)	1 SD
HbA1c	BS-200 ¹	$\mu mol/L$	4.58	4.13 – 5.03	0.15	BS-400 ⁹	$\mu mol/L$	4.57	4.12 – 5.02	0.15
	BS-200E ²	$\mu mol/L$	4.57	4.12 – 5.02	0.15	BS-430 ¹⁰	$\mu mol/L$	4.62	4.16 – 5.08	0.15
	BS-230 ³	$\mu mol/L$	4.62	4.16 – 5.08	0.15	BS-480 ¹¹	$\mu mol/L$	4.62	4.16 – 5.08	0.15
	BS-240E ⁴	$\mu mol/L$	4.62	4.16 – 5.08	0.15	BS-600 ¹²	$\mu mol/L$	4.62	4.16 – 5.08	0.15
	BS-300 ⁵	$\mu mol/L$	4.51	4.06 – 4.96	0.15	BS-600M ¹³	$\mu mol/L$	4.56	4.11 – 5.01	0.15
	BS-330E ⁶	$\mu mol/L$	4.57	4.12 – 5.02	0.15	BS-800 ¹⁴	$\mu mol/L$	4.62	4.16 – 5.08	0.15
	BS-360E ⁷	$\mu mol/L$	4.62	4.16 – 5.08	0.15	BS-2000 ¹⁵	$\mu mol/L$	4.63	4.17 – 5.09	0.15
	BS-380 ⁸	$\mu mol/L$	4.57	4.12 – 5.02	0.15	BS-2800M ¹⁶	$\mu mol/L$	4.63	4.17 – 5.09	0.15
Hb	BS-200 ¹	$\mu mol/L$	126	114 – 138	4	BS-400 ⁹	$\mu mol/L$	126	114 – 138	4
	BS-200E ²	$\mu mol/L$	126	114 – 138	4	BS-430 ¹⁰	$\mu mol/L$	126	114 – 138	4
	BS-230 ³	$\mu mol/L$	126	114 – 138	4	BS-480 ¹¹	$\mu mol/L$	126	114 – 138	4
	BS-240E ⁴	$\mu mol/L$	126	114 – 138	4	BS-600 ¹²	$\mu mol/L$	126	114 – 138	4
	BS-300 ⁵	$\mu mol/L$	127	114 – 140	4	BS-600M ¹³	$\mu mol/L$	127	114 – 140	4
	BS-330E ⁶	$\mu mol/L$	126	114 – 138	4	BS-800 ¹⁴	$\mu mol/L$	126	114 – 138	4
	BS-360E ⁷	$\mu mol/L$	126	114 – 138	4	BS-2000 ¹⁵	$\mu mol/L$	127	114 – 140	4
	BS-380 ⁸	$\mu mol/L$	126	114 – 138	4	BS-2800M ¹⁶	$\mu mol/L$	126	114 – 138	4

HbA1C Control N



Abbreviated name	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	
HbA1c (%) ¹⁷	IFCC	BS-200 ¹	%	3.63	3.27 — 3.99	0.12	BS-400 ⁹	%	3.64	3.28 — 4.00	0.12
		BS-200E ²	%	3.63	3.27 — 3.99	0.12	BS-430 ¹⁰	%	3.68	3.32 — 4.04	0.12
		BS-230 ³	%	3.68	3.32 — 4.04	0.12	BS-480 ¹¹	%	3.68	3.32 — 4.04	0.12
		BS-240E ⁴	%	3.68	3.32 — 4.04	0.12	BS-600 ¹²	%	3.68	3.32 — 4.04	0.12
		BS-300 ⁵	%	3.54	3.19 — 3.89	0.12	BS-600M ¹³	%	3.60	3.24 — 3.96	0.12
		BS-330E ⁶	%	3.63	3.27 — 3.99	0.12	BS-800 ¹⁴	%	3.68	3.32 — 4.04	0.12
		BS-360E ⁷	%	3.68	3.32 — 4.04	0.12	BS-2000 ¹⁵	%	3.65	3.29 — 4.01	0.12
		BS-380 ⁸	%	3.63	3.27 — 3.99	0.12	BS-2800M ¹⁶	%	3.67	3.31 — 4.03	0.12
HbA1c (%) ¹⁷	NGSP / DCCT	BS-200 ¹	%	5.47	4.93 — 6.01	0.18	BS-400 ⁹	%	5.48	4.94 — 6.02	0.18
		BS-200E ²	%	5.47	4.93 — 6.01	0.18	BS-430 ¹⁰	%	5.51	4.96 — 6.06	0.18
		BS-230 ³	%	5.51	4.96 — 6.06	0.18	BS-480 ¹¹	%	5.51	4.96 — 6.06	0.18
		BS-240E ⁴	%	5.51	4.96 — 6.06	0.18	BS-600 ¹²	%	5.51	4.96 — 6.06	0.18
		BS-300 ⁵	%	5.39	4.86 — 5.92	0.18	BS-600M ¹³	%	5.45	4.91 — 5.99	0.18
		BS-330E ⁶	%	5.47	4.93 — 6.01	0.18	BS-800 ¹⁴	%	5.51	4.96 — 6.06	0.18
		BS-360E ⁷	%	5.51	4.96 — 6.06	0.18	BS-2000 ¹⁵	%	5.49	4.95 — 6.03	0.18
		BS-380 ⁸	%	5.47	4.93 — 6.01	0.18	BS-2800M ¹⁶	%	5.51	4.96 — 6.06	0.18
	JCCLS	BS-200 ¹	%	5.11	4.60 — 5.62	0.17	BS-400 ⁹	%	5.12	4.61 — 5.63	0.17
		BS-200E ²	%	5.12	4.61 — 5.63	0.17	BS-430 ¹⁰	%	5.16	4.65 — 5.67	0.17
		BS-230 ³	%	5.16	4.65 — 5.67	0.17	BS-480 ¹¹	%	5.16	4.65 — 5.67	0.17
		BS-240E ⁴	%	5.16	4.65 — 5.67	0.17	BS-600 ¹²	%	5.16	4.65 — 5.67	0.17
		BS-300 ⁵	%	5.03	4.53 — 5.53	0.17	BS-600M ¹³	%	5.09	4.59 — 5.59	0.17
		BS-330E ⁶	%	5.12	4.61 — 5.63	0.17	BS-800 ¹⁴	%	5.16	4.65 — 5.67	0.17
		BS-360E ⁷	%	5.16	4.65 — 5.67	0.17	BS-2000 ¹⁵	%	5.13	4.62 — 5.64	0.17
		BS-380 ⁸	%	5.12	4.61 — 5.63	0.17	BS-2800M ¹⁶	%	5.15	4.64 — 5.66	0.17

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione ±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

	HbA1c	Hb
English	Hemoglobin A1c	Hemoglobin
Русский	гемоглобин A1c	гемоглобин
Português	Hemoglobina A1c	Hemoglobina
Español	hemoglobina A1c	hemoglobina
Italiano	Emoglobina A1c	Emoglobina
Türkçe	Hemoglobin A1c	Hemoglobin