

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- | | |
|---|--|
| 1. BS-200 : BS-200, BS-220; | 8. BS-380 : BS-380, BS-390; |
| 2. BS-200E : BS-200E, BS-220E; | 9. BS-400 : BS-400, BS-420; |
| 3. BS-230 : BS-230, BS-240; | 10. BS-430 : BS-430, BS-450, BS-460; |
| 4. BS-240E : BS-240E, BS-240Pro; | 11. BS-480 : BS-480, BS-490; |
| 5. BS-300 : BS-300, BS-320; | 12. BS-600 : BS-600, BS-620; |
| 6. BS-330E : BS-330E (Serial Number starts with "XQ-"),
BS-350E (Serial Number starts with "XS-"); | 13. BS-800 : BS-800, BS-820, BS-800M, BS-820M, BS-1800,
BS-1800plus; |
| 7. BS-360E : BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00) (Serial Number starts with "W8-" and
software version starts with "35.00"), BS-350E(V35.00)
(Serial Number starts with "W9-" and software version
starts with "35.00"); | 14. BS-2000 : BS-2000, BS-2200, BS-2000M, BS-2200M; |
| | 15. BS-2800M : BS-2600M, BS-2800M. |

16. **IFCC**: $HbA1c(\%) = HbA1c(\mu mol/L) / Hb(\mu mol/L) \times 100$
NGSP/DCCT: $HbA1c(\%) = 91.5 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 2.15$
JCCLS: $HbA1c(\%) = 96.3 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 1.62$

LOT: 052522001

EXP: 2023-08-15

Abbreviated name	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD
HbA1c	BS-200 ¹	μmol/L	10.9	9.8 — 12.0	0.4	BS-400 ⁹	μmol/L	11.0	9.9 — 12.1	0.4
	BS-200E ²	μmol/L	10.9	9.8 — 12.0	0.4	BS-430 ¹⁰	μmol/L	10.9	9.8 — 12.0	0.4
	BS-230 ³	μmol/L	10.9	9.8 — 12.0	0.4	BS-480 ¹¹	μmol/L	10.9	9.8 — 12.0	0.4
	BS-240E ⁴	μmol/L	10.9	9.8 — 12.0	0.4	BS-600 ¹²	μmol/L	10.9	9.8 — 12.0	0.4
	BS-300 ⁵	μmol/L	10.9	9.8 — 12.0	0.4	BS-800 ¹³	μmol/L	10.9	9.8 — 12.0	0.4
	BS-330E ⁶	μmol/L	10.9	9.8 — 12.0	0.4	BS-2000 ¹⁴	μmol/L	10.9	9.8 — 12.0	0.4
	BS-360E ⁷	μmol/L	10.9	9.8 — 12.0	0.4	BS-2800M ¹⁵	μmol/L	10.9	9.8 — 12.0	0.4
	BS-380 ⁸	μmol/L	10.9	9.8 — 12.0	0.4					
Hb	BS-200 ¹	μmol/L	123	111 — 135	4	BS-400 ⁹	μmol/L	125	113 — 137	4
	BS-200E ²	μmol/L	126	114 — 138	4	BS-430 ¹⁰	μmol/L	125	113 — 137	4
	BS-230 ³	μmol/L	125	113 — 137	4	BS-480 ¹¹	μmol/L	125	113 — 137	4
	BS-240E ⁴	μmol/L	125	113 — 137	4	BS-600 ¹²	μmol/L	125	113 — 137	4
	BS-300 ⁵	μmol/L	126	114 — 138	4	BS-800 ¹³	μmol/L	125	113 — 137	4
	BS-330E ⁶	μmol/L	126	114 — 138	4	BS-2000 ¹⁴	μmol/L	125	113 — 137	4
	BS-360E ⁷	μmol/L	125	113 — 137	4	BS-2800M ¹⁵	μmol/L	125	113 — 137	4
	BS-380 ⁸	μmol/L	126	114 — 138	4					

HbA1c Control P



Abbreviated name	Model	Unit	Assay Value	Range(Assay Value \pm 3SD)	1 SD	Model	Unit	Assay Value	Range(Assay Value \pm 3SD)	1 SD	
HbA1c ₁₆ (%)	IFCC	BS-200 ¹	%	8.86	7.98 – 9.74	0.29	BS-400 ⁹	%	8.74	7.87 – 9.61	0.29
		BS-200E ²	%	8.67	7.81 – 9.53	0.29	BS-430 ¹⁰	%	8.69	7.83 – 9.55	0.29
		BS-230 ³	%	8.69	7.83 – 9.55	0.29	BS-480 ¹¹	%	8.69	7.83 – 9.55	0.29
		BS-240E ⁴	%	8.69	7.83 – 9.55	0.29	BS-600 ¹²	%	8.69	7.83 – 9.55	0.29
		BS-300 ⁵	%	8.67	7.81 – 9.53	0.29	BS-800 ¹³	%	8.69	7.83 – 9.55	0.29
		BS-330E ⁶	%	8.67	7.81 – 9.53	0.29	BS-2000 ¹⁴	%	8.70	7.84 – 9.56	0.29
		BS-360E ⁷	%	8.69	7.83 – 9.55	0.29	BS-2800M ¹⁵	%	8.71	7.85 – 9.57	0.29
		BS-380 ⁸	%	8.67	7.81 – 9.53	0.29					
HbA1c ₁₆ (%)	NGSP / DCCT	BS-200 ¹	%	10.3	9.3 – 11.3	0.3	BS-400 ⁹	%	10.1	9.1 – 11.1	0.3
		BS-200E ²	%	10.1	9.1 – 11.1	0.3	BS-430 ¹⁰	%	10.1	9.1 – 11.1	0.3
		BS-230 ³	%	10.1	9.1 – 11.1	0.3	BS-480 ¹¹	%	10.1	9.1 – 11.1	0.3
		BS-240E ⁴	%	10.1	9.1 – 11.1	0.3	BS-600 ¹²	%	10.1	9.1 – 11.1	0.3
		BS-300 ⁵	%	10.1	9.1 – 11.1	0.3	BS-800 ¹³	%	10.1	9.1 – 11.1	0.3
		BS-330E ⁶	%	10.1	9.1 – 11.1	0.3	BS-2000 ¹⁴	%	10.1	9.1 – 11.1	0.3
		BS-360E ⁷	%	10.1	9.1 – 11.1	0.3	BS-2800M ¹⁵	%	10.1	9.1 – 11.1	0.3
		BS-380 ⁸	%	10.1	9.1 – 11.1	0.3					
HbA1c ₁₆ (%)	JCCLS	BS-200 ¹	%	10.2	9.2 – 11.2	0.3	BS-400 ⁹	%	10.0	9.0 – 11.0	0.3
		BS-200E ²	%	10.0	9.0 – 11.0	0.3	BS-430 ¹⁰	%	10.0	9.0 – 11.0	0.3
		BS-230 ³	%	10.0	9.0 – 11.0	0.3	BS-480 ¹¹	%	10.0	9.0 – 11.0	0.3
		BS-240E ⁴	%	10.0	9.0 – 11.0	0.3	BS-600 ¹²	%	10.0	9.0 – 11.0	0.3
		BS-300 ⁵	%	10.0	9.0 – 11.0	0.3	BS-800 ¹³	%	10.0	9.0 – 11.0	0.3
		BS-330E ⁶	%	10.0	9.0 – 11.0	0.3	BS-2000 ¹⁴	%	10.0	9.0 – 11.0	0.3
		BS-360E ⁷	%	10.0	9.0 – 11.0	0.3	BS-2800M ¹⁵	%	10.0	9.0 – 11.0	0.3
		BS-380 ⁸	%	10.0	9.0 – 11.0	0.3					

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа \pm 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise \pm 3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo \pm 3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione \pm 3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri \pm 3SD)

	HbA1c	Hb
English	Hemoglobin A1c	Hemoglobin
Русский	гемоглобин A1c	гемоглобин
Português	Hemoglobina A1c	Hemoglobina
Español	hemoglobina A1c	hemoglobina
Italiano	Emoglobina A1c	Emoglobina
Türkçe	Hemoglobin A1c	Hemoglobin