

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120**: BS-120, BS-130;

7. **BS-380**: BS-380, BS-390;

2. **BS-200**: BS-200, BS-220;

8. **BS-400**: BS-400, BS-420;

3. **BS-200E**: BS-200E, BS-220E;

9. **BS-430**: BS-430, BS-450, BS-460;

4. **BS-230**: BS-230, BS-240;

10. **BS-480**: BS-480, BS-490;

5. **BS-240E**: BS-240E, BS-240Pro;

11. **BS-600**: BS-600, BS-620;

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

12. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

**BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E (V35.00)**

BS-1800, BS-1800plus;

**(Serial Number starts with "W9-" and software version**

13. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

**starts with "35.00");**

14. **BS-2800M**: BS-2600M, BS-2800M.

**LOT**: 053122003

**EXP**: 2023-09-17

Abbreviated name	MALB Control						
	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	
MALB	BS-120 <sup>1</sup>	mg/L	205	174	—	236	10
		$\mu$ mol/L	3.12	2.64	—	3.59	0.15
	BS-200 <sup>2</sup>	mg/L	205	174	—	236	10
		$\mu$ mol/L	3.12	2.64	—	3.59	0.15
	BS-200E <sup>3</sup>	mg/L	195	166	—	224	10
		$\mu$ mol/L	2.96	2.52	—	3.40	0.15
	BS-230 <sup>4</sup>	mg/L	205	174	—	236	10
		$\mu$ mol/L	3.12	2.64	—	3.59	0.15
	BS-240E <sup>5</sup>	mg/L	203	173	—	233	10
		$\mu$ mol/L	3.09	2.63	—	3.54	0.15
	BS-360E <sup>6</sup>	mg/L	203	173	—	233	10
		$\mu$ mol/L	3.09	2.63	—	3.54	0.15
	BS-380 <sup>7</sup>	mg/L	195	166	—	224	10
		$\mu$ mol/L	2.96	2.52	—	3.40	0.15
	BS-400 <sup>8</sup>	mg/L	202	172	—	232	10
		$\mu$ mol/L	3.07	2.61	—	3.53	0.15
	BS-430 <sup>9</sup>	mg/L	205	174	—	236	10
		$\mu$ mol/L	3.12	2.64	—	3.59	0.15
	BS-480 <sup>10</sup>	mg/L	203	173	—	233	10
		$\mu$ mol/L	3.09	2.63	—	3.54	0.15

Abbreviated name	MALB Control						
	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	
MALB	BS-600 <sup>11</sup>	mg/L	203	173	—	233	10
		$\mu$ mol/L	3.09	2.63	—	3.54	0.15
	BS-800 <sup>12</sup>	mg/L	203	173	—	233	10
		$\mu$ mol/L	3.09	2.63	—	3.54	0.15
	BS-2000 <sup>13</sup>	mg/L	207	176	—	238	10
		$\mu$ mol/L	3.15	2.68	—	3.62	0.15
	BS-2800M <sup>14</sup>	mg/L	202	172	—	232	10
		$\mu$ mol/L	3.07	2.61	—	3.53	0.15

English	Abbreviated name	MALB Control	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)
Русский	сокращенное наименование	Контроль микроальбумина	модель	Прибор	Результат анализа	Диапазон (результат анализа $\pm$ 3CO)
Português	Nome abreviado	Controle MALB	Modelo	Unidade	Valores da análise	Faixa (Valores da análise $\pm$ 3SD)
Español	nombre abreviado	Control MALB	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo $\pm$ 3SD)
Italiano	abbreviazione	Controllo MALB	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione $\pm$ 3 DS)
Türkçe	kısaltılmış ad	MALB Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri $\pm$ 3SD)

MALB	
English	Microalbumin
Русский	микроальбумина
Português	Microalbuminúria
Español	Microalbúmina
Italiano	Microalbumina
Türkçe	Mikroalbumin