

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

7.**BS-380**: BS-380, BS-390;

2.**BS-200**: BS-200, BS-220;

8.**BS-400**: BS-400, BS-420;

3.**BS-200E**: BS-200E, BS-220E;

9.**BS-430**: BS-430, BS-450, BS-460;

4.**BS-240**: BS-230, BS-240;

10.**BS-480**: BS-480, BS-490;

5.**BS-240E**: BS-240E, BS-240Pro;

11.**BS-600**: BS-600, BS-620;

6.**BS-360E**: BS-360E,BS-370E,BS-350S,BS-360S,

12.**BS-600M**: BS-600M, BS-620M;

BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)

13.**BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-880, BS-890, BS-1800, BS-1800plus;

(Serial Number starts with "W9-" and software version starts with "35.00");

14.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15.**BS-2800M**: BS-2600M, BS-2800M.

LOT : 053123003

EXP : 2024-09-29

MALB Control													
Abbreviated name	Model	Unit	Urine				CSF						
			Assay Value	Range (Assay Value±3SD)		1 SD	Assay Value	Range (Assay Value±3SD)		1 SD			
MALB	BS-120	1	mg/L	200	170	—	230	10	/	/	—	/	/
			µmol/L	3.04	2.58	—	3.50	0.15	/	/	—	/	/
	BS-200	2	mg/L	201	171	—	231	10	/	/	—	/	/
			µmol/L	3.06	2.60	—	3.51	0.15	/	/	—	/	/
	BS-200E	3	mg/L	198	168	—	228	10	/	/	—	/	/
			µmol/L	3.01	2.55	—	3.47	0.15	/	/	—	/	/
	BS-240	4	mg/L	200	170	—	230	10	/	/	—	/	/
			µmol/L	3.04	2.58	—	3.50	0.15	/	/	—	/	/
	BS-240E	5	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-360E	6	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-380	7	mg/L	198	168	—	228	10	/	/	—	/	/
			µmol/L	3.01	2.55	—	3.47	0.15	/	/	—	/	/
	BS-400	8	mg/L	201	171	—	231	10	/	/	—	/	/
			µmol/L	3.06	2.60	—	3.51	0.15	/	/	—	/	/
	BS-430	9	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-480	10	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/

Abbreviated name	MALB Control												
	Model	Unit	Urine					CSF					
			Assay Value	Range (Assay Value \pm 3SD)		1 SD	Assay Value	Range (Assay Value \pm 3SD)		1 SD			
MALB	BS-600	11	mg/L	199	169	—	229	10	/	/	—	/	/
			μ mol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-600M	12	mg/L	197	167	—	227	10	/	/	—	/	/
			μ mol/L	2.99	2.54	—	3.45	0.15	/	/	—	/	/
	BS-800	13	mg/L	199	169	—	229	10	/	/	—	/	/
			μ mol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-2000	14	mg/L	201	171	—	231	10	/	/	—	/	/
			μ mol/L	3.06	2.60	—	3.51	0.15	/	/	—	/	/
	BS-2800M	15	mg/L	198	168	—	228	10	/	/	—	/	/
			μ mol/L	3.01	2.55	—	3.47	0.15	/	/	—	/	/

English	Abbreviated name	MALB Control	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)
Русский	сокращенное наименование	Контроль микроальбумина	модель	Прибор	Результат анализа	Диапазон (результат анализа \pm 3CO)
Português	Nome abreviado	Controle MALB	Modelo	Unidade	Valores da análise	Faixa (Valores da análise \pm 3SD)
Español	nombre abreviado	Control MALB	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo \pm 3SD)
Italiano	abbreviazione	Controllo MALB	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione \pm 3 DS)
Türkçe	kısaltılmış ad	MALB Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri \pm 3SD)

MALB	
English	Microalbumin
Русский	микроальбумина
Português	Microalbuminúria
Español	Microalbúmina
Italiano	Microalbumina
Türkçe	Mikroalbumin