

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

7.**BS-380**: BS-380, BS-390;

2.**BS-200**: BS-200, BS-220;

8.**BS-400**: BS-400, BS-420;

3.**BS-200E**: BS-200E, BS-220E;

9.**BS-430**: BS-430, BS-450, BS-460;

4.**BS-240**: BS-230, BS-240;

10.**BS-480**: BS-480, BS-490;

5.**BS-240E**: BS-240E, BS-240Pro;

11.**BS-600**: BS-600, BS-620;

6.**BS-360E**: BS-360E,BS-370E,BS-350S,BS-360S,

12.**BS-600M**: BS-600M, BS-620M;

**BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)**

13.**BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-880, BS-890, BS-1800, BS-1800plus;

**(Serial Number starts with "W9-" and software version starts with "35.00");**

14.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15.**BS-2800M**: BS-2600M, BS-2800M.

**LOT : 053123005**

**: 2024-11-23**

MALB Control													
Abbreviated name	Model	Unit	Urine				CSF						
			Assay Value	Range (Assay Value±3SD)		1 SD	Assay Value	Range (Assay Value±3SD)		1 SD			
MALB	BS-120	1	mg/L	200	170	—	230	10	/	/	—	/	/
			µmol/L	3.04	2.58	—	3.50	0.15	/	/	—	/	/
	BS-200	2	mg/L	201	171	—	231	10	/	/	—	/	/
			µmol/L	3.06	2.60	—	3.51	0.15	/	/	—	/	/
	BS-200E	3	mg/L	198	168	—	228	10	/	/	—	/	/
			µmol/L	3.01	2.55	—	3.47	0.15	/	/	—	/	/
	BS-240	4	mg/L	200	170	—	230	10	/	/	—	/	/
			µmol/L	3.04	2.58	—	3.50	0.15	/	/	—	/	/
	BS-240E	5	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-360E	6	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-380	7	mg/L	198	168	—	228	10	/	/	—	/	/
			µmol/L	3.01	2.55	—	3.47	0.15	/	/	—	/	/
	BS-400	8	mg/L	201	171	—	231	10	/	/	—	/	/
			µmol/L	3.06	2.60	—	3.51	0.15	/	/	—	/	/
	BS-430	9	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-480	10	mg/L	199	169	—	229	10	/	/	—	/	/
			µmol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/

Abbreviated name	MALB Control												
	Model	Unit	Urine					CSF					
			Assay Value	Range (Assay Value $\pm$ 3SD)			1 SD	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD		
MALB	BS-600	11	mg/L	199	169	—	229	10	/	/	—	/	/
			$\mu$ mol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-600M	12	mg/L	197	167	—	227	10	/	/	—	/	/
			$\mu$ mol/L	2.99	2.54	—	3.45	0.15	/	/	—	/	/
	BS-800	13	mg/L	199	169	—	229	10	/	/	—	/	/
			$\mu$ mol/L	3.02	2.57	—	3.48	0.15	/	/	—	/	/
	BS-2000	14	mg/L	201	171	—	231	10	/	/	—	/	/
			$\mu$ mol/L	3.06	2.60	—	3.51	0.15	/	/	—	/	/
	BS-2800M	15	mg/L	198	168	—	228	10	/	/	—	/	/
			$\mu$ mol/L	3.01	2.55	—	3.47	0.15	/	/	—	/	/

English	Abbreviated name	MALB Control	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)
Русский	сокращенное наименование	Контроль микроальбумина	модель	Прибор	Результат анализа	Диапазон (результат анализа $\pm$ 3CO)
Português	Nome abreviado	Controle MALB	Modelo	Unidade	Valores da análise	Faixa (Valores da análise $\pm$ 3SD)
Español	nombre abreviado	Control MALB	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo $\pm$ 3SD)
Italiano	abbreviazione	Controllo MALB	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione $\pm$ 3 DS)
Türkçe	kısaltılmış ad	MALB Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri $\pm$ 3SD)

MALB	
English	Microalbumin
Русский	микроальбумина
Português	Microalbuminúria
Español	Microalbúmina
Italiano	Microalbumina
Türkçe	Mikroalbumin