

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası

**The data of each group is same.****Português: A dados de cada grupo é a mesma.****Italiano: la dati di ogni gruppo è la stessa.**

1.BS-120: BS-120, BS-130, BS-180, BS-190;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

3.BS-200E: BS-200E, BS-220E,;

4.BS-240: BS-230, BS-240;

5.BS-240E: BS240E, BS240Pro;

6.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

7.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

8.BS-380: BS-380, BS-390;

9.BS-400: BS-400, BS-420;

Русский: Данные совпадают во всех группах.**Español: la datos de cada grupo es la misma.****Türkçe: her grubun veri aynıdır.**

10.BS-430: BS-430, BS-450,BS-460;

11.BS-480: BS-480, BS-490;

12.BS-600: BS-600, BS-620;

13.BS-600M: BS-600M, BS-620M;

14.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

15.BS-1000M: BS-1000M, BS-1100M;

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

17.BS-2800M:BS-2600M, BS-2800M;

18.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

LOT: 053124005**EXP: 2025-10-06**

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

MALB

English	Microalbumin
Русский	микроальбумина
Português	Microalbuminúria
Español	Microalbúmina
Italiano	Microalbumina
Türkçe	Mikroalbumin

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Assay Value	Range(Assay Value±3SD)	1 SD
MALB	mg/L	BS-120 ¹	204	174 — 234	10	BS-430 ¹⁰	202	172 — 232	10
		BS-200 ²	204	174 — 234	10	BS-480 ¹¹	203	173 — 233	10
		BS-200E ³	202	172 — 232	10	BS-600 ¹²	203	173 — 233	10
		BS-240 ⁴	204	174 — 234	10	BS-600M ¹³	203	173 — 233	10
		BS-240E ⁵	202	172 — 232	10	BS-800 ¹⁴	202	172 — 232	10
	µmol/L	BS-360E ⁷	202	172 — 232	10	BS-1000M ¹⁵	203	173 — 233	10
		BS-380 ⁸	203	173 — 233	10	BS-2000 ¹⁶	205	175 — 235	10
		BS-400 ⁹	203	173 — 233	10	BS-2800M ¹⁷	203	173 — 233	10
		BS-120 ¹	3.10	2.64 — 3.56	0.15	BS-430 ¹⁰	3.07	2.61 — 3.53	0.15
		BS-200 ²	3.10	2.64 — 3.56	0.15	BS-480 ¹¹	3.09	2.63 — 3.54	0.15

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value \pm 3SD)			1 SD	Model	Assay Value	Range(Assay Value \pm 3SD)			1 SD
		BS-380⁸	3.09	2.63	—	3.54	0.15	BS-2000¹⁶	3.12	2.66	—	3.57	0.15
		BS-400⁹	3.09	2.63	—	3.54	0.15	BS-2800M¹⁷	3.09	2.63	—	3.54	0.15

mindray

MALB Control

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 053124005

2025-10-06

