

## Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray

Português : Planilha de valores do Sistema de Medição BS da Mindray

Español : Hoja de valores del sistema de medición Mindray BS

Italiano : Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский : Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

2.**BS-180**: BS-180, BS-190;

3.**BS-200**: BS-200, BS-220;

4.**BS-200E**: BS-200E, BS-220E;

5.**BS-230**: BS-230, BS-240;

6.**BS-240E**: BS240E, BS240Pro;

7.**BS-300**: BS-300, BS-320;

8.**BS-330**: BS-330, BS-350;

9.**BS-330E**:**BS-330E**(Serial Number starts with "XQ-"),

**BS-350E**(Serial Number starts with "XS-");

10.**BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

**BS-330E(V35.00)**(Serial Number starts with "W8-" and software version starts with "35.00"),

**BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

11.**BS-380**: BS-380, BS-390;

12.**BS-400**: BS-400, BS-420;

13.**BS-430**: BS-430, BS-450, BS-460;

14.**BS-480**: BS-480, BS-490;

15.**BS-600**: BS-600, BS-620;

16.**BS-600M**: BS-600M, BS-620M;

17.**BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

18.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

19.**BS-2800M**: BS-2600M, BS-2800M.

**LOT** : 056522002

**EXP** : 2024-02-08

Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD		
ADA	<b>BS-120</b> <sup>1</sup>	U/L	90.1	73.9	—	106.3	5.4	<b>BS-380</b> <sup>11</sup>	U/L	89.8	73.6	—	106.0	5.4
		µkat/L	1.50	1.23	—	1.78	0.09		µkat/L	1.50	1.23	—	1.77	0.09
	<b>BS-180</b> <sup>2</sup>	U/L	90.1	73.9	—	106.3	5.4	<b>BS-400</b> <sup>12</sup>	U/L	89.9	73.7	—	106.1	5.4
		µkat/L	1.50	1.23	—	1.78	0.09		µkat/L	1.50	1.23	—	1.77	0.09
	<b>BS-200</b> <sup>3</sup>	U/L	90.1	73.9	—	106.3	5.4	<b>BS-430</b> <sup>13</sup>	U/L	90.3	74.0	—	106.6	5.4
		µkat/L	1.50	1.23	—	1.78	0.09		µkat/L	1.51	1.24	—	1.78	0.09
	<b>BS-200E</b> <sup>4</sup>	U/L	89.8	73.6	—	106.0	5.4	<b>BS-480</b> <sup>14</sup>	U/L	90.3	74.0	—	106.6	5.4
		µkat/L	1.50	1.23	—	1.77	0.09		µkat/L	1.51	1.24	—	1.78	0.09
	<b>BS-230</b> <sup>5</sup>	U/L	89.5	73.4	—	105.6	5.4	<b>BS-600</b> <sup>15</sup>	U/L	90.3	74.0	—	106.6	5.4
		µkat/L	1.49	1.23	—	1.76	0.09		µkat/L	1.51	1.24	—	1.78	0.09
	<b>BS-240E</b> <sup>6</sup>	U/L	90.3	74.0	—	106.6	5.4	<b>BS-600M</b> <sup>16</sup>	U/L	92.2	75.6	—	108.8	5.5
		µkat/L	1.51	1.24	—	1.78	0.09		µkat/L	1.54	1.26	—	1.82	0.09
	<b>BS-300</b> <sup>7</sup>	U/L	90.1	73.9	—	106.3	5.4	<b>BS-800</b> <sup>17</sup>	U/L	90.3	74.0	—	106.6	5.4
		µkat/L	1.50	1.23	—	1.78	0.09		µkat/L	1.51	1.24	—	1.78	0.09
	<b>BS-330</b> <sup>8</sup>	U/L	90.1	73.9	—	106.3	5.4	<b>BS-2000</b> <sup>18</sup>	U/L	91.8	75.3	—	108.3	5.5
		µkat/L	1.50	1.23	—	1.78	0.09		µkat/L	1.53	1.26	—	1.81	0.09
	<b>BS-330E</b> <sup>9</sup>	U/L	89.8	73.6	—	106.0	5.4	<b>BS-2800M</b> <sup>19</sup>	U/L	91.7	75.2	—	108.2	5.5
		µkat/L	1.50	1.23	—	1.77	0.09		µkat/L	1.53	1.26	—	1.81	0.09
	<b>BS-360E</b> <sup>10</sup>	U/L	90.1	73.9	—	106.3	5.4							
		µkat/L	1.50	1.23	—	1.78	0.09							

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

ADA	
English	Adenosine Deaminase
Русский	является аденозиндезаминаза
Português	Adenosina deaminase
Español	adenosina desaminasa
Italiano	adenosina deamminasi
Türkçe	Adenozin Deaminazdır