

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120**: BS-120, BS-130;

7. **BS-380**: BS-380, BS-390;

2. **BS-200**: BS-200, BS-220;

8. **BS-400**: BS-400, BS-420;

3. **BS-200E**: BS-200E, BS-220E;

9. **BS-480**: BS-480, BS-490;

4. **BS-230**: BS-230, BS-240;

10. **BS-600**: BS-600, BS-620;

5. **BS-240E**: BS-240E, BS-240Pro;

11. **BS-600M**: BS-600M, BS-620M;

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

12. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)**

BS-1800, BS-1800plus;

(Serial Number starts with "W9-" and software version

13. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

starts with "35.00");

14. **BS-2800M**: BS-2600M, BS-2800M.

LOT: 057522004



: 2024-01-03

| Abbreviated name | G6PD Control (L) | | | | | | G6PD Control (H) | | | | | | | |
|------------------|-----------------------------|--------|-------------|-------------------------|---|------|------------------|-----------------------------|-------------|-------------------------|------|------|------|-----|
| | Model | Unit | Assay Value | Range (Assay Value±3SD) | | 1 SD | Model | Unit | Assay Value | Range (Assay Value±3SD) | | 1 SD | | |
| G6PD | BS-120 ¹ | U/L | 1053 | 841 | — | 1265 | 71 | BS-120 ¹ | U/L | 2231 | 1783 | — | 2679 | 149 |
| | | µkat/L | 17.6 | 14.0 | — | 21.1 | 1.2 | | µkat/L | 37.3 | 29.8 | — | 44.7 | 2.5 |
| | BS-200 ² | U/L | 1105 | 883 | — | 1327 | 74 | BS-200 ² | U/L | 2318 | 1852 | — | 2784 | 155 |
| | | µkat/L | 18.5 | 14.7 | — | 22.2 | 1.2 | | µkat/L | 38.7 | 30.9 | — | 46.5 | 2.6 |
| | BS-200E ³ | U/L | 934 | 746 | — | 1122 | 63 | BS-200E ³ | U/L | 1967 | 1572 | — | 2362 | 132 |
| | | µkat/L | 15.6 | 12.5 | — | 18.7 | 1.1 | | µkat/L | 32.8 | 26.3 | — | 39.4 | 2.2 |
| | BS-230 ⁴ | U/L | 1009 | 806 | — | 1212 | 68 | BS-230 ⁴ | U/L | 2144 | 1713 | — | 2575 | 144 |
| | | µkat/L | 16.9 | 13.5 | — | 20.2 | 1.1 | | µkat/L | 35.8 | 28.6 | — | 43.0 | 2.4 |
| | BS-240E ⁵ | U/L | 1054 | 842 | — | 1266 | 71 | BS-240E ⁵ | U/L | 2196 | 1755 | — | 2637 | 147 |
| | | µkat/L | 17.6 | 14.1 | — | 21.1 | 1.2 | | µkat/L | 36.7 | 29.3 | — | 44.0 | 2.5 |
| | BS-360E ⁶ | U/L | 972 | 777 | — | 1167 | 65 | BS-360E ⁶ | U/L | 2028 | 1620 | — | 2436 | 136 |
| | | µkat/L | 16.2 | 13.0 | — | 19.5 | 1.1 | | µkat/L | 33.9 | 27.1 | — | 40.7 | 2.3 |
| | BS-380 ⁷ | U/L | 1007 | 805 | — | 1209 | 67 | BS-380 ⁷ | U/L | 2111 | 1687 | — | 2535 | 141 |
| | | µkat/L | 16.8 | 13.4 | — | 20.2 | 1.1 | | µkat/L | 35.3 | 28.2 | — | 42.3 | 2.4 |
| | BS-400 ⁸ | U/L | 979 | 782 | — | 1176 | 66 | BS-400 ⁸ | U/L | 2086 | 1667 | — | 2505 | 140 |
| | | µkat/L | 16.3 | 13.1 | — | 19.6 | 1.1 | | µkat/L | 34.8 | 27.8 | — | 41.8 | 2.3 |
| | BS-480 ⁹ | U/L | 1032 | 825 | — | 1239 | 69 | BS-480 ⁹ | U/L | 2145 | 1714 | — | 2576 | 144 |
| | | µkat/L | 17.2 | 13.8 | — | 20.7 | 1.2 | | µkat/L | 35.8 | 28.6 | — | 43.0 | 2.4 |

| Abbreviated name | G6PD Control (L) | | | | | | G6PD Control (H) | | | | | | | |
|------------------|-------------------------------|-------------|-------------|-------------------------------|---|------|------------------|-------------------------------|-------------|-------------------------------|------|------|------|-----|
| | Model | Unit | Assay Value | Range (Assay Value \pm 3SD) | | 1 SD | Model | Unit | Assay Value | Range (Assay Value \pm 3SD) | | 1 SD | | |
| G6PD | BS-600 ¹⁰ | U/L | 1032 | 825 | — | 1239 | 69 | BS-600 ¹⁰ | U/L | 2145 | 1714 | — | 2576 | 144 |
| | | μ kat/L | 17.2 | 13.8 | — | 20.7 | 1.2 | | μ kat/L | 35.8 | 28.6 | — | 43.0 | 2.4 |
| | BS-600M ¹¹ | U/L | 1043 | 833 | — | 1253 | 70 | BS-600M ¹¹ | U/L | 2174 | 1737 | — | 2611 | 146 |
| | | μ kat/L | 17.4 | 13.9 | — | 20.9 | 1.2 | | μ kat/L | 36.3 | 29.0 | — | 43.6 | 2.4 |
| | BS-800 ¹² | U/L | 1032 | 825 | — | 1239 | 69 | BS-800 ¹² | U/L | 2145 | 1714 | — | 2576 | 144 |
| | | μ kat/L | 17.2 | 13.8 | — | 20.7 | 1.2 | | μ kat/L | 35.8 | 28.6 | — | 43.0 | 2.4 |
| | BS-2000 ¹³ | U/L | 1066 | 852 | — | 1280 | 71 | BS-2000 ¹³ | U/L | 2176 | 1739 | — | 2613 | 146 |
| | | μ kat/L | 17.8 | 14.2 | — | 21.4 | 1.2 | | μ kat/L | 36.3 | 29.0 | — | 43.6 | 2.4 |
| | BS-2800M ¹⁴ | U/L | 1084 | 866 | — | 1302 | 73 | BS-2800M ¹⁴ | U/L | 2223 | 1776 | — | 2670 | 149 |
| | | μ kat/L | 18.1 | 14.5 | — | 21.7 | 1.2 | | μ kat/L | 37.1 | 29.7 | — | 44.6 | 2.5 |

| English | Abbreviated name | G6PD Control | Model | Unit | Assay Value | Range (Assay Value \pm 3SD) |
|------------------|--------------------------|----------------|---------|---------|--------------------|---|
| Русский | сокращенное наименование | Контроль G6PD | модель | Прибор | Результат анализа | Диапазон(результат анализа \pm 3CO) |
| Português | Nome abreviado | Controle G6PD | Modelo | Unidade | Valores da análise | Faixa (Valores da análise \pm 3SD) |
| Español | nombre abreviado | Control G6PD | modelo | Unidad | Valor de ensayo | Rango (Valor de ensayo \pm 3SD) |
| Italiano | abbreviazione | Controllo G6PD | modelli | Unità | Valori di dosaggio | Intervallo(valore di concentrazione \pm 3 DS) |
| Türkçe | kısaltılmış ad | G6PD Kontrolü | model | Ünite | Tayin Değeri | Aralık (Tayin Değeri \pm 3SD) |

| G6PD | |
|------------------|-----------------------------------|
| English | Glucose-6-Phosphate Dehydrogenase |
| Русский | глюкозо-6-фосфатдегидрогеназа |
| Português | Glicose-6-fosfato desidrogenase |
| Español | Glucosa-6-fosfato deshidrogenasa |
| Italiano | Glucosio-6-fosfato deidrogenasi |
| Türkçe | Glikoz 6 Fosfat Dehidrogenaz |