

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120**: BS-120, BS-130;

7. **BS-380**: BS-380, BS-390;

2. **BS-200**: BS-200, BS-220;

8. **BS-400**: BS-400, BS-420;

3. **BS-200E**: BS-200E, BS-220E;

9. **BS-480**: BS-480, BS-490;

4. **BS-230**: BS-230, BS-240;

10. **BS-600**: BS-600, BS-620;

5. **BS-240E**: BS-240E, BS-240Pro;

11. **BS-600M**: BS-600M, BS-620M;

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

12. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

**BS-330E(V35.00)**(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)**

BS-1800, BS-1800plus;

(Serial Number starts with "W9-" and software version

13. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

starts with "35.00");

14. **BS-2800M**: BS-2600M, BS-2800M.

**LOT**: 057522005



: 2024-01-03

Abbreviated name	G6PD Control (L)						G6PD Control (H)							
	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD		
<b>G6PD</b>	<b>BS-120</b> <sup>1</sup>	U/L	1053	841	—	1265	71	<b>BS-120</b> <sup>1</sup>	U/L	2231	1783	—	2679	149
		µkat/L	17.6	14.0	—	21.1	1.2		µkat/L	37.3	29.8	—	44.7	2.5
	<b>BS-200</b> <sup>2</sup>	U/L	1105	883	—	1327	74	<b>BS-200</b> <sup>2</sup>	U/L	2318	1852	—	2784	155
		µkat/L	18.5	14.7	—	22.2	1.2		µkat/L	38.7	30.9	—	46.5	2.6
	<b>BS-200E</b> <sup>3</sup>	U/L	934	746	—	1122	63	<b>BS-200E</b> <sup>3</sup>	U/L	1967	1572	—	2362	132
		µkat/L	15.6	12.5	—	18.7	1.1		µkat/L	32.8	26.3	—	39.4	2.2
	<b>BS-230</b> <sup>4</sup>	U/L	1009	806	—	1212	68	<b>BS-230</b> <sup>4</sup>	U/L	2144	1713	—	2575	144
		µkat/L	16.9	13.5	—	20.2	1.1		µkat/L	35.8	28.6	—	43.0	2.4
	<b>BS-240E</b> <sup>5</sup>	U/L	1054	842	—	1266	71	<b>BS-240E</b> <sup>5</sup>	U/L	2196	1755	—	2637	147
		µkat/L	17.6	14.1	—	21.1	1.2		µkat/L	36.7	29.3	—	44.0	2.5
	<b>BS-360E</b> <sup>6</sup>	U/L	972	777	—	1167	65	<b>BS-360E</b> <sup>6</sup>	U/L	2028	1620	—	2436	136
		µkat/L	16.2	13.0	—	19.5	1.1		µkat/L	33.9	27.1	—	40.7	2.3
	<b>BS-380</b> <sup>7</sup>	U/L	1007	805	—	1209	67	<b>BS-380</b> <sup>7</sup>	U/L	2111	1687	—	2535	141
		µkat/L	16.8	13.4	—	20.2	1.1		µkat/L	35.3	28.2	—	42.3	2.4
	<b>BS-400</b> <sup>8</sup>	U/L	979	782	—	1176	66	<b>BS-400</b> <sup>8</sup>	U/L	2086	1667	—	2505	140
		µkat/L	16.3	13.1	—	19.6	1.1		µkat/L	34.8	27.8	—	41.8	2.3
	<b>BS-480</b> <sup>9</sup>	U/L	1032	825	—	1239	69	<b>BS-480</b> <sup>9</sup>	U/L	2145	1714	—	2576	144
		µkat/L	17.2	13.8	—	20.7	1.2		µkat/L	35.8	28.6	—	43.0	2.4

Abbreviated name	G6PD Control (L)						G6PD Control (H)							
	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD		
<b>G6PD</b>	<b>BS-600</b> <sup>10</sup>	U/L	1032	825	—	1239	69	<b>BS-600</b> <sup>10</sup>	U/L	2145	1714	—	2576	144
		$\mu$ kat/L	17.2	13.8	—	20.7	1.2		$\mu$ kat/L	35.8	28.6	—	43.0	2.4
	<b>BS-600M</b> <sup>11</sup>	U/L	1043	833	—	1253	70	<b>BS-600M</b> <sup>11</sup>	U/L	2174	1737	—	2611	146
		$\mu$ kat/L	17.4	13.9	—	20.9	1.2		$\mu$ kat/L	36.3	29.0	—	43.6	2.4
	<b>BS-800</b> <sup>12</sup>	U/L	1032	825	—	1239	69	<b>BS-800</b> <sup>12</sup>	U/L	2145	1714	—	2576	144
		$\mu$ kat/L	17.2	13.8	—	20.7	1.2		$\mu$ kat/L	35.8	28.6	—	43.0	2.4
	<b>BS-2000</b> <sup>13</sup>	U/L	1066	852	—	1280	71	<b>BS-2000</b> <sup>13</sup>	U/L	2176	1739	—	2613	146
		$\mu$ kat/L	17.8	14.2	—	21.4	1.2		$\mu$ kat/L	36.3	29.0	—	43.6	2.4
	<b>BS-2800M</b> <sup>14</sup>	U/L	1084	866	—	1302	73	<b>BS-2800M</b> <sup>14</sup>	U/L	2223	1776	—	2670	149
		$\mu$ kat/L	18.1	14.5	—	21.7	1.2		$\mu$ kat/L	37.1	29.7	—	44.6	2.5

English	Abbreviated name	G6PD Control	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)
<b>Русский</b>	сокращенное наименование	Контроль G6PD	модель	Прибор	Результат анализа	Диапазон(результат анализа $\pm$ 3CO)
<b>Português</b>	Nome abreviado	Controle G6PD	Modelo	Unidade	Valores da análise	Faixa (Valores da análise $\pm$ 3SD)
<b>Español</b>	nombre abreviado	Control G6PD	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo $\pm$ 3SD)
<b>Italiano</b>	abbreviazione	Controllo G6PD	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione $\pm$ 3 DS)
<b>Türkçe</b>	kısaltılmış ad	G6PD Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri $\pm$ 3SD)

<b>G6PD</b>	
<b>English</b>	Glucose-6-Phosphate Dehydrogenase
<b>Русский</b>	глюкозо-6-фосфатдегидрогеназа
<b>Português</b>	Glicose-6-fosfato desidrogenase
<b>Español</b>	Glucosa-6-fosfato deshidrogenasa
<b>Italiano</b>	Glucosio-6-fosfato deidrogenasi
<b>Türkçe</b>	Glikoz 6 Fosfat Dehidrogenaz