

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-240: BS-230, BS-240;

8.BS-600: BS-600, BS-620;

2.BS-240E: BS240E, BS240Pro;

9.BS-600M: BS-600M, BS-620M;

3.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

10.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

330E(V35.00)(Serial Number starts with "W8-" and software version

starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-"

and software version starts with "35.00");

11.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

4.BS-380: BS-380, BS-390;

12.BS-2800M:BS-2600M, BS-2800M;

5.BS-400: BS-400, BS-420;

6.BS-430: BS-430, BS-450,BS-460;

13.Applicable models of the chemistry shall be subject to the

7.BS-480: BS-480, BS-490;

parameter list and instructions.

LOT : 057824002

Expires: 2025-04-29

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

D-Dimer

English	D-Dimer
Русский	D-димера
Português	D-Dimer
Español	Dímero D
Italiano	D-Dimero
Türkçe	D-Dimer

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Assay Value	Range (Assay Value±3SD)		1 SD		
D-Dimer	µg/mL	BS-240 ¹	1.48	1.03	—	1.93	0.15	BS-240 ¹	15.3	12.9	—	17.7	0.8
		BS-240E ²	1.52	1.07	—	1.97	0.15	BS-240E ²	15.5	13.1	—	17.9	0.8
		BS-360E ³	1.52	1.07	—	1.97	0.15	BS-360E ³	15.5	13.1	—	17.9	0.8
		BS-380 ⁴	1.50	1.05	—	1.95	0.15	BS-380 ⁴	15.2	12.8	—	17.6	0.8
		BS-400 ⁵	1.50	1.05	—	1.95	0.15	BS-400 ⁵	15.2	12.8	—	17.6	0.8
		BS-430 ⁶	1.44	1.02	—	1.86	0.14	BS-430 ⁶	15.1	12.7	—	17.5	0.8
		BS-480 ⁷	1.52	1.07	—	1.97	0.15	BS-480 ⁷	15.5	13.1	—	17.9	0.8
		BS-600 ⁸	1.52	1.07	—	1.97	0.15	BS-600 ⁸	15.5	13.1	—	17.9	0.8
		BS-600M ⁹	1.40	0.98	—	1.82	0.14	BS-600M ⁹	15.2	12.8	—	17.6	0.8
		BS-800 ¹⁰	1.52	1.07	—	1.97	0.15	BS-800 ¹⁰	15.5	13.1	—	17.9	0.8
		BS-2000 ¹¹	1.43	1.01	—	1.85	0.14	BS-2000 ¹¹	15.1	12.7	—	17.5	0.8
		BS-2800M ¹²	1.40	0.98	—	1.82	0.14	BS-2800M ¹²	15.2	12.8	—	17.6	0.8

mindray

D-Dimer Control

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 057824002

2025-04-29

