

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

1. **BS-120**: BS-120, BS-130;

2. **BS-200**: BS-200, BS-220;

3. **BS-200E**: BS-200E, BS-220E;

4. **BS-230**: BS-230, BS-240;

5. **BS-240E**: BS-240E, BS-240Pro;

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

**BS-330E (V35.00) (Serial Number starts with "W8-" and**

**software version starts with "35.00"), BS-350E(V35.00)**

**(Serial Number starts with "W9-" and software version**

**starts with "35.00");**

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

7. **BS-380**: BS-380, BS-390;

8. **BS-400**: BS-400, BS-420;

9. **BS-430**: BS-430, BS-450, BS-460;

10. **BS-480**: BS-480, BS-490;

11. **BS-600**: BS-600, BS-620;

12. **BS-600M**: BS-600M, BS-620M;

13. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

14. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15. **BS-2800M**: BS-2600M, BS-2800M.

**LOT** : 057922005

**EXP** : 2023-11-30

Abbreviated name	TRF Control (L)						TRF Control (H)							
	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD		
TRF	<b>BS-120</b> <sup>1</sup>	g/L	1.08	0.92	—	1.24	0.05	<b>BS-120</b> <sup>1</sup>	g/L	2.24	1.90	—	2.58	0.11
		$\mu$ mol/L	13.6	11.6	—	15.6	0.6		$\mu$ mol/L	28.2	23.9	—	32.5	1.4
	<b>BS-200</b> <sup>2</sup>	g/L	1.07	0.91	—	1.23	0.05	<b>BS-200</b> <sup>2</sup>	g/L	2.25	1.91	—	2.59	0.11
		$\mu$ mol/L	13.5	11.5	—	15.5	0.6		$\mu$ mol/L	28.4	24.1	—	32.6	1.4
	<b>BS-200E</b> <sup>3</sup>	g/L	1.08	0.92	—	1.24	0.05	<b>BS-200E</b> <sup>3</sup>	g/L	2.28	1.94	—	2.62	0.11
		$\mu$ mol/L	13.6	11.6	—	15.6	0.6		$\mu$ mol/L	28.7	24.4	—	33.0	1.4
	<b>BS-230</b> <sup>4</sup>	g/L	1.08	0.92	—	1.24	0.05	<b>BS-230</b> <sup>4</sup>	g/L	2.26	1.92	—	2.60	0.11
		$\mu$ mol/L	13.6	11.6	—	15.6	0.6		$\mu$ mol/L	28.5	24.2	—	32.8	1.4
	<b>BS-240E</b> <sup>5</sup>	g/L	1.09	0.93	—	1.25	0.05	<b>BS-240E</b> <sup>5</sup>	g/L	2.29	1.95	—	2.63	0.11
		$\mu$ mol/L	13.7	11.7	—	15.8	0.6		$\mu$ mol/L	28.9	24.6	—	33.1	1.4
	<b>BS-360E</b> <sup>6</sup>	g/L	1.09	0.93	—	1.25	0.05	<b>BS-360E</b> <sup>6</sup>	g/L	2.29	1.95	—	2.63	0.11
$\mu$ mol/L		13.7	11.7	—	15.8	0.6	$\mu$ mol/L		28.9	24.6	—	33.1	1.4	
<b>BS-380</b> <sup>7</sup>	g/L	1.08	0.92	—	1.24	0.05	<b>BS-380</b> <sup>7</sup>	g/L	2.28	1.94	—	2.62	0.11	
	$\mu$ mol/L	13.6	11.6	—	15.6	0.6		$\mu$ mol/L	28.7	24.4	—	33.0	1.4	
<b>BS-400</b> <sup>8</sup>	g/L	1.08	0.92	—	1.24	0.05	<b>BS-400</b> <sup>8</sup>	g/L	2.27	1.93	—	2.61	0.11	
	$\mu$ mol/L	13.6	11.6	—	15.6	0.6		$\mu$ mol/L	28.6	24.3	—	32.9	1.4	
<b>BS-430</b> <sup>9</sup>	g/L	1.08	0.92	—	1.24	0.05	<b>BS-430</b> <sup>9</sup>	g/L	2.29	1.95	—	2.63	0.11	
	$\mu$ mol/L	13.6	11.6	—	15.6	0.6		$\mu$ mol/L	28.9	24.6	—	33.1	1.4	
<b>BS-480</b> <sup>10</sup>	g/L	1.09	0.93	—	1.25	0.05	<b>BS-480</b> <sup>10</sup>	g/L	2.29	1.95	—	2.63	0.11	
	$\mu$ mol/L	13.7	11.7	—	15.8	0.6		$\mu$ mol/L	28.9	24.6	—	33.1	1.4	
<b>BS-600</b> <sup>11</sup>	g/L	1.09	0.93	—	1.25	0.05	<b>BS-600</b> <sup>11</sup>	g/L	2.29	1.95	—	2.63	0.11	
	$\mu$ mol/L	13.7	11.7	—	15.8	0.6		$\mu$ mol/L	28.9	24.6	—	33.1	1.4	

Abbreviated name	TRF Control (L)							TRF Control (H)						
	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)			1 SD	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)			1 SD
TRF	BS-600M <sup>12</sup>	g/L	1.10	0.94	—	1.27	0.06	BS-600M <sup>12</sup>	g/L	2.30	1.96	—	2.65	0.12
		$\mu$ mol/L	13.9	11.8	—	16.0	0.8		$\mu$ mol/L	29.0	24.7	—	33.4	1.5
	BS-800 <sup>13</sup>	g/L	1.09	0.93	—	1.25	0.05	BS-800 <sup>13</sup>	g/L	2.29	1.95	—	2.63	0.11
		$\mu$ mol/L	13.7	11.7	—	15.8	0.6		$\mu$ mol/L	28.9	24.6	—	33.1	1.4
	BS-2000 <sup>14</sup>	g/L	1.09	0.93	—	1.25	0.05	BS-2000 <sup>14</sup>	g/L	2.28	1.94	—	2.62	0.11
		$\mu$ mol/L	13.7	11.7	—	15.8	0.6		$\mu$ mol/L	28.7	24.4	—	33.0	1.4
	BS-2800M <sup>15</sup>	g/L	1.10	0.94	—	1.27	0.06	BS-2800M <sup>15</sup>	g/L	2.30	1.96	—	2.65	0.12
		$\mu$ mol/L	13.9	11.8	—	16.0	0.8		$\mu$ mol/L	29.0	24.7	—	33.4	1.5

English	Abbreviated name	TRF Control	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)
Русский	сокращенное наименование	Контроль трансферрина	модель	Прибор	Результат анализа	Диапазон (результат анализа $\pm$ 3CO)
Português	Nome abreviado	Controle TRF	Modelo	Unidade	Valores da análise	Faixa (Valores da análise $\pm$ 3SD)
Español	nombre abreviado	Control TRF	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo $\pm$ 3SD)
Italiano	abbreviazione	Controllo TRF	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione $\pm$ 3 DS)
Türkçe	kısaltılmış ad	TRF Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri $\pm$ 3SD)

TRF	
English	Transferrin
Русский	трансферрина
Português	Transferrina
Español	Transferrina
Italiano	Transferrina
Türkçe	Transferin