

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

6. **BS-400**: BS-400, BS-420;

2. **BS-230**: BS-230, BS-240, BS-280;

7. **BS-430**: BS-430, BS-450, BS-460;

3. **BS-240E**: BS-240E, BS-240Pro;

8. **BS-480**: BS-480, BS-490;

4. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**

9. **BS-600**: BS-600, BS-620;

(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

10. **BS-600M**: BS-600M, BS-620M;

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

5. **BS-380**: BS-380, BS-390;

**LOT**: 058023007

**有效期**: 2024-12-18

Abbreviated name	Multimmun Control (L)						Multimmun Control (H)						
	Model	Unit	Assay Value	Range (Assay Value ± 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value ± 3SD)		1 SD	
<b>BS-200E</b> <sup>1</sup>	ng/mL	124	106	—	142	6	<b>BS-200E</b> <sup>1</sup>	ng/mL	466	397	—	535	23
	pmol/L	279	238	—	319	13		pmol/L	1047	892	—	1202	52
<b>BS-230</b> <sup>2</sup>	ng/mL	125	107	—	143	6	<b>BS-230</b> <sup>2</sup>	ng/mL	464	395	—	533	23
	pmol/L	281	240	—	321	13		pmol/L	1043	888	—	1198	52
<b>BS-240E</b> <sup>3</sup>	ng/mL	124	106	—	142	6	<b>BS-240E</b> <sup>3</sup>	ng/mL	465	396	—	534	23
	pmol/L	279	238	—	319	13		pmol/L	1045	890	—	1200	52
<b>BS-360E</b> <sup>4</sup>	ng/mL	124	106	—	142	6	<b>BS-360E</b> <sup>4</sup>	ng/mL	465	396	—	534	23
	pmol/L	279	238	—	319	13		pmol/L	1045	890	—	1200	52
<b>BS-380</b> <sup>5</sup>	ng/mL	124	106	—	142	6	<b>BS-380</b> <sup>5</sup>	ng/mL	466	397	—	535	23
	pmol/L	279	238	—	319	13		pmol/L	1047	892	—	1202	52
<b>BS-400</b> <sup>6</sup>	ng/mL	124	106	—	142	6	<b>BS-400</b> <sup>6</sup>	ng/mL	466	397	—	535	23
	pmol/L	279	238	—	319	13		pmol/L	1047	892	—	1202	52
<b>BS-430</b> <sup>7</sup>	ng/mL	124	106	—	142	6	<b>BS-430</b> <sup>7</sup>	ng/mL	465	396	—	534	23
	pmol/L	279	238	—	319	13		pmol/L	1045	890	—	1200	52
<b>BS-480</b> <sup>8</sup>	ng/mL	124	106	—	142	6	<b>BS-480</b> <sup>8</sup>	ng/mL	465	396	—	534	23
	pmol/L	279	238	—	319	13		pmol/L	1045	890	—	1200	52
<b>BS-600</b> <sup>9</sup>	ng/mL	124	106	—	142	6	<b>BS-600</b> <sup>9</sup>	ng/mL	465	396	—	534	23
	pmol/L	279	238	—	319	13		pmol/L	1045	890	—	1200	52
<b>BS-600M</b> <sup>10</sup>	ng/mL	123	105	—	141	6	<b>BS-600M</b> <sup>10</sup>	ng/mL	459	390	—	528	23
	pmol/L	276	236	—	317	13		pmol/L	1031	876	—	1186	52
<b>BS-800</b> <sup>11</sup>	ng/mL	124	106	—	142	6	<b>BS-800</b> <sup>11</sup>	ng/mL	465	396	—	534	23
	pmol/L	279	238	—	319	13		pmol/L	1045	890	—	1200	52
<b>BS-2000</b> <sup>12</sup>	ng/mL	121	103	—	139	6	<b>BS-2000</b> <sup>12</sup>	ng/mL	449	383	—	515	22
	pmol/L	272	231	—	312	13		pmol/L	1009	861	—	1157	49
<b>BS-2800M</b> <sup>13</sup>	ng/mL	123	105	—	141	6	<b>BS-2800M</b> <sup>13</sup>	ng/mL	459	390	—	528	23
	pmol/L	276	236	—	317	13		pmol/L	1031	876	—	1186	52

**FER**  
(Remark 2:  
The target value of FER in BS-230 is only applicable to BS-230 and BS-240)

Abbreviated name	Multimmun Control (L)					Multimmun Control (H)								
	Model	Unit	Assay Value	Range (AssayValue±3SD)		1 SD	Model	Unit	Assay Value	Range (AssayValue±3SD)		1 SD		
<b>MYO</b> (Remark 2: The target value of MYO in BS-240 is only applicable to BS-240)	<b>BS-240</b> <sup>2</sup>	ng/mL	79.2	67.2	—	91.2	4.0	<b>BS-240</b> <sup>2</sup>	ng/mL	294	249	—	339	15
	<b>BS-240E</b> <sup>3</sup>	ng/mL	76.6	65.2	—	88.0	3.8	<b>BS-240E</b> <sup>3</sup>	ng/mL	286	244	—	328	14
	<b>BS-360E</b> <sup>4</sup>	ng/mL	76.6	65.2	—	88.0	3.8	<b>BS-360E</b> <sup>4</sup>	ng/mL	286	244	—	328	14
	<b>BS-380</b> <sup>5</sup>	ng/mL	79.9	67.9	—	91.9	4.0	<b>BS-380</b> <sup>5</sup>	ng/mL	300	255	—	345	15
	<b>BS-400</b> <sup>6</sup>	ng/mL	79.9	67.9	—	91.9	4.0	<b>BS-400</b> <sup>6</sup>	ng/mL	300	255	—	345	15
	<b>BS-430</b> <sup>7</sup>	ng/mL	79.6	67.6	—	91.6	4.0	<b>BS-430</b> <sup>7</sup>	ng/mL	292	247	—	337	15
	<b>BS-480</b> <sup>8</sup>	ng/mL	76.6	65.2	—	88.0	3.8	<b>BS-480</b> <sup>8</sup>	ng/mL	286	244	—	328	14
	<b>BS-600</b> <sup>9</sup>	ng/mL	76.6	65.2	—	88.0	3.8	<b>BS-600</b> <sup>9</sup>	ng/mL	286	244	—	328	14
	<b>BS-600M</b> <sup>10</sup>	ng/mL	79.8	67.8	—	91.8	4.0	<b>BS-600M</b> <sup>10</sup>	ng/mL	289	247	—	331	14
	<b>BS-800</b> <sup>11</sup>	ng/mL	76.6	65.2	—	88.0	3.8	<b>BS-800</b> <sup>11</sup>	ng/mL	286	244	—	328	14
	<b>BS-2000</b> <sup>12</sup>	ng/mL	80.1	68.1	—	92.1	4.0	<b>BS-2000</b> <sup>12</sup>	ng/mL	294	249	—	339	15
	<b>BS-2800M</b> <sup>13</sup>	ng/mL	79.8	67.8	—	91.8	4.0	<b>BS-2800M</b> <sup>13</sup>	ng/mL	289	247	—	331	14
	<b>IgE</b> (Remark 2: The target value of IgE in BS-240 is only applicable to BS-240 and BS-280)	<b>BS-240</b> <sup>2</sup>	IU/mL	79.6	55.6	—	103.6	8.0	<b>BS-240</b> <sup>2</sup>	IU/mL	530	371	—	689
<b>BS-240E</b> <sup>3</sup>		IU/mL	79.5	55.5	—	103.5	8.0	<b>BS-240E</b> <sup>3</sup>	IU/mL	523	367	—	679	52
<b>BS-360E</b> <sup>4</sup>		IU/mL	79.8	55.8	—	103.8	8.0	<b>BS-360E</b> <sup>4</sup>	IU/mL	510	357	—	663	51
<b>BS-380</b> <sup>5</sup>		IU/mL	76.6	53.5	—	99.7	7.7	<b>BS-380</b> <sup>5</sup>	IU/mL	528	369	—	687	53
<b>BS-400</b> <sup>6</sup>		IU/mL	77.7	54.3	—	101.1	7.8	<b>BS-400</b> <sup>6</sup>	IU/mL	532	373	—	691	53
<b>BS-480</b> <sup>8</sup>		IU/mL	78.3	54.9	—	101.7	7.8	<b>BS-480</b> <sup>8</sup>	IU/mL	518	362	—	674	52
<b>BS-600</b> <sup>9</sup>		IU/mL	78.3	54.9	—	101.7	7.8	<b>BS-600</b> <sup>9</sup>	IU/mL	518	362	—	674	52
<b>BS-600M</b> <sup>10</sup>		IU/mL	80.9	56.6	—	105.2	8.1	<b>BS-600M</b> <sup>10</sup>	IU/mL	540	378	—	702	54
<b>BS-800</b> <sup>11</sup>		IU/mL	78.3	54.9	—	101.7	7.8	<b>BS-800</b> <sup>11</sup>	IU/mL	518	362	—	674	52
<b>BS-2000</b> <sup>12</sup>		IU/mL	80.2	56.2	—	104.2	8.0	<b>BS-2000</b> <sup>12</sup>	IU/mL	535	373	—	697	54
<b>BS-2800M</b> <sup>13</sup>	IU/mL	80.9	56.6	—	105.2	8.1	<b>BS-2800M</b> <sup>13</sup>	IU/mL	540	378	—	702	54	
English	Abbreviated name	Multimmun Control	Model	Unit	Assay Value	Range (Assay Value±3SD)								
<b>Русский</b>	сокращенное наименование	Контроль мультииммун	модель	Прибор	Результат анализа	Диапазон (результат анализа ± 3CO)								
<b>Português</b>	Nome abreviado	Controle Multimmun	Modelo	Прибор	Valores da análise	Faixa (Valores da análise ±3SD)								
<b>Español</b>	nombre abreviado	Control multimmun	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo ±3SD)								
<b>Italiano</b>	abbreviazione	Controllo Multimmun	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione ±3 DS)								
<b>Türkçe</b>	kısaltılmış ad	Multimmun Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)								
		FER			MYO	IgE								
<b>English</b>		Ferritin			Myoglobin	Immunoglobulin E								
<b>Русский</b>		ферритин			Миоглобин	Иммуноглобулин E								
<b>Português</b>		Ferritina			Mioglobina	Imunoglobulina E								
<b>Español</b>		Ferritina			Mioglobina	Inmunoglobulina E								
<b>Italiano</b>		Ferritina			Mioglobina	Immunoglobulina E								
<b>Türkçe</b>		Ferritin			Myoglobin	İmmünglobulin E								