

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

5. **BS-380**: BS-380, BS-390;

2. **BS-230**: BS-230, BS-240;

6. **BS-400**: BS-400, BS-420;

3. **BS-240E**: BS240E, BS240Pro;

7. **BS-430**: BS-430, BS-450, BS-460;

4. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

8. **BS-480**: BS-480, BS-490;

BS-330E(V35.00)(Serial Number starts with "W8-"

9. **BS-600**: BS-600, BS-620;

and software version starts with "35.00"),

10. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

BS-350E(V35.00) (Serial Number starts with "W9-"

11. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

and software version starts with "35.00");

12. **BS-2800M**: BS-2600M, BS-2800M.

LOT: 059122001

⌚: 2023-05-09

Abbreviated name	β2-Microglobulin Control (L)						β2-Microglobulin Control (H)							
	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD		
β2-MGII	BS-200E ¹	mg/L	1.02	0.87	—	1.17	0.05	BS-200E ¹	mg/L	2.93	2.49	—	3.37	0.15
		nmol/L	86.4	73.7	—	99.1	4.2		nmol/L	248	211	—	285	13
	BS-230 ²	mg/L	1.07	0.91	—	1.23	0.05	BS-230 ²	mg/L	2.96	2.52	—	3.40	0.15
		nmol/L	90.6	77.1	—	104.2	4.2		nmol/L	251	213	—	288	13
	BS-240E ³	mg/L	1.04	0.88	—	1.20	0.05	BS-240E ³	mg/L	2.97	2.52	—	3.42	0.15
		nmol/L	88.1	74.5	—	101.6	4.2		nmol/L	252	213	—	290	13
	BS-360E ⁴	mg/L	1.06	0.90	—	1.22	0.05	BS-360E ⁴	mg/L	2.97	2.52	—	3.42	0.15
		nmol/L	89.8	76.2	—	103.3	4.2		nmol/L	252	213	—	290	13
	BS-380 ⁵	mg/L	1.07	0.91	—	1.23	0.05	BS-380 ⁵	mg/L	2.97	2.52	—	3.42	0.15
		nmol/L	90.6	77.1	—	104.2	4.2		nmol/L	252	213	—	290	13
	BS-400 ⁶	mg/L	1.05	0.89	—	1.21	0.05	BS-400 ⁶	mg/L	2.96	2.52	—	3.40	0.15
		nmol/L	88.9	75.4	—	102.5	4.2		nmol/L	251	213	—	288	13
	BS-430 ⁷	mg/L	1.07	0.91	—	1.23	0.05	BS-430 ⁷	mg/L	2.96	2.52	—	3.40	0.15
		nmol/L	90.6	77.1	—	104.2	4.2		nmol/L	251	213	—	288	13
	BS-480 ⁸	mg/L	1.05	0.89	—	1.21	0.05	BS-480 ⁸	mg/L	3.00	2.55	—	3.45	0.15
		nmol/L	88.9	75.4	—	102.5	4.2		nmol/L	254	216	—	292	13
	BS-600 ⁹	mg/L	1.04	0.88	—	1.20	0.05	BS-600 ⁹	mg/L	2.94	2.50	—	3.38	0.15
		nmol/L	88.1	74.5	—	101.6	4.2		nmol/L	249	212	—	286	13
	BS-800 ¹⁰	mg/L	1.08	0.92	—	1.24	0.05	BS-800 ¹⁰	mg/L	2.97	2.52	—	3.42	0.15
		nmol/L	91.5	77.9	—	105.0	4.2		nmol/L	252	213	—	290	13
	BS-2000 ¹¹	mg/L	1.09	0.93	—	1.25	0.05	BS-2000 ¹¹	mg/L	3.05	2.59	—	3.51	0.15
		nmol/L	92.3	78.8	—	105.9	4.2		nmol/L	258	219	—	297	13
	BS-2800M ¹²	mg/L	1.06	0.90	—	1.22	0.05	BS-2800M ¹²	mg/L	3.01	2.56	—	3.46	0.15
		nmol/L	89.8	76.2	—	103.3	4.2		nmol/L	255	217	—	293	13

β2-Microglobulin Control

English	Abbreviated name	β2-Microglobulin Control	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	Контроль β2-микрोगлобулина	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Controlo β2-Microglobulina	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	Control β2-microglobulina	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	Controllo β2 Microglobulina	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	β2-Mikroglobulin Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

β2-MG	
English	β2-Microglobulin
Русский	β2-микрोगлобулина
Português	β2-Microglobulina
Español	β2-microglobulina
Italiano	β2-Microglobulina
Türkçe	β2-Mikroglobulin