

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

5. **BS-380**: BS-380, BS-390;

2. **BS-230**: BS-230, BS-240;

6. **BS-400**: BS-400, BS-420;

3. **BS-240E**: BS240E, BS240Pro;

7. **BS-430**: BS-430, BS-450, BS-460;

4. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

8. **BS-480**: BS-480, BS-490;

**BS-330E(V35.00)**(Serial Number starts with "W8-"

9. **BS-600**: BS-600, BS-620;

and software version starts with "35.00"),

10. **BS-600M**: BS-600M, BS-620M;

**BS-350E(V35.00)** (Serial Number starts with "W9-"

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

and software version starts with "35.00");

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

**LOT**: 059123004

**EXP**: 2024-11-10

Abbreviated name	β2-Microglobulin Control (L)						β2-Microglobulin Control (H)							
	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD		
β2-MGII	BS-200E <sup>1</sup>	mg/L	1.01	0.86	—	1.16	0.05	BS-200E <sup>1</sup>	mg/L	2.90	2.47	—	3.34	0.15
		nmol/L	85.5	72.8	—	98.3	4.2		nmol/L	246	209	—	283	13
	BS-230 <sup>2</sup>	mg/L	1.03	0.88	—	1.18	0.05	BS-230 <sup>2</sup>	mg/L	2.90	2.47	—	3.34	0.15
		nmol/L	87.2	74.5	—	99.9	4.2		nmol/L	246	209	—	283	13
	BS-240E <sup>3</sup>	mg/L	1.03	0.88	—	1.18	0.05	BS-240E <sup>3</sup>	mg/L	2.89	2.46	—	3.32	0.14
		nmol/L	87.2	74.5	—	99.9	4.2		nmol/L	245	208	—	281	12
	BS-360E <sup>4</sup>	mg/L	1.04	0.88	—	1.20	0.05	BS-360E <sup>4</sup>	mg/L	2.86	2.43	—	3.29	0.14
		nmol/L	88.1	74.5	—	101.6	4.2		nmol/L	242	206	—	279	12
	BS-380 <sup>5</sup>	mg/L	1.01	0.86	—	1.16	0.05	BS-380 <sup>5</sup>	mg/L	2.90	2.47	—	3.34	0.15
		nmol/L	85.5	72.8	—	98.3	4.2		nmol/L	246	209	—	283	13
	BS-400 <sup>6</sup>	mg/L	1.01	0.86	—	1.16	0.05	BS-400 <sup>6</sup>	mg/L	2.90	2.47	—	3.34	0.15
		nmol/L	85.5	72.8	—	98.3	4.2		nmol/L	246	209	—	283	13
	BS-430 <sup>7</sup>	mg/L	1.03	0.88	—	1.18	0.05	BS-430 <sup>7</sup>	mg/L	2.88	2.45	—	3.31	0.14
		nmol/L	87.2	74.5	—	99.9	4.2		nmol/L	244	208	—	280	12
	BS-480 <sup>8</sup>	mg/L	1.05	0.89	—	1.21	0.05	BS-480 <sup>8</sup>	mg/L	2.88	2.45	—	3.31	0.14
		nmol/L	88.9	75.4	—	102.5	4.2		nmol/L	244	208	—	280	12
	BS-600 <sup>9</sup>	mg/L	1.04	0.88	—	1.20	0.05	BS-600 <sup>9</sup>	mg/L	2.87	2.44	—	3.30	0.14
		nmol/L	88.1	74.5	—	101.6	4.2		nmol/L	243	207	—	280	12
	BS-600M <sup>10</sup>	mg/L	1.06	0.90	—	1.22	0.05	BS-600M <sup>10</sup>	mg/L	2.86	2.43	—	3.29	0.14
		nmol/L	89.8	76.2	—	103.3	4.2		nmol/L	242	206	—	279	12
BS-800 <sup>11</sup>	mg/L	1.05	0.89	—	1.21	0.05	BS-800 <sup>11</sup>	mg/L	2.87	2.44	—	3.30	0.14	
	nmol/L	88.9	75.4	—	102.5	4.2		nmol/L	243	207	—	280	12	
BS-2000 <sup>12</sup>	mg/L	1.05	0.89	—	1.21	0.05	BS-2000 <sup>12</sup>	mg/L	2.97	2.52	—	3.42	0.15	
	nmol/L	88.9	75.4	—	102.5	4.2		nmol/L	252	213	—	290	13	
BS-2800M <sup>13</sup>	mg/L	1.06	0.90	—	1.22	0.05	BS-2800M <sup>13</sup>	mg/L	2.86	2.43	—	3.29	0.14	
	nmol/L	89.8	76.2	—	103.3	4.2		nmol/L	242	206	—	279	12	

# β2-Microglobulin Control

English	Abbreviated name	β2-Microglobulin Control	Model	Unit	Assay Value	Range (Assay Value±3SD)
<b>Русский</b>	сокращенное наименование	Контроль β2-микροглобулина	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
<b>Português</b>	Nome abreviado	Controlo β2-Microglobulina	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
<b>Español</b>	nombre abreviado	Control β2-microglobulina	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
<b>Italiano</b>	abbreviazione	Controllo β2 Microglobulina	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione±3 DS)
<b>Türkçe</b>	kısaltılmış ad	β2-Mikroglobulin Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

β2-MG	
<b>English</b>	β2-Microglobulin
<b>Русский</b>	β2-микροглобулина
<b>Português</b>	β2-Microglobulina
<b>Español</b>	β2-microglobulina
<b>Italiano</b>	β2-Microglobulina
<b>Türkçe</b>	β2-Mikroglobulin