

## Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray

Português : Planilha de valores do Sistema de Medição BS da Mindray

Español : Hoja de valores del sistema de medición Mindray BS

Italiano : Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский : Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.**BS-200E**: BS-200E, BS-220E;

5.**BS-380**: BS-380, BS-390;

2.**BS-230**: BS-230, BS-240;

6.**BS-400**: BS-400, BS-420;

3.**BS-240E**: BS240E, BS240Pro;

7.**BS-430**: BS-430, BS-450, BS-460;

4.**BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

8.**BS-480**: BS-480, BS-490;

**BS-330E(V35.00)**(Serial Number starts with "W8-"

9.**BS-600**: BS-600, BS-620;

and software version starts with "35.00"),

10.**BS-600M**: BS-600M, BS-620M;

**BS-350E(V35.00)** (Serial Number starts with "W9-"

11.**BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

and software version starts with "35.00");

12.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13.**BS-2800M**: BS-2600M, BS-2800M.

**LOT** : 059123007

**EXP** : 2025-03-14

Abbreviated name	β2-Microglobulin Control (L)					β2-Microglobulin Control (H)				
	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD
β2-MGII	<b>BS-200E</b> <sup>1</sup>	mg/L	1.05	0.90 – 1.20	0.05	<b>BS-200E</b> <sup>1</sup>	mg/L	2.90	2.45 – 3.35	0.15
		nmol/L	88.9	76.2 – 101.6	4.2		nmol/L	246	208 – 284	13
	<b>BS-230</b> <sup>2</sup>	mg/L	1.06	0.91 – 1.21	0.05	<b>BS-230</b> <sup>2</sup>	mg/L	2.89	2.47 – 3.31	0.14
		nmol/L	89.8	77.1 – 102.5	4.2		nmol/L	245	209 – 280	12
	<b>BS-240E</b> <sup>3</sup>	mg/L	1.06	0.91 – 1.21	0.05	<b>BS-240E</b> <sup>3</sup>	mg/L	2.87	2.45 – 3.29	0.14
		nmol/L	89.8	77.1 – 102.5	4.2		nmol/L	243	208 – 279	12
	<b>BS-360E</b> <sup>4</sup>	mg/L	1.04	0.89 – 1.19	0.05	<b>BS-360E</b> <sup>4</sup>	mg/L	2.91	2.46 – 3.36	0.15
		nmol/L	88.1	75.4 – 100.8	4.2		nmol/L	246	208 – 285	13
	<b>BS-380</b> <sup>5</sup>	mg/L	1.05	0.90 – 1.20	0.05	<b>BS-380</b> <sup>5</sup>	mg/L	2.90	2.45 – 3.35	0.15
		nmol/L	88.9	76.2 – 101.6	4.2		nmol/L	246	208 – 284	13
	<b>BS-400</b> <sup>6</sup>	mg/L	1.05	0.90 – 1.20	0.05	<b>BS-400</b> <sup>6</sup>	mg/L	2.90	2.45 – 3.35	0.15
		nmol/L	88.9	76.2 – 101.6	4.2		nmol/L	246	208 – 284	13
	<b>BS-430</b> <sup>7</sup>	mg/L	1.06	0.91 – 1.21	0.05	<b>BS-430</b> <sup>7</sup>	mg/L	2.94	2.49 – 3.39	0.15
		nmol/L	89.8	77.1 – 102.5	4.2		nmol/L	249	211 – 287	13
	<b>BS-480</b> <sup>8</sup>	mg/L	1.04	0.89 – 1.19	0.05	<b>BS-480</b> <sup>8</sup>	mg/L	2.90	2.45 – 3.35	0.15
		nmol/L	88.1	75.4 – 100.8	4.2		nmol/L	246	208 – 284	13
	<b>BS-600</b> <sup>9</sup>	mg/L	1.01	0.86 – 1.16	0.05	<b>BS-600</b> <sup>9</sup>	mg/L	2.91	2.46 – 3.36	0.15
		nmol/L	85.5	72.8 – 98.3	4.2		nmol/L	246	208 – 285	13
	<b>BS-600M</b> <sup>10</sup>	mg/L	1.06	0.91 – 1.21	0.05	<b>BS-600M</b> <sup>10</sup>	mg/L	2.90	2.45 – 3.35	0.15
		nmol/L	89.8	77.1 – 102.5	4.2		nmol/L	246	208 – 284	13
<b>BS-800</b> <sup>11</sup>	mg/L	1.08	0.93 – 1.23	0.05	<b>BS-800</b> <sup>11</sup>	mg/L	2.96	2.51 – 3.41	0.15	
	nmol/L	91.5	78.8 – 104.2	4.2		nmol/L	251	213 – 289	13	
<b>BS-2000</b> <sup>12</sup>	mg/L	1.07	0.92 – 1.22	0.05	<b>BS-2000</b> <sup>12</sup>	mg/L	3.01	2.56 – 3.46	0.15	
	nmol/L	90.6	77.9 – 103.3	4.2		nmol/L	255	217 – 293	13	
<b>BS-2800M</b> <sup>13</sup>	mg/L	1.06	0.91 – 1.21	0.05	<b>BS-2800M</b> <sup>13</sup>	mg/L	2.90	2.45 – 3.35	0.15	
	nmol/L	89.8	77.1 – 102.5	4.2		nmol/L	246	208 – 284	13	

# β2-Microglobulin Control

English	Abbreviated name	β2-Microglobulin Control	Model	Unit	Assay Value	Range (Assay Value±3SD)
<b>Русский</b>	сокращенное наименование	Контроль β2-микροглобулина	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
<b>Português</b>	Nome abreviado	Controlo β2-Microglobulina	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
<b>Español</b>	nombre abreviado	Control β2-microglobulina	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
<b>Italiano</b>	abbreviazione	Controllo β2 Microglobulina	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione±3 DS)
<b>Türkçe</b>	kısaltılmış ad	β2-Mikroglobulin Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

## β2-MG

<b>English</b>	β2-Microglobulin
<b>Русский</b>	β2-микροглобулина
<b>Português</b>	β2-Microglobulina
<b>Español</b>	β2-microglobulina
<b>Italiano</b>	β2-Microglobulina
<b>Türkçe</b>	β2-Mikroglobulin