

**Value sheet of Mindray BS Measurement System**

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- 1.BS-200E: BS-200E, BS-220E,;
- 2.BS-240: BS-230, BS-240;
- 3.BS-240E: BS240E, BS240Pro;
- 4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");
- 5.BS-380: BS-380, BS-390;
- 6.BS-400: BS-400, BS-420;
- 7.BS-430: BS-430, BS-450,BS-460;

- 8.BS-480: BS-480, BS-490;
- 9.BS-600: BS-600, BS-620;
- 10.BS-600M: BS-600M, BS-620M;
- 11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
- 12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
- 13.BS-2800M:BS-2600M, BS-2800M;
- 14.For applicable analyzers of the analyte, please subject to the parameter sheet and instrument.

**LOT** : 059124001

**Expires**: 2025-04-30

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

**β2-MG**

English	β2-Microglobulin
Русский	β2-микροглобулина
Português	β2-Microglobulina
Español	β2-microglobulina
Italiano	β2-Microglobulina
Türkçe	β2-Mikroglöbulin

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Assay Value	Range (Assay Value±3SD)		1 SD		
β2-MG II	mg/L	BS-200E <sup>1</sup>	1.11	0.93	—	1.29	0.06	BS-200E <sup>1</sup>	2.98	2.53	—	3.43	0.15
		BS-240 <sup>2</sup>	1.07	0.92	—	1.22	0.05	BS-240 <sup>2</sup>	3.11	2.63	—	3.59	0.16
		BS-240E <sup>3</sup>	1.05	0.90	—	1.20	0.05	BS-240E <sup>3</sup>	2.98	2.53	—	3.43	0.15
		BS-360E <sup>4</sup>	1.06	0.91	—	1.21	0.05	BS-360E <sup>4</sup>	2.92	2.47	—	3.37	0.15
		BS-380 <sup>5</sup>	1.11	0.93	—	1.29	0.06	BS-380 <sup>5</sup>	2.98	2.53	—	3.43	0.15
		BS-400 <sup>6</sup>	1.11	0.93	—	1.29	0.06	BS-400 <sup>6</sup>	2.98	2.53	—	3.43	0.15
		BS-430 <sup>7</sup>	1.09	0.94	—	1.24	0.05	BS-430 <sup>7</sup>	2.97	2.52	—	3.42	0.15
		BS-480 <sup>8</sup>	1.09	0.94	—	1.24	0.05	BS-480 <sup>8</sup>	2.97	2.52	—	3.42	0.15
		BS-600 <sup>9</sup>	1.08	0.93	—	1.23	0.05	BS-600 <sup>9</sup>	2.96	2.51	—	3.41	0.15
		BS-600M <sup>10</sup>	1.04	0.89	—	1.19	0.05	BS-600M <sup>10</sup>	2.96	2.51	—	3.41	0.15
		BS-800 <sup>11</sup>	1.08	0.93	—	1.23	0.05	BS-800 <sup>11</sup>	3.00	2.55	—	3.45	0.15
		BS-2000 <sup>12</sup>	1.09	0.94	—	1.24	0.05	BS-2000 <sup>12</sup>	3.05	2.60	—	3.50	0.15
		BS-2800M <sup>13</sup>	1.04	0.89	—	1.19	0.05	BS-2800M <sup>13</sup>	2.96	2.51	—	3.41	0.15
BS-200E <sup>1</sup>	94.0	78.8	—	109.3	5.1	BS-200E <sup>1</sup>	252	214	—	291	13		
BS-240 <sup>2</sup>	90.6	77.9	—	103.3	4.2	BS-240 <sup>2</sup>	263	223	—	304	14		
BS-240E <sup>3</sup>	88.9	76.2	—	101.6	4.2	BS-240E <sup>3</sup>	252	214	—	291	13		
BS-360E <sup>4</sup>	89.8	77.1	—	102.5	4.2	BS-360E <sup>4</sup>	247	209	—	285	13		
BS-380 <sup>5</sup>	94.0	78.8	—	109.3	5.1	BS-380 <sup>5</sup>	252	214	—	291	13		

## β2-Microglobulin Control



nmol/L	<b>BS-400<sup>6</sup></b>	94.0	78.8	—	109.3	5.1	<b>BS-400<sup>6</sup></b>	252	214	—	291	13
	<b>BS-430<sup>7</sup></b>	92.3	79.6	—	105.0	4.2	<b>BS-430<sup>7</sup></b>	252	213	—	290	13
	<b>BS-480<sup>8</sup></b>	92.3	79.6	—	105.0	4.2	<b>BS-480<sup>8</sup></b>	252	213	—	290	13
	<b>BS-600<sup>9</sup></b>	91.5	78.8	—	104.2	4.2	<b>BS-600<sup>9</sup></b>	251	213	—	289	13
	<b>BS-600M<sup>10</sup></b>	88.1	75.4	—	100.8	4.2	<b>BS-600M<sup>10</sup></b>	251	213	—	289	13
	<b>BS-800<sup>11</sup></b>	91.5	78.8	—	104.2	4.2	<b>BS-800<sup>11</sup></b>	254	216	—	292	13
	<b>BS-2000<sup>12</sup></b>	92.3	79.6	—	105.0	4.2	<b>BS-2000<sup>12</sup></b>	258	220	—	296	13
	<b>BS-2800M<sup>13</sup></b>	88.1	75.4	—	100.8	4.2	<b>BS-2800M<sup>13</sup></b>	251	213	—	289	13

# mindray

## $\beta$ 2-Microglobulin Control

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

LOT 059124001

🕒 2025-04-30

