

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

1. **BS-120**: BS-120, BS-130;

2. **BS-200**: BS-200, BS-220;

3. **BS-230**: BS-230, BS-240;

4. **BS-240E**: BS240E, BS240Pro;

5. **BS-330**: BS-330, BS-350;

6. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),

BS-350E(Serial Number starts with "XS-");

7. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

8. **BS-380**: BS-380, BS-390;

9. **BS-400**: BS-400, BS-420;

10. **BS-430**: BS-430, BS-450, BS-460;

11. **BS-480**: BS-480, BS-490;

12. **BS-600**: BS-600, BS-620;

13. **BS-600M**: BS-600M, BS-620M;

14. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

15. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

16. **BS-2800M**: BS-2600M, BS-2800M.

LOT: 060823005

2024-08-28

| Abbreviated name | HCY Control (L) | | | | | | HCY Control (H) | | | | | | | |
|------------------|-------------------------------|-------------------|-------------|-------------------------------|---|------|-----------------|-------------------------------|-------------------|-------------------------------|------|------|------|-----|
| | Model | Unit | Assay Value | Range (Assay Value \pm 3SD) | | 1 SD | Model | Unit | Assay Value | Range (Assay Value \pm 3SD) | | 1 SD | | |
| HCY | BS-120 ¹ | $\mu\text{mol/L}$ | 16.0 | 12.8 | — | 19.2 | 1.1 | BS-120 ¹ | $\mu\text{mol/L}$ | 30.9 | 24.7 | — | 37.1 | 2.1 |
| | BS-200 ² | $\mu\text{mol/L}$ | 15.0 | 12.0 | — | 18.0 | 1.0 | BS-200 ² | $\mu\text{mol/L}$ | 30.6 | 24.5 | — | 36.7 | 2.0 |
| | BS-230 ³ | $\mu\text{mol/L}$ | 15.6 | 12.5 | — | 18.7 | 1.0 | BS-230 ³ | $\mu\text{mol/L}$ | 30.8 | 24.6 | — | 37.0 | 2.1 |
| | BS-240E ⁴ | $\mu\text{mol/L}$ | 15.7 | 12.6 | — | 18.8 | 1.0 | BS-240E ⁴ | $\mu\text{mol/L}$ | 31.0 | 24.8 | — | 37.2 | 2.1 |
| | BS-330 ⁵ | $\mu\text{mol/L}$ | 15.0 | 12.0 | — | 18.0 | 1.0 | BS-330 ⁵ | $\mu\text{mol/L}$ | 30.6 | 24.5 | — | 36.7 | 2.0 |
| | BS-330E ⁶ | $\mu\text{mol/L}$ | 15.2 | 12.2 | — | 18.2 | 1.0 | BS-330E ⁶ | $\mu\text{mol/L}$ | 30.1 | 24.1 | — | 36.1 | 2.0 |
| | BS-360E ⁷ | $\mu\text{mol/L}$ | 15.7 | 12.6 | — | 18.8 | 1.0 | BS-360E ⁷ | $\mu\text{mol/L}$ | 31.0 | 24.8 | — | 37.2 | 2.1 |
| | BS-380 ⁸ | $\mu\text{mol/L}$ | 15.5 | 12.4 | — | 18.6 | 1.0 | BS-380 ⁸ | $\mu\text{mol/L}$ | 30.5 | 24.4 | — | 36.6 | 2.0 |
| | BS-400 ⁹ | $\mu\text{mol/L}$ | 15.5 | 12.4 | — | 18.6 | 1.0 | BS-400 ⁹ | $\mu\text{mol/L}$ | 30.5 | 24.4 | — | 36.6 | 2.0 |
| | BS-430 ¹⁰ | $\mu\text{mol/L}$ | 15.7 | 12.6 | — | 18.8 | 1.0 | BS-430 ¹⁰ | $\mu\text{mol/L}$ | 31.0 | 24.8 | — | 37.2 | 2.1 |
| | BS-480 ¹¹ | $\mu\text{mol/L}$ | 15.4 | 12.3 | — | 18.5 | 1.0 | BS-480 ¹¹ | $\mu\text{mol/L}$ | 30.3 | 24.2 | — | 36.4 | 2.0 |
| | BS-600 ¹² | $\mu\text{mol/L}$ | 15.7 | 12.6 | — | 18.8 | 1.0 | BS-600 ¹² | $\mu\text{mol/L}$ | 31.0 | 24.8 | — | 37.2 | 2.1 |
| | BS-600M ¹³ | $\mu\text{mol/L}$ | 16.1 | 12.9 | — | 19.3 | 1.1 | BS-600M ¹³ | $\mu\text{mol/L}$ | 32.2 | 25.8 | — | 38.6 | 2.1 |
| | BS-800 ¹⁴ | $\mu\text{mol/L}$ | 15.7 | 12.6 | — | 18.8 | 1.0 | BS-800 ¹⁴ | $\mu\text{mol/L}$ | 31.0 | 24.8 | — | 37.2 | 2.1 |
| | BS-2000 ¹⁵ | $\mu\text{mol/L}$ | 15.9 | 12.7 | — | 19.1 | 1.1 | BS-2000 ¹⁵ | $\mu\text{mol/L}$ | 31.9 | 25.5 | — | 38.3 | 2.1 |
| | BS-2800M ¹⁶ | $\mu\text{mol/L}$ | 16.0 | 12.8 | — | 19.2 | 1.1 | BS-2800M ¹⁶ | $\mu\text{mol/L}$ | 32.0 | 25.6 | — | 38.4 | 2.1 |

| English | Abbreviated name | HCY Control | Model | Unit | Assay Value | Range (Assay Value \pm 3SD) |
|-----------|--------------------------|-----------------------|---------|---------|--------------------|--|
| Русский | сокращенное наименование | Контроль гомоцистеина | модель | Прибор | Результат анализа | Диапазон (результат анализа \pm 3CO) |
| Português | Nome abreviado | Controle HCY | Modelo | Unidade | Valores da análise | Faixa (Valores da análise \pm 3SD) |
| Español | nombre abreviado | Control HCY | modelo | Unidad | Valor de ensayo | Rango (Valor de ensayo \pm 3SD) |
| Italiano | abbreviazione | Controllo HCY | modelli | Unità | Valori di dosaggio | Intervallo (valore di concentrazione \pm 3 DS) |
| Türkçe | kısaltılmış ad | HCY Kontrolü | model | Ünite | Tayin Değeri | Aralık (Tayin Değeri \pm 3SD) |

HCY

| | |
|-----------|--------------------|
| English | Homocysteine (HCY) |
| Русский | Гомоцистеин (HCY) |
| Português | Homocisteína (HCY) |
| Español | homocisteína (HCY) |
| Italiano | Omocisteina (HCY) |
| Türkçe | Homosistain (HCY) |