

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

10.BS-400: BS-400, BS-420;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

11.BS-430: BS-430, BS-450,BS-460;

3.BS-200E: BS-200E, BS-220E,;

12.BS-480: BS-480, BS-490;

4.BS-240: BS-230, BS-240;

13.BS-600: BS-600, BS-620;

5.BS-240E: BS240E, BS240Pro;

14.BS-600M: BS-600M, BS-620M;

6.BS-300: BS-300, BS-320;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

17.BS-2800M:BS-2600M, BS-2800M;

330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

18.For applicable analyzers of the analyte, please subject to the parameter sheet and instrument.

9.BS-380: BS-380, BS-390;

LOT : 061124001

: 2025-08-24

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

Lp(a) II

English	Lp(a) II Calibrator
Русский	Калибратор Lp(a) II
Português	Calibrador Lp(a) II
Español	Calibrador Lp(a) II
Italiano	Calibratore Lp(a) II
Türkçe	Lp(a) II Kalibratörü

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Assay Value	Range (Assay Value±3SD)		1 SD		
Lp(a) II	nmol/L	BS-120 ¹	40.2	26.7	—	53.7	4.5	BS-120 ¹	120	81	—	159	13
		BS-200 ²	39.8	26.6	—	53.0	4.4	BS-200 ²	121	82	—	160	13
		BS-200E ³	41.7	27.9	—	55.5	4.6	BS-200E ³	123	81	—	165	14
		BS-240 ⁴	41.4	27.6	—	55.2	4.6	BS-240 ⁴	119	80	—	158	13
		BS-240E ⁵	41.9	27.8	—	56.0	4.7	BS-240E ⁵	116	77	—	155	13
		BS-300 ⁶	45.8	30.5	—	61.1	5.1	BS-300 ⁶	126	84	—	168	14
		BS-330E ⁷	41.7	27.9	—	55.5	4.6	BS-330E ⁷	123	81	—	165	14
		BS-360E ⁸	40.1	26.6	—	53.6	4.5	BS-360E ⁸	113	74	—	152	13
		BS-380 ⁹	41.0	27.2	—	54.8	4.6	BS-380 ⁹	120	81	—	159	13
		BS-400 ¹⁰	41.1	27.3	—	54.9	4.6	BS-400 ¹⁰	124	82	—	166	14
		BS-430 ¹¹	41.9	27.8	—	56.0	4.7	BS-430 ¹¹	122	80	—	164	14
		BS-480 ¹²	41.9	27.8	—	56.0	4.7	BS-480 ¹²	122	80	—	164	14
		BS-600 ¹³	41.9	27.8	—	56.0	4.7	BS-600 ¹³	122	80	—	164	14
		BS-600M ¹⁴	42.3	28.2	—	56.4	4.7	BS-600M ¹⁴	124	82	—	166	14
		BS-800 ¹⁵	41.9	27.8	—	56.0	4.7	BS-800 ¹⁵	122	80	—	164	14

Lipoprotein (a) Control

mindray

BS-2000¹⁶	41.4	27.6	—	55.2	4.6	BS-2000¹⁶	122	80	—	164	14
BS-2800M¹⁷	42.1	28.0	—	56.2	4.7	BS-2800M¹⁷	123	81	—	165	14

mindray

Lipoprotein (a) Control

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 061124001

🕒 2025-08-24

