

# CO2 and TBA Multi Control

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- |   |   |
|---|---|
| <p>1. <b>BS-120</b>: BS-120, BS-130;</p> <p>2. <b>BS-180</b>: BS-180, BS-190;</p> <p>3. <b>BS-200</b>: BS-200, BS-220;</p> <p>4. <b>BS-200E</b>: BS-200E, BS-220E;</p> <p>5. <b>BS-230</b>: BS-230, BS-240, BS-280;</p> <p>6. <b>BS-240E</b>: BS-240E, BS-240Pro;</p> <p>7. <b>BS-300</b>: BS-300, BS-320;</p> <p>8. <b>BS-330</b>: BS-330, BS-350;</p> <p>9. <b>BS-330E</b>: BS-330E (Serial Number starts with "XQ-"),<br/>BS-350E (Serial Number starts with "XS-");</p> | <p>10. <b>BS-360E</b>: BS-360E, BS-370E, BS-350S, BS-360S,<br/><b>BS-330E(V35.00)</b> (Serial Number starts with "W8-"<br/>and software version starts with "35.00"),<br/><b>BS-350E(V35.00)</b> (Serial Number starts with "W9-"<br/>and software version starts with "35.00");</p> <p>11. <b>BS-380</b>: BS-380, BS-390;</p> <p>12. <b>BS-400</b>: BS-400, BS-420;</p> <p>13. <b>BS-430</b>: BS-410, BS-430, BS-450, BS-460, BS-470;</p> <p>14. <b>BS-480</b>: BS-480, BS-490;</p> <p>15. <b>BS-600</b>: BS-600, BS-620;</p> <p>16. <b>BS-600M</b>: BS-600M, BS-620M;</p> <p>17. <b>BS-800</b>: BS-800, BS-820, BS-800M, BS-820M,<br/>BS-1800, BS-1800plus;</p> <p>18. <b>BS-830</b>: BS-830, BS-830S, BS-840, BS-850, BS-860,<br/>BS-870;</p> <p>19. <b>BS-2000</b>: BS-2000, BS-2200, BS-2000M, BS-2200M.</p> <p>20. <b>BS-2800M</b>: BS-2600M, BS-2800M.</p> |
|---|---|

**LOT**: 061822002

: 2023-04-28

Abbreviate d name	CO2 Control(L)						CO2 Control(H)				
	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD	
CO2	<b>BS-120</b>	<sup>1</sup> mmol/L	/	/ - /	/	<b>BS-120</b>	<sup>1</sup> mmol/L	/	/ - /	/	
	<b>BS-180</b>	<sup>2</sup> mmol/L	/	/ - /	/	<b>BS-180</b>	<sup>2</sup> mmol/L	/	/ - /	/	
	<b>BS-200</b>	<sup>3</sup> mmol/L	/	/ - /	/	<b>BS-200</b>	<sup>3</sup> mmol/L	/	/ - /	/	
	<b>BS-200E</b>	<sup>4</sup> mmol/L	/	/ - /	/	<b>BS-200E</b>	<sup>4</sup> mmol/L	/	/ - /	/	
	<b>BS-230</b>	<sup>5</sup> mmol/L	/	/ - /	/	<b>BS-230</b>	<sup>5</sup> mmol/L	/	/ - /	/	
	<b>BS-240E</b>	<sup>6</sup> mmol/L	/	/ - /	/	<b>BS-240E</b>	<sup>6</sup> mmol/L	/	/ - /	/	
	<b>BS-300</b>	<sup>7</sup> mmol/L	/	/ - /	/	<b>BS-300</b>	<sup>7</sup> mmol/L	/	/ - /	/	
	<b>BS-330</b>	<sup>8</sup> mmol/L	/	/ - /	/	<b>BS-330</b>	<sup>8</sup> mmol/L	/	/ - /	/	
	<b>BS-330E</b>	<sup>9</sup> mmol/L	/	/ - /	/	<b>BS-330E</b>	<sup>9</sup> mmol/L	/	/ - /	/	
	<b>BS-360E</b>	<sup>10</sup> mmol/L	/	/ - /	/	<b>BS-360E</b>	<sup>10</sup> mmol/L	/	/ - /	/	
	<b>BS-380</b>	<sup>11</sup> mmol/L	/	/ - /	/	<b>BS-380</b>	<sup>11</sup> mmol/L	/	/ - /	/	
	<b>BS-400</b>	<sup>12</sup> mmol/L	/	/ - /	/	<b>BS-400</b>	<sup>12</sup> mmol/L	/	/ - /	/	
	<b>BS-430</b>	<sup>13</sup> mmol/L	/	/ - /	/	<b>BS-430</b>	<sup>13</sup> mmol/L	/	/ - /	/	
	<b>BS-480</b>	<sup>14</sup> mmol/L	/	/ - /	/	<b>BS-480</b>	<sup>14</sup> mmol/L	/	/ - /	/	
	<b>BS-600</b>	<sup>15</sup> mmol/L	/	/ - /	/	<b>BS-600</b>	<sup>15</sup> mmol/L	/	/ - /	/	
	<b>BS-600M</b>	<sup>16</sup> mmol/L	15.4	12.3 - 18.5	1.0	<b>BS-600M</b>	<sup>16</sup> mmol/L	15.7	12.5 - 18.9	1.1	
	<b>BS-800</b>	<sup>17</sup> mmol/L	/	/ - /	/	<b>BS-800</b>	<sup>17</sup> mmol/L	/	/ - /	/	
	<b>BS-830</b>	<sup>18</sup> mmol/L	/	/ - /	/	<b>BS-830</b>	<sup>18</sup> mmol/L	/	/ - /	/	
	<b>BS-2000</b>	<sup>19</sup> mmol/L	/	/ - /	/	<b>BS-2000</b>	<sup>19</sup> mmol/L	/	/ - /	/	
	<b>BS-2800M</b>	<sup>20</sup> mmol/L	/	/ - /	/	<b>BS-2800M</b>	<sup>20</sup> mmol/L	/	/ - /	/	

# CO2 and TBA Multi Control



TBA Control(L)							TBA Control(H)									
Abbreviated name	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD				
TBA	BS-120	<sup>1</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-120	<sup>1</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-180	<sup>2</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-180	<sup>2</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-200	<sup>3</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-200	<sup>3</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-200E	<sup>4</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-200E	<sup>4</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-230	<sup>5</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-230	<sup>5</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-240E	<sup>6</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-240E	<sup>6</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-300	<sup>7</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-300	<sup>7</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-330	<sup>8</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-330	<sup>8</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-330E	<sup>9</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-330E	<sup>9</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-360E	<sup>10</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-360E	<sup>10</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-380	<sup>11</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-380	<sup>11</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-400	<sup>12</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-400	<sup>12</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-430	<sup>13</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-430	<sup>13</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-480	<sup>14</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-480	<sup>14</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-600	<sup>15</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-600	<sup>15</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-600M	<sup>16</sup>	$\mu\text{mol/L}$	21.1	16.9	-	25.3	1.4	BS-600M	<sup>16</sup>	$\mu\text{mol/L}$	39.1	31.2	-	47.0	2.6
	BS-800	<sup>17</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-800	<sup>17</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-830	<sup>18</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-830	<sup>18</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-2000	<sup>19</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-2000	<sup>19</sup>	$\mu\text{mol/L}$	/	/	-	/	/
	BS-2800M	<sup>20</sup>	$\mu\text{mol/L}$	/	/	-	/	/	BS-2800M	<sup>20</sup>	$\mu\text{mol/L}$	/	/	-	/	/

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа $\pm$ 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise $\pm$ 3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo $\pm$ 3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione $\pm$ 3 DS)
Türkçe	kisaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri $\pm$ 3SD)

CO2		TBA	
English	Carbon Dioxide	Total Bile Acids	
Русский	Двуокись углерода	общие желчные кислоты	
Português	Dióxido de Carbono	ácidos biliareos totales	
Español	Dióxido de carbono	ácidos biliareos totales	
Italiano	Anidride carbonica	acidi biliari totali	
Türkçe	Karbon Dioksit	total safra asitleridir	