

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

10.BS-400: BS-400, BS-420;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

11.BS-430: BS-430, BS-450,BS-460;

3.BS-200E: BS-200E, BS-220E,;

12.BS-480: BS-480, BS-490;

4.BS-240: BS-230, BS-240;

13.BS-600: BS-600, BS-620;

5.BS-240E: BS240E, BS240Pro;

14.BS-600M: BS-600M, BS-620M;

6.BS-300: BS-300, BS-320;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

17.BS-2800M:BS-2600M, BS-2800M;

9.BS-380: BS-380, BS-390;

18.For applicable analyzers of the analyte, please subject to the parameter sheet and instrument.

LOT : 061824003

: 2025-09-04

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

CO2			TBA		
English	Carbon Dioxide		Total Bile Acids		
Русский	Двуокись углерода		желчные кислоты		
Português	Dióxido de Carbono		Ácido Biliar Total		
Español	Dióxido de carbono		ácidos biliares totales		
Italiano	Anidride carbonica		acidi biliari totali		
Türkçe	Karbon Dioksit		Total Safra Asitleridir		

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Assay Value	Range (Assay Value±3SD)		1 SD		
CO2	mmol/L	BS-200 <sup>2</sup>	12.8	10.1	—	15.5	0.9	BS-200 <sup>2</sup>	17.2	13.6	—	20.8	1.2
		BS-200E <sup>3</sup>	13.6	10.9	—	16.3	0.9	BS-200E <sup>3</sup>	18.0	14.4	—	21.6	1.2
		BS-240 <sup>4</sup>	13.0	10.3	—	15.7	0.9	BS-240 <sup>4</sup>	17.4	13.8	—	21.0	1.2
		BS-240E <sup>5</sup>	13.5	10.8	—	16.2	0.9	BS-240E <sup>5</sup>	17.9	14.3	—	21.5	1.2
		BS-300 <sup>6</sup>	13.1	10.4	—	15.8	0.9	BS-300 <sup>6</sup>	17.6	14.0	—	21.2	1.2
		BS-330E <sup>7</sup>	13.6	10.9	—	16.3	0.9	BS-330E <sup>7</sup>	18.0	14.4	—	21.6	1.2
		BS-360E <sup>8</sup>	13.5	10.8	—	16.2	0.9	BS-360E <sup>8</sup>	17.9	14.3	—	21.5	1.2
		BS-380 <sup>9</sup>	13.6	10.9	—	16.3	0.9	BS-380 <sup>9</sup>	18.0	14.4	—	21.6	1.2
		BS-400 <sup>10</sup>	13.6	10.9	—	16.3	0.9	BS-400 <sup>10</sup>	18.1	14.5	—	21.7	1.2
		BS-430 <sup>11</sup>	13.5	10.8	—	16.2	0.9	BS-430 <sup>11</sup>	17.9	14.3	—	21.5	1.2
		BS-480 <sup>12</sup>	13.5	10.8	—	16.2	0.9	BS-480 <sup>12</sup>	17.9	14.3	—	21.5	1.2
		BS-600 <sup>13</sup>	13.5	10.8	—	16.2	0.9	BS-600 <sup>13</sup>	17.9	14.3	—	21.5	1.2
		BS-600M <sup>14</sup>	14.0	11.3	—	16.7	0.9	BS-600M <sup>14</sup>	18.6	15.0	—	22.2	1.2
		BS-800 <sup>15</sup>	13.5	10.8	—	16.2	0.9	BS-800 <sup>15</sup>	17.9	14.3	—	21.5	1.2
		BS-2000 <sup>16</sup>	13.3	10.6	—	16.0	0.9	BS-2000 <sup>16</sup>	17.8	14.2	—	21.4	1.2

**BS-2800M<sup>17</sup>** 13.8 11.1 — 16.5 0.9 **BS-2800M<sup>17</sup>** 18.4 14.8 — 22.0 1.2

Abbreviated name	Unit	Level: L/N				Level: H/P			
		Model	Assay Value	Range (Assay Value±3SD)	1 SD	Model	Assay Value	Range (Assay Value±3SD)	1 SD
TBA	μmol/L	<b>BS-120<sup>1</sup></b>	20.8	16.6 — 25.0	1.4	<b>BS-120<sup>1</sup></b>	40.5	32.4 — 48.6	2.7
		<b>BS-200<sup>2</sup></b>	20.7	16.5 — 24.9	1.4	<b>BS-200<sup>2</sup></b>	40.6	32.5 — 48.7	2.7
		<b>BS-200E<sup>3</sup></b>	21.3	17.1 — 25.5	1.4	<b>BS-200E<sup>3</sup></b>	41.5	33.1 — 49.9	2.8
		<b>BS-240<sup>4</sup></b>	21.1	16.9 — 25.3	1.4	<b>BS-240<sup>4</sup></b>	40.7	32.6 — 48.8	2.7
		<b>BS-240E<sup>5</sup></b>	21.2	17.0 — 25.4	1.4	<b>BS-240E<sup>5</sup></b>	41.8	33.4 — 50.2	2.8
		<b>BS-300<sup>6</sup></b>	21.8	17.3 — 26.3	1.5	<b>BS-300<sup>6</sup></b>	42.4	34.0 — 50.8	2.8
		<b>BS-330E<sup>7</sup></b>	21.3	17.1 — 25.5	1.4	<b>BS-330E<sup>7</sup></b>	41.5	33.1 — 49.9	2.8
		<b>BS-360E<sup>8</sup></b>	21.0	16.8 — 25.2	1.4	<b>BS-360E<sup>8</sup></b>	41.0	32.9 — 49.1	2.7
		<b>BS-380<sup>9</sup></b>	21.7	17.2 — 26.2	1.5	<b>BS-380<sup>9</sup></b>	41.8	33.4 — 50.2	2.8
		<b>BS-400<sup>10</sup></b>	21.4	17.2 — 25.6	1.4	<b>BS-400<sup>10</sup></b>	41.7	33.3 — 50.1	2.8
		<b>BS-430<sup>11</sup></b>	21.4	17.2 — 25.6	1.4	<b>BS-430<sup>11</sup></b>	41.4	33.0 — 49.8	2.8
		<b>BS-480<sup>12</sup></b>	21.5	17.3 — 25.7	1.4	<b>BS-480<sup>12</sup></b>	41.6	33.2 — 50.0	2.8
		<b>BS-600<sup>13</sup></b>	21.2	17.0 — 25.4	1.4	<b>BS-600<sup>13</sup></b>	41.3	32.9 — 49.7	2.8
		<b>BS-600M<sup>14</sup></b>	21.5	17.3 — 25.7	1.4	<b>BS-600M<sup>14</sup></b>	41.7	33.3 — 50.1	2.8
		<b>BS-800<sup>15</sup></b>	21.8	17.3 — 26.3	1.5	<b>BS-800<sup>15</sup></b>	42.3	33.9 — 50.7	2.8
		<b>BS-2000<sup>16</sup></b>	21.9	17.4 — 26.4	1.5	<b>BS-2000<sup>16</sup></b>	42.8	34.1 — 51.5	2.9
		<b>BS-2800M<sup>17</sup></b>	21.5	17.3 — 25.7	1.4	<b>BS-2800M<sup>17</sup></b>	41.7	33.3 — 50.1	2.8

# mindray

## CO2 and TBA Multi Control

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

LOT 061824003

🕒 2025-09-04

