

## Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray  
 Português : Planilha de valores do Sistema de Medição BS da Mindray  
 Español : Hoja de valores del sistema de medición Mindray BS  
 Italiano : Scheda dei valori del sistema di misurazione BS di Mindray  
 Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



**The data of each group is same.**

**Português : A dados de cada grupo é a mesma.**

**Italiano : la dati di ogni gruppo è la stessa.**

- 1.BS-120: BS-120, BS-130, BS-180, BS-190;
- 2.BS-200: BS-200, BS-220, BS-330, BS-350;
- 3.BS-200E: BS-200E, BS-220E,;
- 4.BS-240: BS-230, BS-240;
- 5.BS-240E: BS240E, BS240Pro;
- 6.BS-300: BS-300, BS-320;
- 7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")
- 8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");
- 9.BS-380: BS-380, BS-390;

**Русский: Данные совпадают во всех группах.**

**Español : la datos de cada grupo es la misma.**

**Türkçe : her grubun veri aynıdır.**

- 10.BS-400: BS-400, BS-420;
- 11.BS-430: BS-430, BS-450,BS-460;
- 12.BS-480: BS-480, BS-490;
- 13.BS-600: BS-600, BS-620;
- 14.BS-600M: BS-600M, BS-620M;
- 15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
- 16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
- 17.BS-2800M:BS-2600M, BS-2800M;
- 18.Applicable models of the chemistry shall be subject to the parameter list and instructions.

**LOT : 145323010**

**🕒 : 2025-04-16**

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
<b>Русский</b>	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
<b>Português</b>	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
<b>Español</b>	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
<b>Italiano</b>	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
<b>Türkçe</b>	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)
<b>Fe</b>					
<b>English</b>			Iron		
<b>Русский</b>			Железо		
<b>Português</b>			Ferro		
<b>Español</b>			Hierro		
<b>Italiano</b>			Ferro		
<b>Türkçe</b>			Demir		

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Assay Value	Range(Assay Value±3SD)	1 SD
Fe	µmol/L	BS-120 <sup>1</sup>	19.3	15.4 — 23.2	1.3	BS-400 <sup>10</sup>	19.9	16.0 — 23.8	1.3
		BS-200 <sup>2</sup>	19.9	16.0 — 23.8	1.3	BS-430 <sup>11</sup>	19.6	15.7 — 23.5	1.3
		BS-200E <sup>3</sup>	19.9	16.0 — 23.8	1.3	BS-480 <sup>12</sup>	19.7	15.8 — 23.6	1.3
		BS-240 <sup>4</sup>	19.2	15.3 — 23.1	1.3	BS-600 <sup>13</sup>	19.6	15.7 — 23.5	1.3
		BS-240E <sup>5</sup>	20.1	16.2 — 24.0	1.3	BS-600M <sup>14</sup>	19.5	15.6 — 23.4	1.3
	mg/L	BS-300 <sup>6</sup>	19.5	15.6 — 23.4	1.3	BS-800 <sup>15</sup>	19.0	15.1 — 22.9	1.3
		BS-330E <sup>7</sup>	19.9	16.0 — 23.8	1.3	BS-2000 <sup>16</sup>	19.6	15.7 — 23.5	1.3
		BS-360E <sup>8</sup>	19.8	15.9 — 23.7	1.3	BS-2800M <sup>17</sup>	19.5	15.6 — 23.4	1.3
		BS-380 <sup>9</sup>	20.1	16.2 — 24.0	1.3				
		BS-120 <sup>1</sup>	1.08	0.86 — 1.30	0.07	BS-400 <sup>10</sup>	1.11	0.89 — 1.33	0.07
BS-200 <sup>2</sup>	1.11	0.89 — 1.33	0.07	BS-430 <sup>11</sup>	1.09	0.88 — 1.31	0.07		
BS-200E <sup>3</sup>	1.11	0.89 — 1.33	0.07	BS-480 <sup>12</sup>	1.10	0.88 — 1.32	0.07		
BS-240 <sup>4</sup>	1.07	0.85 — 1.29	0.07	BS-600 <sup>13</sup>	1.09	0.88 — 1.31	0.07		
BS-240E <sup>5</sup>	1.12	0.91 — 1.34	0.07	BS-600M <sup>14</sup>	1.09	0.87 — 1.31	0.07		

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value $\pm$ 3SD)			1 SD	Model	Assay Value	Range(Assay Value $\pm$ 3SD)			1 SD
Fe	mg/L	BS-300 <sup>6</sup>	1.09	0.87	—	1.31	0.07	BS-800 <sup>15</sup>	1.06	0.84	—	1.28	0.07
		BS-330E <sup>7</sup>	1.11	0.89	—	1.33	0.07	BS-2000 <sup>16</sup>	1.09	0.88	—	1.31	0.07
		BS-360E <sup>8</sup>	1.11	0.89	—	1.32	0.07	BS-2800M <sup>17</sup>	1.09	0.87	—	1.31	0.07
		BS-380 <sup>9</sup>	1.12	0.91	—	1.34	0.07						