

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray  
 Português: Planilha de valores do Sistema de Medição BS da Mindray  
 Español: Hoja de valores del sistema de medición Mindray BS  
 Italiano: Scheda dei valori del sistema di misurazione BS di Mindray  
 Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



**The data of each group is same.**

**Português: A dados de cada grupo é a mesma.**

**Italiano: la dati di ogni gruppo è la stessa.**

- 1.BS-120: BS-120, BS-130, BS-180, BS-190;
- 2.BS-200: BS-200, BS-220, BS-330, BS-350;
- 3.BS-200E: BS-200E, BS-220E,;
- 4.BS-240: BS-230, BS-240;
- 5.BS-240E: BS240E, BS240Pro;
- 6.BS-300: BS-300, BS-320;
- 7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")
- 8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");
- 9.BS-380: BS-380, BS-390;

**Русский: Данные совпадают во всех группах.**

**Español: la datos de cada grupo es la misma.**

**Türkçe: her grubun veri aynıdır.**

- 10.BS-400: BS-400, BS-420;
- 11.BS-430: BS-430, BS-450,BS-460;
- 12.BS-480: BS-480, BS-490;
- 13.BS-600: BS-600, BS-620;
- 14.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
- 15.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
- 16.BS-2800M:BS-2600M, BS-2800M;
- 17.Applicable models of the chemistry shall be subject to the parameter list and instructions.

**LOT: 145423010**

**🕒: 2025-03-02**

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
<b>Русский</b>	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
<b>Português</b>	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
<b>Español</b>	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
<b>Italiano</b>	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
<b>Türkçe</b>	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

**FUN**

<b>English</b>	Fructosamine
<b>Русский</b>	Фруктозамин
<b>Português</b>	Frutosamina
<b>Español</b>	Fructosamina
<b>Italiano</b>	Fruttosamina
<b>Türkçe</b>	Fruktozamin

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Assay Value	Range(Assay Value±3SD)	1 SD
<b>FUN</b>	µmol/L	<b>BS-120<sup>1</sup></b>	459	390 — 528	23	<b>BS-380<sup>9</sup></b>	444	378 — 510	22
		<b>BS-200<sup>2</sup></b>	455	386 — 524	23	<b>BS-400<sup>10</sup></b>	457	388 — 526	23
		<b>BS-200E<sup>3</sup></b>	478	406 — 550	24	<b>BS-430<sup>11</sup></b>	471	399 — 543	24
		<b>BS-240<sup>4</sup></b>	455	386 — 524	23	<b>BS-480<sup>12</sup></b>	467	398 — 536	23
		<b>BS-240E<sup>5</sup></b>	464	395 — 533	23	<b>BS-600<sup>13</sup></b>	464	395 — 533	23
		<b>BS-300<sup>6</sup></b>	467	398 — 536	23	<b>BS-800<sup>14</sup></b>	467	398 — 536	23
		<b>BS-330E<sup>7</sup></b>	478	406 — 550	24	<b>BS-2000<sup>15</sup></b>	472	400 — 544	24
		<b>BS-360E<sup>8</sup></b>	471	399 — 543	24	<b>BS-2800M<sup>16</sup></b>	456	387 — 525	23

**mindray**  
**FUN Control P**

For use on: BS-2000, BS-2800M,  
BS-800

LOT 145423010

2025-03-02

