

Coagulation Control

Reference Value Table

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation/Analyte	Unit	Reference Value	Scope
Coagulation control (level 2)		CX-9000, CX-9010, CX-9200 and CX-9210 Auto Coagulation Analyzers of Shenzhen Mindray Bio-medical Electronics Co., Ltd	PT	s	30.48	24.38 ~ 36.58
	8723060118		APTT	s	47.06	37.65 ~ 56.47
			TT	s	/	/
	2024-06-13		FIB	g/L	/	/
				mg/dL	/	/

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation/Analyte	Unit	Reference Value	Scope
Coagulation control (level 2)		C3510 and ExC810 Auto Coagulation Analyzers of Beijing Mindray Medical Instrument Co., Ltd	PT	s	27.12	21.70 ~ 32.54
	8723060118		APTT	s	51.28	41.02 ~ 61.54
			TT	s	/	/
	2024-06-13		FIB	g/L	/	/
				mg/dL	/	/