

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

## ASSAY VALUES AND EXPECTED RANGES

**LOT**  
**BC2405B**  
2024-07-10

<b>Instrument</b>	<b>Parameter</b>	<b>Low</b>		<b>Normal</b>		<b>High</b>		<b>+</b>
		<b>LOT</b>	<b>BC2405BL</b>	<b>LOT</b>	<b>BC2405BN</b>	<b>LOT</b>	<b>BC2405BH</b>	
<b>BC-5800, BC-5600</b>	WBC $\times 10^9/L$	3.41	$\pm$ 0.50	7.95	$\pm$ 1.00	17.96	$\pm$ 2.50	
<b>QC Mode</b>	Neu# $\times 10^9/L$	1.65	$\pm$ 0.31	4.40	$\pm$ 0.72	11.40	$\pm$ 1.62	
	Lym# $\times 10^9/L$	1.34	$\pm$ 0.31	2.39	$\pm$ 0.64	3.74	$\pm$ 1.44	
	Mon# $\times 10^9/L$	0.24	$\pm$ 0.18	0.56	$\pm$ 0.40	1.15	$\pm$ 0.90	
	Eos# $\times 10^9/L$	0.15	$\pm$ 0.11	0.52	$\pm$ 0.40	1.49	$\pm$ 1.08	
	Bas# $\times 10^9/L$	0.03	$\pm$ 0.03	0.08	$\pm$ 0.08	0.18	$\pm$ 0.18	
	Neu%	48.5	$\pm$ 9.0	55.4	$\pm$ 9.0	63.5	$\pm$ 9.0	
	Lym%	39.2	$\pm$ 9.0	30.0	$\pm$ 8.0	20.8	$\pm$ 8.0	
	Mon%	6.9	$\pm$ 5.0	7.1	$\pm$ 5.0	6.4	$\pm$ 5.0	
	Eos%	4.4	$\pm$ 3.0	6.5	$\pm$ 5.0	8.3	$\pm$ 6.0	
	Bas%	1.0	$\pm$ 1.0	1.0	$\pm$ 1.0	1.0	$\pm$ 1.0	
	RBC $\times 10^{12}/L$	2.17	$\pm$ 0.18	4.05	$\pm$ 0.24	4.96	$\pm$ 0.30	
	HGB g/L	59	$\pm$ 4	129	$\pm$ 6	167	$\pm$ 8	
	HCT %	18.3	$\pm$ 1.5	39.7	$\pm$ 2.0	52.4	$\pm$ 2.4	
	MCV fL	84.2	$\pm$ 5.0	98.0	$\pm$ 5.0	105.7	$\pm$ 5.0	
	MCH pg	27.2	$\pm$ 2.5	31.9	$\pm$ 2.5	33.7	$\pm$ 2.5	
	MCHC g/L	323	$\pm$ 30	325	$\pm$ 30	319	$\pm$ 30	
	RDW-CV %	15.6	$\pm$ 3.0	13.5	$\pm$ 3.0	13.5	$\pm$ 3.0	
	RDW-SD fL	47.6	$\pm$ 10.0	49.5	$\pm$ 10.0	53.1	$\pm$ 10.0	
	PLT $\times 10^9/L$	53	$\pm$ 20	255	$\pm$ 40	500	$\pm$ 60	
	MPV fL	8.1	$\pm$ 3.0	9.3	$\pm$ 3.0	9.7	$\pm$ 3.0	
	PCT %*	0.043	$\pm$ 0.043	0.237	$\pm$ 0.100	0.485	$\pm$ 0.200	
	PDW*	15.9	$\pm$ 3.0	16.1	$\pm$ 3.0	16.3	$\pm$ 3.0	
	P-LCC $\times 10^9/L$	10	$\pm$ 10	69	$\pm$ 25	155	$\pm$ 35	
	P-LCR %	19.1	$\pm$ 10.0	27.0	$\pm$ 10.0	31.0	$\pm$ 10.0	
<b>BC-5390</b>	WBC $\times 10^9/L$	3.25	$\pm$ 0.50	7.75	$\pm$ 1.00	17.15	$\pm$ 2.50	
<b>QC Mode</b>	Neu# $\times 10^9/L$	1.66	$\pm$ 0.30	4.42	$\pm$ 0.70	11.23	$\pm$ 1.55	
	Lym# $\times 10^9/L$	1.24	$\pm$ 0.30	2.17	$\pm$ 0.70	3.34	$\pm$ 1.37	
	Mon# $\times 10^9/L$	0.20	$\pm$ 0.17	0.50	$\pm$ 0.39	0.86	$\pm$ 0.69	
	Eos# $\times 10^9/L$	0.16	$\pm$ 0.13	0.66	$\pm$ 0.47	1.72	$\pm$ 1.38	
	Bas# $\times 10^9/L$	0.80	$\pm$ 0.33	2.13	$\pm$ 0.78	5.23	$\pm$ 1.72	
	Neu%	51.0	$\pm$ 9.0	57.0	$\pm$ 9.0	65.5	$\pm$ 9.0	
	Lym%	38.0	$\pm$ 9.0	28.0	$\pm$ 9.0	19.5	$\pm$ 8.0	
	Mon%	6.0	$\pm$ 5.0	6.5	$\pm$ 5.0	5.0	$\pm$ 4.0	
	Eos%	5.0	$\pm$ 4.0	8.5	$\pm$ 6.0	10.0	$\pm$ 8.0	
	Bas%	24.5	$\pm$ 10.0	27.5	$\pm$ 10.0	30.5	$\pm$ 10.0	
	RBC $\times 10^{12}/L$	2.11	$\pm$ 0.18	3.98	$\pm$ 0.24	4.92	$\pm$ 0.30	
	HGB g/L	54	$\pm$ 4	120	$\pm$ 6	156	$\pm$ 8	
	HCT %	17.3	$\pm$ 1.5	38.2	$\pm$ 2.0	50.4	$\pm$ 2.4	
	MCV fL	82.0	$\pm$ 5.0	96.0	$\pm$ 5.0	102.5	$\pm$ 5.0	
	MCH pg	25.6	$\pm$ 2.5	30.2	$\pm$ 2.5	31.7	$\pm$ 2.5	
	MCHC g/L	312	$\pm$ 30	314	$\pm$ 30	309	$\pm$ 30	
	RDW-CV %	15.5	$\pm$ 3.0	13.5	$\pm$ 3.0	13.5	$\pm$ 3.0	
	RDW-SD fL	49.5	$\pm$ 8.0	50.0	$\pm$ 8.0	52.5	$\pm$ 8.0	
	PLT $\times 10^9/L$	51	$\pm$ 20	252	$\pm$ 40	492	$\pm$ 60	
	MPV fL	11.2	$\pm$ 3.0	12.2	$\pm$ 3.0	12.4	$\pm$ 3.0	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

## ASSAY VALUES AND EXPECTED RANGES

**LOT**  


**BC2405B**  
2024-07-10

<b>Instrument</b>	<b>Parameter</b>	<b>Low</b>		<b>Normal</b>		<b>High</b>		<b>++</b>
		<b>LOT</b>	<b>BC2405BL</b>	<b>LOT</b>	<b>BC2405BN</b>	<b>LOT</b>	<b>BC2405BH</b>	
<b>BC-5390 CRP</b>	$\text{WBC} \times 10^9/\text{L}$	3.23	$\pm$ 0.50	7.76	$\pm$ 1.00	17.39	$\pm$ 2.50	
<b>BC-5310 CRP</b>	$\text{Neu\#} \times 10^9/\text{L}$	1.65	$\pm$ 0.30	4.44	$\pm$ 0.70	11.37	$\pm$ 1.57	
<b>QC Mode</b>	$\text{Lym\#} \times 10^9/\text{L}$	1.25	$\pm$ 0.30	2.26	$\pm$ 0.63	3.43	$\pm$ 1.40	
	$\text{Mon\#} \times 10^9/\text{L}$	0.17	$\pm$ 0.13	0.40	$\pm$ 0.31	0.89	$\pm$ 0.70	
	$\text{Eos\#} \times 10^9/\text{L}$	0.16	$\pm$ 0.13	0.66	$\pm$ 0.47	1.70	$\pm$ 1.39	
	$\text{Bas\#} \times 10^9/\text{L}$	0.79	$\pm$ 0.33	2.13	$\pm$ 0.78	5.29	$\pm$ 1.74	
	$\text{Neu\%}$	50.9	$\pm$ 9.0	57.2	$\pm$ 9.0	65.4	$\pm$ 9.0	
	$\text{Lym\%}$	38.7	$\pm$ 9.0	29.1	$\pm$ 8.0	19.7	$\pm$ 8.0	
	$\text{Mon\%}$	5.3	$\pm$ 4.0	5.2	$\pm$ 4.0	5.1	$\pm$ 4.0	
	$\text{Eos\%}$	5.1	$\pm$ 4.0	8.5	$\pm$ 6.0	9.8	$\pm$ 8.0	
	$\text{Bas\%}$	24.5	$\pm$ 10.0	27.5	$\pm$ 10.0	30.4	$\pm$ 10.0	
	$\text{RBC} \times 10^{12}/\text{L}$	2.11	$\pm$ 0.18	4.01	$\pm$ 0.24	4.91	$\pm$ 0.30	
	$\text{HGB g/L}$	55	$\pm$ 4	120	$\pm$ 6	157	$\pm$ 8	
	$\text{HCT \%}$	17.3	$\pm$ 1.5	38.8	$\pm$ 2.0	51.5	$\pm$ 2.4	
	$\text{MCV fL}$	82.2	$\pm$ 5.0	96.8	$\pm$ 5.0	104.8	$\pm$ 5.0	
	$\text{MCH pg}$	26.1	$\pm$ 2.5	29.9	$\pm$ 2.5	32.0	$\pm$ 2.5	
	$\text{MCHC g/L}$	317	$\pm$ 30	309	$\pm$ 30	305	$\pm$ 30	
	$\text{RDW-CV \%}$	16.1	$\pm$ 3.0	14.1	$\pm$ 3.0	13.9	$\pm$ 3.0	
	$\text{RDW-SD fL}$	47.4	$\pm$ 8.0	48.6	$\pm$ 8.0	51.9	$\pm$ 8.0	
	$\text{PLT} \times 10^9/\text{L}$	46	$\pm$ 20	246	$\pm$ 40	488	$\pm$ 60	
	$\text{MPV fL}$	9.3	$\pm$ 3.0	10.1	$\pm$ 3.0	10.3	$\pm$ 3.0	
	$\text{PCT \%*}$	0.043	$\pm$ 0.043	0.248	$\pm$ 0.100	0.503	$\pm$ 0.200	
	$\text{PDW*}$	15.6	$\pm$ 3.0	16.1	$\pm$ 3.0	16.2	$\pm$ 3.0	
	$\text{P-LCC} \times 10^9/\text{L}$	9	$\pm$ 9	63	$\pm$ 25	142	$\pm$ 35	
	$\text{P-LCR \%}$	20.3	$\pm$ 10.0	25.8	$\pm$ 10.0	29.0	$\pm$ 10.0	
<b>BC-5300, BC-5100</b>	$\text{WBC} \times 10^9/\text{L}$	3.25	$\pm$ 0.50	7.75	$\pm$ 1.00	17.50	$\pm$ 2.50	
<b>BC-5380, BC-5180</b>	$\text{Neu\#} \times 10^9/\text{L}$	1.71	$\pm$ 0.30	4.61	$\pm$ 0.70	11.81	$\pm$ 1.58	
<b>QC Mode</b>	$\text{Lym\#} \times 10^9/\text{L}$	1.27	$\pm$ 0.30	2.25	$\pm$ 0.70	3.50	$\pm$ 1.40	
(Software version lower than 1.24.00.16860)	$\text{Mon\#} \times 10^9/\text{L}$	0.08	$\pm$ 0.07	0.27	$\pm$ 0.24	0.44	$\pm$ 0.44	
	$\text{Eos\#} \times 10^9/\text{L}$	0.20	$\pm$ 0.17	0.62	$\pm$ 0.47	1.75	$\pm$ 1.40	
	$\text{Bas\#} \times 10^9/\text{L}$	1.80	$\pm$ 0.33	5.27	$\pm$ 0.78	13.48	$\pm$ 1.75	
	$\text{Neu\%}$	52.5	$\pm$ 9.0	59.5	$\pm$ 9.0	67.5	$\pm$ 9.0	
	$\text{Lym\%}$	39.0	$\pm$ 9.0	29.0	$\pm$ 9.0	20.0	$\pm$ 8.0	
	$\text{Mon\%}$	2.5	$\pm$ 2.0	3.5	$\pm$ 3.0	2.5	$\pm$ 2.5	
	$\text{Eos\%}$	6.0	$\pm$ 5.0	8.0	$\pm$ 6.0	10.0	$\pm$ 8.0	
	$\text{Bas\%}$	55.5	$\pm$ 10.0	68.0	$\pm$ 10.0	77.0	$\pm$ 10.0	
	$\text{RBC} \times 10^{12}/\text{L}$	2.11	$\pm$ 0.18	4.02	$\pm$ 0.24	4.89	$\pm$ 0.30	
	$\text{HGB g/L}$	56	$\pm$ 4	120	$\pm$ 6	156	$\pm$ 8	
	$\text{HCT \%}$	18.0	$\pm$ 1.5	39.8	$\pm$ 2.0	52.3	$\pm$ 2.4	
	$\text{MCV fL}$	85.5	$\pm$ 5.0	99.0	$\pm$ 5.0	107.0	$\pm$ 5.0	
	$\text{MCH pg}$	26.5	$\pm$ 2.5	29.9	$\pm$ 2.5	31.9	$\pm$ 2.5	
	$\text{MCHC g/L}$	310	$\pm$ 30	302	$\pm$ 30	298	$\pm$ 30	
	$\text{RDW-CV \%}$	15.5	$\pm$ 3.0	13.5	$\pm$ 3.0	13.0	$\pm$ 3.0	
	$\text{RDW-SD fL}$	60.0	$\pm$ 8.0	61.0	$\pm$ 8.0	64.5	$\pm$ 8.0	
	$\text{PLT} \times 10^9/\text{L}$	47	$\pm$ 20	240	$\pm$ 40	466	$\pm$ 60	
	$\text{MPV fL}$	9.0	$\pm$ 3.0	9.7	$\pm$ 3.0	9.9	$\pm$ 3.0	
	$\text{PCT \%*}$	0.042	$\pm$ 0.042	0.231	$\pm$ 0.100	0.463	$\pm$ 0.200	
	$\text{PDW*}$	15.6	$\pm$ 3.0	16.0	$\pm$ 3.0	16.2	$\pm$ 3.0	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

## ASSAY VALUES AND EXPECTED RANGES

LOT

BC2405B



2024-07-10

Instrument	Parameter	Low		Normal		High		+++
		LOT	BC2405BL	LOT	BC2405BN	LOT	BC2405BH	
<b>BC-5300, BC-5100</b>	WBC $\times 10^9/L$	3.22	$\pm$ 0.50	7.78	$\pm$ 1.00	17.22	$\pm$ 2.50	
<b>BC-5380, BC-5180</b>	Neu# $\times 10^9/L$	1.70	$\pm$ 0.29	4.62	$\pm$ 0.70	11.49	$\pm$ 1.55	
<b>QC Mode</b> (Software version 1.24.00.16860 or higher)	Lym# $\times 10^9/L$	1.25	$\pm$ 0.29	2.25	$\pm$ 0.70	3.44	$\pm$ 1.38	
	Mon# $\times 10^9/L$	0.09	$\pm$ 0.09	0.23	$\pm$ 0.23	0.45	$\pm$ 0.45	
	Eos# $\times 10^9/L$	0.18	$\pm$ 0.13	0.68	$\pm$ 0.47	1.84	$\pm$ 1.38	
	Bas# $\times 10^9/L$	1.78	$\pm$ 0.33	5.24	$\pm$ 0.78	13.35	$\pm$ 1.73	
	Neu%	52.9	$\pm$ 9.0	59.5	$\pm$ 9.0	66.7	$\pm$ 9.0	
	Lym%	38.7	$\pm$ 9.0	28.9	$\pm$ 9.0	20.0	$\pm$ 8.0	
	Mon%	2.7	$\pm$ 2.5	2.9	$\pm$ 2.9	2.6	$\pm$ 2.6	
	Eos%	5.7	$\pm$ 4.0	8.7	$\pm$ 6.0	10.7	$\pm$ 8.0	
	Bas%	55.2	$\pm$ 10.0	67.4	$\pm$ 10.0	77.5	$\pm$ 10.0	
	RBC $\times 10^{12}/L$	2.15	$\pm$ 0.18	4.02	$\pm$ 0.24	4.89	$\pm$ 0.30	
	HGB g/L	56	$\pm$ 4	120	$\pm$ 6	156	$\pm$ 8	
	HCT %	18.1	$\pm$ 1.5	39.5	$\pm$ 2.0	52.0	$\pm$ 2.4	
	MCV fL	84.0	$\pm$ 5.0	98.3	$\pm$ 5.0	106.4	$\pm$ 5.0	
	MCH pg	26.0	$\pm$ 2.5	29.9	$\pm$ 2.5	31.9	$\pm$ 2.5	
	MCHC g/L	310	$\pm$ 30	304	$\pm$ 30	300	$\pm$ 30	
	RDW-CV %	15.8	$\pm$ 3.0	14.0	$\pm$ 3.0	13.7	$\pm$ 3.0	
	RDW-SD fL	56.8	$\pm$ 8.0	58.0	$\pm$ 8.0	62.3	$\pm$ 8.0	
	PLT $\times 10^9/L$	48	$\pm$ 20	248	$\pm$ 40	486	$\pm$ 60	
	MPV fL	8.6	$\pm$ 3.0	9.7	$\pm$ 3.0	9.8	$\pm$ 3.0	
	PCT %*	0.041	$\pm$ 0.041	0.241	$\pm$ 0.100	0.476	$\pm$ 0.200	
	PDW*	15.6	$\pm$ 3.0	16.2	$\pm$ 3.0	16.3	$\pm$ 3.0	
<b>BC-5000, BC-5150, BC-5120</b>	WBC $\times 10^9/L$	3.35	$\pm$ 0.50	7.85	$\pm$ 1.00	17.40	$\pm$ 2.50	
<b>BC-5130, BC-5140, BC-5000VET</b>	Neu# $\times 10^9/L$	1.62	$\pm$ 0.41	4.35	$\pm$ 0.95	11.34	$\pm$ 2.09	
<b>QC Mode</b>	Lym# $\times 10^9/L$	1.27	$\pm$ 0.30	2.17	$\pm$ 0.63	3.36	$\pm$ 1.22	
	Mon# $\times 10^9/L$	0.23	$\pm$ 0.23	0.54	$\pm$ 0.54	0.82	$\pm$ 0.82	
	Eos# $\times 10^9/L$	0.20	$\pm$ 0.20	0.67	$\pm$ 0.67	1.62	$\pm$ 1.62	
	Bas# $\times 10^9/L$	0.03	$\pm$ 0.03	0.12	$\pm$ 0.12	0.26	$\pm$ 0.26	
	Neu%	48.2	$\pm$ 12.0	55.4	$\pm$ 12.0	65.2	$\pm$ 12.0	
	Lym%	38.0	$\pm$ 9.0	27.7	$\pm$ 8.0	19.3	$\pm$ 7.0	
	Mon%	6.9	$\pm$ 6.9	6.9	$\pm$ 6.9	4.7	$\pm$ 4.7	
	Eos%	5.9	$\pm$ 5.9	8.5	$\pm$ 8.5	9.3	$\pm$ 9.3	
	Bas%	1.0	$\pm$ 1.0	1.5	$\pm$ 1.5	1.5	$\pm$ 1.5	
	RBC $\times 10^{12}/L$	2.14	$\pm$ 0.18	4.06	$\pm$ 0.24	5.00	$\pm$ 0.30	
	HGB g/L	56	$\pm$ 4	122	$\pm$ 6	160	$\pm$ 8	
	HCT %	17.9	$\pm$ 1.5	38.9	$\pm$ 2.0	50.8	$\pm$ 2.4	
	MCV fL	83.6	$\pm$ 5.0	95.9	$\pm$ 5.0	101.6	$\pm$ 5.0	
	MCH pg	26.2	$\pm$ 2.5	30.0	$\pm$ 2.5	32.0	$\pm$ 2.5	
	MCHC g/L	313	$\pm$ 30	313	$\pm$ 30	315	$\pm$ 30	
	RDW-CV %	19.3	$\pm$ 3.0	16.4	$\pm$ 3.0	16.1	$\pm$ 3.0	
	RDW-SD fL	59.3	$\pm$ 8.0	58.3	$\pm$ 8.0	60.8	$\pm$ 8.0	
	PLT $\times 10^9/L$	50	$\pm$ 20	253	$\pm$ 40	502	$\pm$ 60	
	MPV fL	10.6	$\pm$ 3.0	11.6	$\pm$ 3.0	11.8	$\pm$ 3.0	
	PCT %*	0.053	$\pm$ 0.053	0.293	$\pm$ 0.100	0.592	$\pm$ 0.200	
	PDW*	15.5	$\pm$ 3.0	16.1	$\pm$ 3.0	16.5	$\pm$ 3.0	
	P-LCC $\times 10^9/L$ **	16	$\pm$ 15	93	$\pm$ 25	200	$\pm$ 35	
	P-LCR %**	32.7	$\pm$ 10.0	36.6	$\pm$ 10.0	39.9	$\pm$ 10.0	

\* For Research Use Only

\*\* These parameters are not provided on BC-5000/BC-5000 Vet analyzers

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

**BC-5D**  
**HEMATOLOGY CONTROLS**  
**CONTROL**

ASSAY VALUES AND EXPECTED RANGES

**LOT** BC2405B

2024-07-10

Instrument	Parameter	Low		Normal		High		++++
		LOT	BC2405BL	LOT	BC2405BN	LOT	BC2405BH	
BC-5300Vet, BC-5100Vet QC Mode	WBC $\times 10^3/L$	3.25	$\pm$ 0.50	7.75	$\pm$ 1.00	17.50	$\pm$ 2.50	
	Neu# $\times 10^9/L$	1.71	$\pm$ 0.30	4.61	$\pm$ 0.70	11.81	$\pm$ 1.58	
	Lym# $\times 10^9/L$	1.27	$\pm$ 0.30	2.25	$\pm$ 0.70	3.50	$\pm$ 1.40	
	Mon# $\times 10^9/L$	0.08	$\pm$ 0.07	0.27	$\pm$ 0.24	0.44	$\pm$ 0.44	
	Eos# $\times 10^9/L$	0.20	$\pm$ 0.17	0.62	$\pm$ 0.47	1.75	$\pm$ 1.40	
	Neu%	52.5	$\pm$ 9.0	59.5	$\pm$ 9.0	67.5	$\pm$ 9.0	
	Lym%	39.0	$\pm$ 9.0	29.0	$\pm$ 9.0	20.0	$\pm$ 8.0	
	Mon%	2.5	$\pm$ 2.0	3.5	$\pm$ 3.0	2.5	$\pm$ 2.5	
	Eos%	6.0	$\pm$ 5.0	8.0	$\pm$ 6.0	10.0	$\pm$ 8.0	
	RBC $\times 10^{12}/L$	2.11	$\pm$ 0.18	4.02	$\pm$ 0.24	4.89	$\pm$ 0.30	
	HGB g/L	56	$\pm$ 4	120	$\pm$ 6	156	$\pm$ 8	
	HCT %	18.0	$\pm$ 1.5	39.8	$\pm$ 2.0	52.3	$\pm$ 2.4	
	MCV fL	85.5	$\pm$ 5.0	99.0	$\pm$ 5.0	107.0	$\pm$ 5.0	
	MCH pg	26.5	$\pm$ 2.5	29.9	$\pm$ 2.5	31.9	$\pm$ 2.5	
	MCHC g/L	310	$\pm$ 30	302	$\pm$ 30	298	$\pm$ 30	
	RDW-CV %	15.5	$\pm$ 3.0	13.5	$\pm$ 3.0	13.0	$\pm$ 3.0	
	RDW-SD fL	60.0	$\pm$ 8.0	61.0	$\pm$ 8.0	64.5	$\pm$ 8.0	
	PLT $\times 10^9/L$	47	$\pm$ 20	240	$\pm$ 40	466	$\pm$ 60	
	MPV fL	9.0	$\pm$ 3.0	9.7	$\pm$ 3.0	9.9	$\pm$ 3.0	
	PCT %*	0.042	$\pm$ 0.042	0.231	$\pm$ 0.100	0.463	$\pm$ 0.200	
	PDW*	15.6	$\pm$ 3.0	16.0	$\pm$ 3.0	16.2	$\pm$ 3.0	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.



Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China

Tel: +86 755 81888998

Fax: +86 755 26582680

**EC REP**

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726