

# BC-6D

## HEMATOLOGY CONTROLS

**CONTROL**

ASSAY VALUES AND EXPECTED RANGES

**LOT** MB0524A

 2024-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0524AL	LOT	MB0524AN	LOT	MB0524AH
BC-6800,BC-6600	WBC $\times 10^9/L$		3.64 $\pm$ 0.50		7.29 $\pm$ 1.00		19.88 $\pm$ 2.50
QC Mode	Neu# $\times 10^9/L$		2.10 $\pm$ 0.37		3.55 $\pm$ 0.73		10.50 $\pm$ 1.99
	Lym# $\times 10^9/L$		0.89 $\pm$ 0.30		2.66 $\pm$ 0.59		6.40 $\pm$ 1.59
	Mon# $\times 10^9/L$		0.12 $\pm$ 0.12		0.39 $\pm$ 0.39		1.01 $\pm$ 1.01
	Eos# $\times 10^9/L$		0.48 $\pm$ 0.33		0.60 $\pm$ 0.48		1.71 $\pm$ 1.71
	Bas# $\times 10^9/L$		0.05 $\pm$ 0.20		0.09 $\pm$ 0.35		0.26 $\pm$ 1.00
	Neu%		57.8 $\pm$ 10.0		48.7 $\pm$ 10.0		52.8 $\pm$ 10.0
	Lym%		24.4 $\pm$ 8.0		36.5 $\pm$ 8.0		32.2 $\pm$ 8.0
	Mon%		3.2 $\pm$ 3.2		5.3 $\pm$ 5.3		5.1 $\pm$ 5.1
	Eos%		13.3 $\pm$ 9.0		8.2 $\pm$ 6.5		8.6 $\pm$ 8.6
	Bas%		1.3 $\pm$ 5.0		1.3 $\pm$ 5.0		1.3 $\pm$ 5.0
	RBC $\times 10^{12}/L$		2.33 $\pm$ 0.18		4.37 $\pm$ 0.24		5.00 $\pm$ 0.30
	HGB g/dL		6.0 $\pm$ 0.4		12.6 $\pm$ 0.6		15.5 $\pm$ 0.8
	HCT %		19.5 $\pm$ 1.5		40.2 $\pm$ 2.0		49.9 $\pm$ 2.4
	MCV fL		83.8 $\pm$ 5.0		91.9 $\pm$ 5.0		99.7 $\pm$ 5.0
	MCH pg		25.8 $\pm$ 2.5		28.8 $\pm$ 2.5		31.0 $\pm$ 2.5
	MCHC g/dL		30.7 $\pm$ 3.0		31.4 $\pm$ 3.0		31.1 $\pm$ 3.0
	RDW-CV %		16.5 $\pm$ 3.0		15.2 $\pm$ 3.0		14.9 $\pm$ 3.0
	RDW-SD fL		47.7 $\pm$ 6.0		49.8 $\pm$ 6.0		52.0 $\pm$ 8.0
	PLT $\times 10^9/L$		56 $\pm$ 20		196 $\pm$ 40		397 $\pm$ 60
	MPV fL		11.0 $\pm$ 3.0		10.4 $\pm$ 3.0		10.9 $\pm$ 3.0
	PDW		16.2 $\pm$ 3.0		16.4 $\pm$ 3.0		16.6 $\pm$ 3.0
	PCT %		0.062 $\pm$ 0.050		0.204 $\pm$ 0.100		0.433 $\pm$ 0.200
	P-LCR %		35.3 $\pm$ 10.0		29.7 $\pm$ 10.0		32.7 $\pm$ 10.0
	P-LCC $\times 10^9/L$		20 $\pm$ 8		58 $\pm$ 24		130 $\pm$ 50
	IMG#** $\times 10^9/L$		0.06 $\pm$ 0.38		0.09 $\pm$ 0.75		0.38 $\pm$ 2.00
	IMG%**		1.6 $\pm$ 10.0		1.2 $\pm$ 10.0		1.9 $\pm$ 10.0
	NRBC# $\times 10^9/L$		0.390 $\pm$ 0.390		0.353 $\pm$ 0.353		0.531 $\pm$ 0.531
	NRBC% /100WBC		10.72 $\pm$ 10.72		4.84 $\pm$ 4.84		2.67 $\pm$ 2.67

Before using, refer to the instruction sheet for mixing directions.

\*Research use only.

\*\*Software version 1.09 or higher.

# BC-6D

## HEMATOLOGY CONTROLS

**CONTROL**

ASSAY VALUES AND EXPECTED RANGES

**LOT** MB0524A

 2024-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0524AL	LOT	MB0524AN	LOT	MB0524AH
BC-6000	WBC $\times 10^9/L$	3.57	$\pm$ 0.50	7.52	$\pm$ 1.00	20.60	$\pm$ 2.50
BC-6100	Neu# $\times 10^9/L$	2.08	$\pm$ 0.36	3.65	$\pm$ 0.76	10.75	$\pm$ 2.06
BC-6200	Lym# $\times 10^9/L$	0.84	$\pm$ 0.29	2.79	$\pm$ 0.61	6.90	$\pm$ 1.65
BC-6000Plus	Mon# $\times 10^9/L$	0.12	$\pm$ 0.12	0.38	$\pm$ 0.38	0.95	$\pm$ 0.95
BC-6100Plus	Eos# $\times 10^9/L$	0.48	$\pm$ 0.24	0.59	$\pm$ 0.53	1.71	$\pm$ 1.71
QC Mode	Bas# $\times 10^9/L$	0.05	$\pm$ 0.05	0.11	$\pm$ 0.11	0.29	$\pm$ 0.29
	Neu%	58.4	$\pm$ 10.0	48.6	$\pm$ 10.0	52.2	$\pm$ 10.0
	Lym%	23.4	$\pm$ 8.0	37.1	$\pm$ 8.0	33.5	$\pm$ 8.0
	Mon%	3.3	$\pm$ 3.3	5.1	$\pm$ 5.1	4.6	$\pm$ 4.6
	Eos%	13.4	$\pm$ 6.5	7.8	$\pm$ 7.0	8.3	$\pm$ 8.3
	Bas%	1.5	$\pm$ 1.5	1.4	$\pm$ 1.4	1.4	$\pm$ 1.4
	RBC $\times 10^{12}/L$	2.36	$\pm$ 0.18	4.43	$\pm$ 0.24	5.07	$\pm$ 0.30
	HGB g/dL	6.0	$\pm$ 0.4	12.6	$\pm$ 0.6	15.7	$\pm$ 0.8
	HCT %	20.2	$\pm$ 1.5	41.3	$\pm$ 2.0	50.6	$\pm$ 2.4
	MCV fL	85.5	$\pm$ 5.0	93.2	$\pm$ 5.0	99.9	$\pm$ 5.0
	MCH pg	25.4	$\pm$ 2.5	28.4	$\pm$ 2.5	31.0	$\pm$ 2.5
	MCHC g/dL	29.7	$\pm$ 3.0	30.5	$\pm$ 3.0	31.0	$\pm$ 3.0
	RDW-CV %	16.8	$\pm$ 3.0	15.5	$\pm$ 3.0	14.6	$\pm$ 3.0
	RDW-SD fL	51.8	$\pm$ 6.0	52.7	$\pm$ 6.0	53.4	$\pm$ 8.0
	PLT $\times 10^9/L$	59	$\pm$ 20	212	$\pm$ 40	429	$\pm$ 60
	MPV fL	12.6	$\pm$ 3.0	11.2	$\pm$ 3.0	11.4	$\pm$ 3.0
	PDW	16.2	$\pm$ 3.0	16.4	$\pm$ 3.0	16.6	$\pm$ 3.0
	PCT %	0.074	$\pm$ 0.050	0.237	$\pm$ 0.100	0.489	$\pm$ 0.200
	P-LCR %	45.8	$\pm$ 10.0	35.0	$\pm$ 10.0	36.7	$\pm$ 10.0
	P-LCC $\times 10^9/L$	27	$\pm$ 8	74	$\pm$ 24	157	$\pm$ 50
	IMG# $\times 10^9/L$	0.04	$\pm$ 0.34	0.10	$\pm$ 0.77	0.27	$\pm$ 2.08
	IMG%	1.2	$\pm$ 10.0	1.3	$\pm$ 10.0	1.3	$\pm$ 10.0
	NRBC# $\times 10^9/L$	0.322	$\pm$ 0.322	0.307	$\pm$ 0.307	0.428	$\pm$ 0.428
	NRBC% /100WBC	9.03	$\pm$ 9.03	4.08	$\pm$ 4.08	2.08	$\pm$ 2.08
	Micro%	3.3	$\pm$ 15.0	2.5	$\pm$ 4.0	2.0	$\pm$ 2.5
	Macro%	4.5	$\pm$ 3.0	5.7	$\pm$ 5.0	7.4	$\pm$ 10.0
	Neu-X	396.6	$\pm$ 120.0	407.1	$\pm$ 120.0	408.5	$\pm$ 120.0
	Neu-Y	514.9	$\pm$ 120.0	515.3	$\pm$ 150.0	515.9	$\pm$ 150.0
	Mon-X	236.2	$\pm$ 120.0	219.0	$\pm$ 120.0	220.9	$\pm$ 120.0
	PLT-I $\times 10^9/L$	59	$\pm$ 30	212	$\pm$ 60	429	$\pm$ 90

Before using, refer to the instruction sheet for mixing directions.

# BC-6D

## HEMATOLOGY CONTROLS

**CONTROL**

ASSAY VALUES AND EXPECTED RANGES

**LOT** MB0524A

 2024-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0524AL	LOT	MB0524AN	LOT	MB0524AH
BC-6800Plus	WBC $\times 10^9/L$		3.82 $\pm$ 0.50		7.54 $\pm$ 1.00		20.99 $\pm$ 2.50
BC-6700Plus	Neu# $\times 10^9/L$		2.23 $\pm$ 0.39		3.71 $\pm$ 0.76		11.14 $\pm$ 2.11
BC-6600Plus	Lym# $\times 10^9/L$		0.91 $\pm$ 0.31		2.76 $\pm$ 0.61		6.88 $\pm$ 1.68
QC Mode	Mon# $\times 10^9/L$		0.13 $\pm$ 0.13		0.38 $\pm$ 0.38		0.94 $\pm$ 0.94
	Eos# $\times 10^9/L$		0.49 $\pm$ 0.35		0.58 $\pm$ 0.53		1.74 $\pm$ 1.74
	Bas# $\times 10^9/L$		0.06 $\pm$ 0.06		0.11 $\pm$ 0.11		0.29 $\pm$ 0.29
	Neu%		58.6 $\pm$ 10.0		49.2 $\pm$ 10.0		53.0 $\pm$ 10.0
	Lym%		23.8 $\pm$ 8.0		36.6 $\pm$ 8.0		32.8 $\pm$ 8.0
	Mon%		3.3 $\pm$ 3.3		5.1 $\pm$ 5.1		4.5 $\pm$ 4.5
	Eos%		12.8 $\pm$ 9.0		7.7 $\pm$ 7.0		8.3 $\pm$ 8.3
	Bas%		1.5 $\pm$ 1.5		1.4 $\pm$ 1.4		1.4 $\pm$ 1.4
	RBC $\times 10^{12}/L$		2.36 $\pm$ 0.18		4.43 $\pm$ 0.24		5.04 $\pm$ 0.30
	HGB g/dL		6.0 $\pm$ 0.4		12.5 $\pm$ 0.6		15.4 $\pm$ 0.8
	HCT %		19.5 $\pm$ 1.5		40.7 $\pm$ 2.0		50.4 $\pm$ 2.4
	MCV fL		82.7 $\pm$ 5.0		91.9 $\pm$ 5.0		100.0 $\pm$ 5.0
	MCH pg		25.4 $\pm$ 2.5		28.2 $\pm$ 2.5		30.6 $\pm$ 2.5
	MCHC g/dL		30.7 $\pm$ 3.0		30.7 $\pm$ 3.0		30.6 $\pm$ 3.0
	RDW-CV %		16.4 $\pm$ 3.0		15.3 $\pm$ 3.0		14.6 $\pm$ 3.0
	RDW-SD fL		46.8 $\pm$ 6.0		48.9 $\pm$ 6.0		50.3 $\pm$ 8.0
	PLT $\times 10^9/L$		55 $\pm$ 20		203 $\pm$ 40		407 $\pm$ 60
	PLT-H $\times 10^9/L$		56 $\pm$ 20		202 $\pm$ 40		406 $\pm$ 60
	MPV fL		10.9 $\pm$ 3.0		10.4 $\pm$ 3.0		11.0 $\pm$ 3.0
	PDW		15.8 $\pm$ 3.0		16.3 $\pm$ 3.0		16.6 $\pm$ 3.0
	PCT %		0.060 $\pm$ 0.050		0.211 $\pm$ 0.100		0.448 $\pm$ 0.200
	P-LCR %		35.6 $\pm$ 10.0		30.0 $\pm$ 10.0		33.3 $\pm$ 10.0
	P-LCC $\times 10^9/L$		20 $\pm$ 8		61 $\pm$ 24		136 $\pm$ 50
	IMG# $\times 10^9/L$		0.05 $\pm$ 0.42		0.09 $\pm$ 0.75		0.27 $\pm$ 2.08
	IMG%		1.2 $\pm$ 10.0		1.2 $\pm$ 10.0		1.3 $\pm$ 10.0
	NRBC# $\times 10^9/L$		0.310 $\pm$ 0.310		0.274 $\pm$ 0.274		0.443 $\pm$ 0.443
	NRBC% /100WBC		8.12 $\pm$ 8.12		3.63 $\pm$ 3.63		2.11 $\pm$ 2.11
	Micro%		3.7 $\pm$ 15.0		2.7 $\pm$ 4.0		2.2 $\pm$ 2.5
	Macro%		4.1 $\pm$ 3.0		4.5 $\pm$ 5.0		5.2 $\pm$ 10.0
	Neu-X		386.4 $\pm$ 120.0		380.6 $\pm$ 120.0		385.6 $\pm$ 120.0
	Neu-Y		506.0 $\pm$ 120.0		507.2 $\pm$ 150.0		509.5 $\pm$ 150.0
	Mon-X		237.4 $\pm$ 120.0		210.9 $\pm$ 120.0		215.0 $\pm$ 120.0
	PLT-I $\times 10^9/L$		55 $\pm$ 30		203 $\pm$ 60		407 $\pm$ 90

Before using, refer to the instruction sheet for mixing directions.

# BC-6D

## HEMATOLOGY CONTROLS

**CONTROL**

ASSAY VALUES AND EXPECTED RANGES

**LOT** MB0524A

 2024-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0524AL	LOT	MB0524AN	LOT	MB0524AH
BC-7500[B] CRP	WBC $\times 10^9/L$	3.61	$\pm$ 0.50	7.47	$\pm$ 1.00	20.41	$\pm$ 2.50
BC-7500[N] CRP	Neu# $\times 10^9/L$	2.10	$\pm$ 0.37	3.65	$\pm$ 0.75	10.75	$\pm$ 2.04
BC-7500[R] CRP	Lym# $\times 10^9/L$	0.86	$\pm$ 0.29	2.76	$\pm$ 0.60	6.78	$\pm$ 1.64
BC-7500[NR] CRP	Mon# $\times 10^9/L$	0.12	$\pm$ 0.12	0.38	$\pm$ 0.38	0.92	$\pm$ 0.92
BC-7300[B] CRP	Eos# $\times 10^9/L$	0.48	$\pm$ 0.33	0.58	$\pm$ 0.58	1.67	$\pm$ 1.67
BC-7300[N] CRP	Bas# $\times 10^9/L$	0.05	$\pm$ 0.05	0.10	$\pm$ 0.10	0.29	$\pm$ 0.29
BC-7300[R] CRP	Neu%	58.0	$\pm$ 10.0	48.9	$\pm$ 10.0	52.7	$\pm$ 10.0
BC-7300[NR] CRP	Lym%	23.8	$\pm$ 8.0	36.9	$\pm$ 8.0	33.2	$\pm$ 8.0
BC-7500[B] CS	Mon%	3.3	$\pm$ 3.3	5.1	$\pm$ 5.1	4.5	$\pm$ 4.5
BC-7500[N] CS	Eos%	13.4	$\pm$ 9.0	7.7	$\pm$ 7.7	8.2	$\pm$ 8.2
BC-7500[R] CS	Bas%	1.5	$\pm$ 1.5	1.4	$\pm$ 1.4	1.4	$\pm$ 1.4
BC-7500[NR] CS	RBC $\times 10^{12}/L$	2.29	$\pm$ 0.18	4.38	$\pm$ 0.24	5.05	$\pm$ 0.30
QC Mode	HGB g/dL	5.9	$\pm$ 0.4	12.6	$\pm$ 0.6	15.7	$\pm$ 0.8
	HCT %	19.0	$\pm$ 1.5	40.4	$\pm$ 2.0	50.7	$\pm$ 2.4
	MCV fL	83.1	$\pm$ 5.0	92.3	$\pm$ 5.0	100.4	$\pm$ 5.0
	MCH pg	25.8	$\pm$ 2.5	28.8	$\pm$ 2.5	31.1	$\pm$ 2.5
	MCHC g/dL	31.0	$\pm$ 3.0	31.2	$\pm$ 3.0	31.0	$\pm$ 3.0
	RDW-CV %	17.6	$\pm$ 3.0	15.4	$\pm$ 3.0	14.4	$\pm$ 3.0
	RDW-SD fL	51.7	$\pm$ 6.0	50.6	$\pm$ 6.0	51.1	$\pm$ 8.0
	PLT $\times 10^9/L$	57	$\pm$ 20	200	$\pm$ 40	404	$\pm$ 60
	MPV fL	11.0	$\pm$ 3.0	10.5	$\pm$ 3.0	11.0	$\pm$ 3.0
	PDW	16.3	$\pm$ 3.0	16.5	$\pm$ 3.0	16.8	$\pm$ 3.0
PCT %	0.063	$\pm$ 0.050	0.210	$\pm$ 0.100	0.444	$\pm$ 0.200	
P-LCR %	36.8	$\pm$ 10.0	30.9	$\pm$ 10.0	33.8	$\pm$ 10.0	
P-LCC $\times 10^9/L$	21	$\pm$ 8	62	$\pm$ 24	137	$\pm$ 50	
IMG# $\times 10^9/L$	0.04	$\pm$ 0.34	0.10	$\pm$ 0.77	0.27	$\pm$ 2.08	
IMG%	1.2	$\pm$ 10.0	1.3	$\pm$ 10.0	1.3	$\pm$ 10.0	
NRBC# $\times 10^9/L$	0.320	$\pm$ 0.320	0.311	$\pm$ 0.311	0.408	$\pm$ 0.408	
NRBC% /100WBC	8.87	$\pm$ 8.87	4.17	$\pm$ 4.17	2.00	$\pm$ 2.00	

Before using, refer to the instruction sheet for mixing directions.

# BC-6D

## HEMATOLOGY CONTROLS

**CONTROL**

ASSAY VALUES AND EXPECTED RANGES

**LOT** MB0524A

 2024-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0524AL	LOT	MB0524AN	LOT	MB0524AH
BC-700[B]	WBC $\times 10^9/L$	3.66	$\pm$ 0.50	7.67	$\pm$ 1.00	21.04	$\pm$ 2.50
BC-760[B]	Neu# $\times 10^9/L$	2.24	$\pm$ 0.37	3.90	$\pm$ 0.77	11.55	$\pm$ 2.11
BC-700[B] CS	Lym# $\times 10^9/L$	0.84	$\pm$ 0.30	2.81	$\pm$ 0.62	6.82	$\pm$ 1.69
BC-760[B] CS	Mon# $\times 10^9/L$	0.12	$\pm$ 0.12	0.39	$\pm$ 0.39	0.95	$\pm$ 0.95
BC-700[R]	Eos# $\times 10^9/L$	0.41	$\pm$ 0.33	0.46	$\pm$ 0.46	1.43	$\pm$ 1.43
BC-720[R]	Bas# $\times 10^9/L$	0.05	$\pm$ 0.18	0.11	$\pm$ 0.40	0.29	$\pm$ 1.04
BC-760[R]	Neu%	61.0	$\pm$ 10.0	50.9	$\pm$ 10.0	54.9	$\pm$ 10.0
BC-780[R]	Lym%	23.0	$\pm$ 8.0	36.6	$\pm$ 8.0	32.4	$\pm$ 8.0
BC-700[R] CS	Mon%	3.3	$\pm$ 3.3	5.1	$\pm$ 5.1	4.5	$\pm$ 4.5
BC-760[R] CS	Eos%	11.3	$\pm$ 9.0	6.0	$\pm$ 6.0	6.8	$\pm$ 6.8
BP 200n	Bas%	1.4	$\pm$ 5.0	1.4	$\pm$ 5.0	1.4	$\pm$ 5.0
BP 260n	RBC $\times 10^{12}/L$	2.33	$\pm$ 0.18	4.40	$\pm$ 0.24	5.06	$\pm$ 0.30
QC Mode	HGB g/dL	5.9	$\pm$ 0.4	12.5	$\pm$ 0.6	15.5	$\pm$ 0.8
	HCT %	19.7	$\pm$ 1.5	41.1	$\pm$ 2.0	51.4	$\pm$ 2.4
	MCV fL	84.4	$\pm$ 5.0	93.4	$\pm$ 5.0	101.6	$\pm$ 5.0
	MCH pg	25.3	$\pm$ 2.5	28.4	$\pm$ 2.5	30.6	$\pm$ 2.5
	MCHC g/dL	30.0	$\pm$ 3.0	30.4	$\pm$ 3.0	30.2	$\pm$ 3.0
	RDW-CV %	17.4	$\pm$ 3.0	15.5	$\pm$ 3.0	14.4	$\pm$ 3.0
	RDW-SD fL	51.1	$\pm$ 6.0	50.4	$\pm$ 6.0	50.8	$\pm$ 8.0
	PLT $\times 10^9/L$	59	$\pm$ 20	206	$\pm$ 40	416	$\pm$ 60
	PLT-H $\times 10^9/L$	60	$\pm$ 20	208	$\pm$ 40	422	$\pm$ 60
	MPV fL	11.1	$\pm$ 3.0	10.2	$\pm$ 3.0	10.8	$\pm$ 3.0
	PDW	16.2	$\pm$ 3.0	16.5	$\pm$ 3.0	16.9	$\pm$ 3.0
	PCT %	0.065	$\pm$ 0.050	0.210	$\pm$ 0.100	0.449	$\pm$ 0.200
	P-LCR %	36.6	$\pm$ 10.0	29.5	$\pm$ 10.0	32.7	$\pm$ 10.0
	P-LCC $\times 10^9/L$	22	$\pm$ 8	61	$\pm$ 24	136	$\pm$ 50
	IMG# $\times 10^9/L$	0.05	$\pm$ 0.39	0.10	$\pm$ 0.77	0.25	$\pm$ 2.09
	IMG%	1.3	$\pm$ 10.0	1.3	$\pm$ 10.0	1.2	$\pm$ 10.0
	NRBC# $\times 10^9/L$	0.088	$\pm$ 0.088	0.144	$\pm$ 0.144	0.448	$\pm$ 0.448
	NRBC% /100WBC	2.41	$\pm$ 2.41	1.88	$\pm$ 1.88	2.13	$\pm$ 2.13
	IPF %	2.5	$\pm$ 5.0	2.3	$\pm$ 5.0	2.5	$\pm$ 5.0
	ESR mm/h	38.12	$\pm$ 10.00	14.67	$\pm$ 4.00	/	$\pm$ /

The ESR is not applicable for High level, and only ESR is involved in BP 200n and BP 260n.

Before using, refer to the instruction sheet for mixing directions.



Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen 518057, P.R.China

Tel: +86 755 26582479 26582888

Fax: +86 755 26582934 26582500



Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726



P/N: 046-003618-00 (13.0)