

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0324A

 2024-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0324AL	LOT	MB0324AN	LOT	MB0324AH
BC-6800,BC-6600	WBC $\times 10^9/L$		3.61 \pm 0.50		7.34 \pm 1.00		20.32 \pm 2.50
QC Mode	Neu# $\times 10^9/L$		2.02 \pm 0.36		3.55 \pm 0.74		11.31 \pm 2.04
	Lym# $\times 10^9/L$		0.92 \pm 0.29		2.67 \pm 0.59		6.56 \pm 1.63
	Mon# $\times 10^9/L$		0.13 \pm 0.13		0.39 \pm 0.39		0.91 \pm 0.91
	Eos# $\times 10^9/L$		0.49 \pm 0.33		0.63 \pm 0.48		1.32 \pm 1.32
	Bas# $\times 10^9/L$		0.05 \pm 0.20		0.10 \pm 0.39		0.22 \pm 1.00
	Neu%		56.1 \pm 10.0		48.4 \pm 10.0		55.6 \pm 10.0
	Lym%		25.5 \pm 8.0		36.4 \pm 8.0		32.3 \pm 8.0
	Mon%		3.6 \pm 3.6		5.3 \pm 5.3		4.5 \pm 4.5
	Eos%		13.5 \pm 9.0		8.6 \pm 6.5		6.5 \pm 6.5
	Bas%		1.3 \pm 5.0		1.3 \pm 5.0		1.1 \pm 5.0
	RBC $\times 10^{12}/L$		2.36 \pm 0.18		4.26 \pm 0.24		5.20 \pm 0.30
	HGB g/dL		6.1 \pm 0.4		12.4 \pm 0.6		16.2 \pm 0.8
	HCT %		19.7 \pm 1.5		39.7 \pm 2.0		51.9 \pm 2.4
	MCV fL		83.6 \pm 5.0		93.3 \pm 5.0		99.8 \pm 5.0
	MCH pg		25.8 \pm 2.5		29.1 \pm 2.5		31.2 \pm 2.5
	MCHC g/dL		30.9 \pm 3.0		31.2 \pm 3.0		31.2 \pm 3.0
	RDW-CV %		16.5 \pm 3.0		14.7 \pm 3.0		14.3 \pm 3.0
	RDW-SD fL		49.3 \pm 6.0		47.9 \pm 6.0		50.8 \pm 8.0
	PLT $\times 10^9/L$		53 \pm 20		193 \pm 40		387 \pm 60
	MPV fL		11.6 \pm 3.0		10.9 \pm 3.0		11.5 \pm 3.0
	PDW		15.8 \pm 3.0		16.8 \pm 3.0		16.5 \pm 3.0
	PCT %		0.061 \pm 0.050		0.210 \pm 0.100		0.445 \pm 0.200
	P-LCR %		40.8 \pm 10.0		32.3 \pm 10.0		37.2 \pm 10.0
	P-LCC $\times 10^9/L$		22 \pm 8		62 \pm 24		144 \pm 50
	IMG#** $\times 10^9/L$		0.05 \pm 0.36		0.10 \pm 0.72		0.26 \pm 2.00
	IMG%**		1.4 \pm 10.0		1.4 \pm 10.0		1.3 \pm 10.0
	NRBC# $\times 10^9/L$		0.329 \pm 0.329		0.363 \pm 0.363		0.608 \pm 0.608
	NRBC% /100WBC		9.10 \pm 9.10		4.95 \pm 4.95		2.99 \pm 2.99

Before using, refer to the instruction sheet for mixing directions.

*Research use only.

**Software version 1.09 or higher.

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0324A

 2024-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0324AL	LOT	MB0324AN	LOT	MB0324AH
BC-6000	WBC $\times 10^9/L$	3.51	\pm 0.50	7.37	\pm 1.00	21.17	\pm 2.50
BC-6100	Neu# $\times 10^9/L$	1.97	\pm 0.36	3.58	\pm 0.74	11.69	\pm 2.12
BC-6200	Lym# $\times 10^9/L$	0.92	\pm 0.29	2.71	\pm 0.59	6.90	\pm 1.70
BC-6000Plus	Mon# $\times 10^9/L$	0.13	\pm 0.13	0.38	\pm 0.38	0.95	\pm 0.95
BC-6100Plus	Eos# $\times 10^9/L$	0.44	\pm 0.23	0.60	\pm 0.52	1.35	\pm 1.35
QC Mode	Bas# $\times 10^9/L$	0.05	\pm 0.05	0.10	\pm 0.08	0.28	\pm 0.28
	Neu%	56.1	\pm 10.0	48.5	\pm 10.0	55.2	\pm 10.0
	Lym%	26.1	\pm 8.0	36.8	\pm 8.0	32.6	\pm 8.0
	Mon%	3.7	\pm 3.7	5.1	\pm 5.1	4.5	\pm 4.5
	Eos%	12.6	\pm 6.5	8.2	\pm 7.0	6.4	\pm 6.4
	Bas%	1.5	\pm 1.5	1.4	\pm 1.0	1.3	\pm 1.3
	RBC $\times 10^{12}/L$	2.38	\pm 0.18	4.29	\pm 0.24	5.31	\pm 0.30
	HGB g/dL	6.1	\pm 0.4	12.4	\pm 0.6	16.3	\pm 0.8
	HCT %	20.5	\pm 1.5	40.5	\pm 2.0	53.4	\pm 2.4
	MCV fL	86.0	\pm 5.0	94.5	\pm 5.0	100.5	\pm 5.0
	MCH pg	25.6	\pm 2.5	28.9	\pm 2.5	30.7	\pm 2.5
	MCHC g/dL	29.8	\pm 3.0	30.6	\pm 3.0	30.5	\pm 3.0
	RDW-CV %	17.3	\pm 3.0	15.1	\pm 3.0	14.5	\pm 3.0
	RDW-SD fL	53.4	\pm 6.0	52.2	\pm 6.0	53.4	\pm 8.0
	PLT $\times 10^9/L$	53	\pm 20	202	\pm 40	402	\pm 60
	MPV fL	13.1	\pm 3.0	11.6	\pm 3.0	11.9	\pm 3.0
	PDW	16.0	\pm 3.0	16.5	\pm 3.0	16.5	\pm 3.0
	PCT %	0.069	\pm 0.050	0.234	\pm 0.100	0.478	\pm 0.200
	P-LCR %	50.0	\pm 10.0	37.0	\pm 10.0	40.6	\pm 10.0
	P-LCC $\times 10^9/L$	27	\pm 8	75	\pm 24	163	\pm 50
	IMG# $\times 10^9/L$	0.05	\pm 0.39	0.10	\pm 0.77	0.28	\pm 2.16
	IMG%	1.3	\pm 10.0	1.3	\pm 10.0	1.3	\pm 10.0
	NRBC# $\times 10^9/L$	0.241	\pm 0.241	0.298	\pm 0.298	0.576	\pm 0.576
	NRBC% /100WBC	6.87	\pm 6.87	4.05	\pm 4.05	2.72	\pm 2.72
	Micro%	3.3	\pm 15.0	2.3	\pm 4.0	1.9	\pm 2.5
	Macro%	4.6	\pm 3.0	5.9	\pm 5.0	7.6	\pm 10.0
	Neu-X	397.3	\pm 120.0	410.6	\pm 120.0	395.7	\pm 120.0
	Neu-Y	518.3	\pm 120.0	518.3	\pm 150.0	514.6	\pm 150.0
	Mon-X	222.1	\pm 120.0	218.1	\pm 120.0	204.4	\pm 120.0
	PLT-I $\times 10^9/L$	53	\pm 30	202	\pm 60	402	\pm 90

Before using, refer to the instruction sheet for mixing directions.

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0324A

 2024-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0324AL	LOT	MB0324AN	LOT	MB0324AH
BC-6800Plus	WBC $\times 10^9/L$	3.77	\pm 0.50	7.44	\pm 1.00	21.19	\pm 2.50
BC-6700Plus	Neu# $\times 10^9/L$	2.11	\pm 0.38	3.64	\pm 0.75	11.74	\pm 2.12
BC-6600Plus	Lym# $\times 10^9/L$	0.97	\pm 0.31	2.73	\pm 0.60	6.89	\pm 1.70
QC Mode	Mon# $\times 10^9/L$	0.14	\pm 0.14	0.38	\pm 0.38	0.95	\pm 0.95
	Eos# $\times 10^9/L$	0.49	\pm 0.34	0.59	\pm 0.53	1.31	\pm 1.31
	Bas# $\times 10^9/L$	0.06	\pm 0.06	0.10	\pm 0.10	0.30	\pm 0.30
	Neu%	56.1	\pm 10.0	48.9	\pm 10.0	55.4	\pm 10.0
	Lym%	25.7	\pm 8.0	36.7	\pm 8.0	32.5	\pm 8.0
	Mon%	3.6	\pm 3.6	5.1	\pm 5.1	4.5	\pm 4.5
	Eos%	13.1	\pm 9.0	7.9	\pm 7.0	6.2	\pm 6.2
RBC $\times 10^{12}/L$	Bas%	1.5	\pm 1.5	1.4	\pm 1.4	1.4	\pm 1.4
	HGB g/dL	2.38	\pm 0.18	4.30	\pm 0.24	5.22	\pm 0.30
	HCT %	6.0	\pm 0.4	12.2	\pm 0.6	16.0	\pm 0.8
	MCV fL	19.8	\pm 1.5	40.1	\pm 2.0	52.4	\pm 2.4
	MCH pg	83.2	\pm 5.0	93.2	\pm 5.0	100.4	\pm 5.0
	MCHC g/dL	25.2	\pm 2.5	28.4	\pm 2.5	30.7	\pm 2.5
	RDW-CV %	30.3	\pm 3.0	30.4	\pm 3.0	30.5	\pm 3.0
	RDW-SD fL	16.8	\pm 3.0	14.9	\pm 3.0	14.6	\pm 3.0
	PLT $\times 10^9/L$	48.5	\pm 6.0	48.2	\pm 6.0	50.6	\pm 8.0
	PLT-H $\times 10^9/L$	50	\pm 20	193	\pm 40	380	\pm 60
	MPV fL	53	\pm 20	196	\pm 40	386	\pm 60
	PDW	11.6	\pm 3.0	10.9	\pm 3.0	11.4	\pm 3.0
	PCT %	15.5	\pm 3.0	16.5	\pm 3.0	16.5	\pm 3.0
P-LCR %	P-LCR %	0.058	\pm 0.050	0.210	\pm 0.100	0.433	\pm 0.200
	P-LCC $\times 10^9/L$	40.4	\pm 10.0	32.4	\pm 10.0	37.1	\pm 10.0
	IMG# $\times 10^9/L$	20	\pm 8	63	\pm 24	141	\pm 50
	IMG%	0.05	\pm 0.39	0.10	\pm 0.77	0.28	\pm 2.16
	NRBC# $\times 10^9/L$	1.3	\pm 10.0	1.3	\pm 10.0	1.3	\pm 10.0
	NRBC% /100WBC	0.236	\pm 0.236	0.289	\pm 0.289	0.557	\pm 0.557
	Micro%	6.27	\pm 6.27	3.88	\pm 3.88	2.63	\pm 2.63
	Macro%	3.7	\pm 15.0	2.6	\pm 4.0	2.1	\pm 2.5
	Neu-X	4.1	\pm 3.0	4.5	\pm 5.0	5.1	\pm 10.0
	Neu-Y	380.9	\pm 120.0	387.8	\pm 120.0	383.4	\pm 120.0
Mon-X	Mon-X	511.5	\pm 120.0	510.1	\pm 150.0	506.5	\pm 150.0
	PLT-I $\times 10^9/L$	216.1	\pm 120.0	216.0	\pm 120.0	196.5	\pm 120.0
Mon-X	51	\pm 30	196	\pm 60	387	\pm 90	

Before using, refer to the instruction sheet for mixing directions.

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0324A

 2024-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0324AL	LOT	MB0324AN	LOT	MB0324AH
BC-7500[B] CRP	WBC $\times 10^9/L$	3.61	\pm 0.50	7.32	\pm 1.00	21.17	\pm 2.50
BC-7500[N] CRP	Neu# $\times 10^9/L$	2.03	\pm 0.37	3.56	\pm 0.74	11.69	\pm 2.12
BC-7500[R] CRP	Lym# $\times 10^9/L$	0.94	\pm 0.29	2.69	\pm 0.59	6.90	\pm 1.70
BC-7500[NR] CRP	Mon# $\times 10^9/L$	0.13	\pm 0.13	0.37	\pm 0.37	0.95	\pm 0.95
BC-7300[B] CRP	Eos# $\times 10^9/L$	0.46	\pm 0.33	0.60	\pm 0.60	1.33	\pm 1.33
BC-7300[N] CRP	Bas# $\times 10^9/L$	0.05	\pm 0.05	0.10	\pm 0.10	0.30	\pm 0.30
BC-7300[R] CRP	Neu%	56.2	\pm 10.0	48.5	\pm 10.0	55.2	\pm 10.0
BC-7300[NR] CRP	Lym%	26.0	\pm 8.0	36.8	\pm 8.0	32.6	\pm 8.0
BC-7500[B] CS	Mon%	3.6	\pm 3.6	5.1	\pm 5.1	4.5	\pm 4.5
BC-7500[N] CS	Eos%	12.7	\pm 9.0	8.2	\pm 8.2	6.3	\pm 6.3
BC-7500[R] CS	Bas%	1.5	\pm 1.5	1.4	\pm 1.4	1.4	\pm 1.4
BC-7500[NR] CS	RBC $\times 10^{12}/L$	2.31	\pm 0.18	4.25	\pm 0.24	5.29	\pm 0.30
QC Mode	HGB g/dL	5.9	\pm 0.4	12.3	\pm 0.6	16.3	\pm 0.8
	HCT %	19.3	\pm 1.5	39.9	\pm 2.0	53.5	\pm 2.4
	MCV fL	83.7	\pm 5.0	93.8	\pm 5.0	101.1	\pm 5.0
	MCH pg	25.5	\pm 2.5	28.9	\pm 2.5	30.8	\pm 2.5
	MCHC g/dL	30.5	\pm 3.0	30.9	\pm 3.0	30.5	\pm 3.0
	RDW-CV %	17.9	\pm 3.0	15.0	\pm 3.0	14.3	\pm 3.0
	RDW-SD fL	53.0	\pm 6.0	49.9	\pm 6.0	50.9	\pm 8.0
	PLT $\times 10^9/L$	53	\pm 20	193	\pm 40	380	\pm 60
	MPV fL	11.7	\pm 3.0	10.9	\pm 3.0	11.5	\pm 3.0
	PDW	16.2	\pm 3.0	16.8	\pm 3.0	16.7	\pm 3.0
PCT %	0.062	\pm 0.050	0.210	\pm 0.100	0.437	\pm 0.200	
P-LCR %	41.8	\pm 10.0	33.3	\pm 10.0	38.2	\pm 10.0	
P-LCC $\times 10^9/L$	22	\pm 8	64	\pm 24	145	\pm 50	
IMG# $\times 10^9/L$	0.05	\pm 0.39	0.10	\pm 0.77	0.28	\pm 2.16	
IMG%	1.3	\pm 10.0	1.3	\pm 10.0	1.3	\pm 10.0	
NRBC# $\times 10^9/L$	0.235	\pm 0.235	0.300	\pm 0.300	0.550	\pm 0.550	
NRBC% /100WBC	6.51	\pm 6.51	4.10	\pm 4.10	2.60	\pm 2.60	

Before using, refer to the instruction sheet for mixing directions.

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0324A

 2024-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0324AL	LOT	MB0324AN	LOT	MB0324AH
BC-700[B]	WBC $\times 10^9/L$	3.65	\pm 0.50	7.63	\pm 1.00	21.11	\pm 2.50
BC-760[B]	Neu# $\times 10^9/L$	2.13	\pm 0.37	3.84	\pm 0.77	12.20	\pm 2.12
BC-700[B] CS	Lym# $\times 10^9/L$	0.92	\pm 0.30	2.80	\pm 0.61	6.69	\pm 1.69
BC-760[B] CS	Mon# $\times 10^9/L$	0.13	\pm 0.13	0.38	\pm 0.38	0.93	\pm 0.93
BC-700[R]	Eos# $\times 10^9/L$	0.42	\pm 0.33	0.50	\pm 0.50	0.99	\pm 0.99
BC-720[R]	Bas# $\times 10^9/L$	0.05	\pm 0.18	0.11	\pm 0.40	0.30	\pm 1.08
BC-760[R]	Neu%	58.3	\pm 10.0	50.3	\pm 10.0	57.8	\pm 10.0
BC-780[R]	Lym%	25.2	\pm 8.0	36.7	\pm 8.0	31.7	\pm 8.0
BC-700[R] CS	Mon%	3.6	\pm 3.6	5.0	\pm 5.0	4.4	\pm 4.4
BC-760[R] CS	Eos%	11.5	\pm 9.0	6.6	\pm 6.6	4.7	\pm 4.7
BP 200n	Bas%	1.4	\pm 5.0	1.4	\pm 5.0	1.4	\pm 5.0
BP 260n	RBC $\times 10^{12}/L$	2.31	\pm 0.18	4.25	\pm 0.24	5.24	\pm 0.30
QC Mode	HGB g/dL	5.9	\pm 0.4	12.3	\pm 0.6	16.2	\pm 0.8
	HCT %	19.6	\pm 1.5	40.4	\pm 2.0	53.6	\pm 2.4
	MCV fL	85.0	\pm 5.0	95.1	\pm 5.0	102.3	\pm 5.0
	MCH pg	25.5	\pm 2.5	28.9	\pm 2.5	30.9	\pm 2.5
	MCHC g/dL	30.0	\pm 3.0	30.4	\pm 3.0	30.2	\pm 3.0
	RDW-CV %	17.9	\pm 3.0	15.1	\pm 3.0	14.4	\pm 3.0
	RDW-SD fL	52.9	\pm 6.0	50.1	\pm 6.0	51.2	\pm 8.0
	PLT $\times 10^9/L$	54	\pm 20	197	\pm 40	387	\pm 60
	PLT-H $\times 10^9/L$	55	\pm 20	199	\pm 40	392	\pm 60
	MPV fL	12.0	\pm 3.0	10.8	\pm 3.0	11.2	\pm 3.0
	PDW	15.9	\pm 3.0	16.8	\pm 3.0	16.7	\pm 3.0
	PCT %	0.065	\pm 0.050	0.213	\pm 0.100	0.433	\pm 0.200
	P-LCR %	42.4	\pm 10.0	32.4	\pm 10.0	36.6	\pm 10.0
	P-LCC $\times 10^9/L$	23	\pm 8	64	\pm 24	142	\pm 50
	IMG# $\times 10^9/L$	0.04	\pm 0.34	0.10	\pm 0.39	0.25	\pm 2.09
	IMG%	1.2	\pm 10.0	1.3	\pm 10.0	1.2	\pm 10.0
	NRBC# $\times 10^9/L$	0.081	\pm 0.081	0.150	\pm 0.150	0.488	\pm 0.488
	NRBC% /100WBC	2.22	\pm 2.22	1.96	\pm 1.96	2.31	\pm 2.31
	IPF %	2.4	\pm 5.0	2.5	\pm 5.0	2.6	\pm 5.0
	ESR mm/h	38.03	\pm 10.00	15.51	\pm 4.00	/	\pm /

The ESR is not applicable for High level, and only ESR is involved in BP 200n and BP 260n.

Before using, refer to the instruction sheet for mixing directions.



Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen 518057, P.R.China

Tel: +86 755 26582479 26582888

Fax: +86 755 26582934 26582500



Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726



P/N: 046-003618-00 (13.0)