

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;

10.BS-480: BS-480, BS-490;

2.BS-240: BS-230, BS-240;

11.BS-600: BS-600, BS-620;

3.BS-240E: BS240E, BS240Pro;

12.BS-600M: BS-600M, BS-620M;

4.BS-300: BS-300, BS-320;

5.BS-330E:BS-330E(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-")13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800,
BS-1800plus;6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-
330E(V35.00)(Serial Number starts with "W8-" and
software version starts with "35.00"), BS-
350E(V35.00) (Serial Number starts with"W9-" and
software version starts with "35.00");

14.BS-1000M: BS-1000M, BS-1100M;

7.BS-380: BS-380, BS-390;

15.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-400: BS-400, BS-420;

16.BS-2800M:BS-2600M, BS-2800M;

9.BS-430: BS-430, BS-450,BS-460;

17.For applicable models of the analyte, please refer to the
reagent parameter sheet and instruction.

LOT : 043425002

: 2026-09-05

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzye	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

HS-CRP

English	High Sensitivity C-reaction Protein
Русский	C-реактивный белок
Português	Proteína C Reativa de Alta Sensibilidade
Español	proteína C reactiva de alta sensibilidad
Italiano	proteina C reattiva ad alta sensibilità
Türkçe	Yüksek Duyarlıklılı C Reaksiyon Proteindir

Abbreviated name		HS-CRP		Calibration Rule		spline			
Model		BS-200E ¹		BS-240 ²		BS-240E ³		BS-300 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.85	0.3	4.85	0.3	7.23	0.5	4.85	0.3

HS-CRP Calibrator



b		18.8	1.3	18.8	1.3	18.6	1.3	18.8	1.3
c	mg/L	38.7	2.6	38.7	2.6	38.0	2.5	38.7	2.6
d		154	10	154	10	156	10	154	10
e		322	21	322	21	318	21	322	21
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a		46.2	3	46.2	3	68.8	5	46.2	3
b	nmol/L	179	13	179	13	177	12	179	13
c		368	24	368	24	362	24	368	24
d		1466	98	1466	98	1485	99	1466	98
e		3065	205	3065	205	3027	202	3065	205
Model		BS-330E ⁵		BS-360E ⁶		BS-380 ⁷		BS-400 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.85	0.3	4.85	0.3	4.80	0.3	4.80	0.3
b	mg/L	18.8	1.3	18.8	1.3	18.8	1.3	18.8	1.3
c		38.7	2.6	38.7	2.6	39.1	2.6	39.1	2.6
d		154	10	154	10	155	10	155	10
e		322	21	322	21	322	21	322	21
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a		46.2	3	46.2	3	45.7	3	45.7	3
b	nmol/L	179	13	179	13	179	13	179	13
c		368	24	368	24	372	25	372	25
d		1466	98	1466	98	1476	99	1476	99
e		3065	205	3065	205	3065	205	3065	205
Model		BS-430 ⁹		BS-480 ¹⁰		BS-600 ¹¹		BS-600M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.85	0.3	4.80	0.3	4.80	0.3	4.80	0.3
b	mg/L	18.8	1.3	18.8	1.3	18.8	1.3	18.8	1.3
c		38.7	2.6	39.1	2.6	39.1	2.6	39.1	2.6
d		154	10	155	10	155	10	155	10
e		322	21	322	21	322	21	322	21
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a		46.2	3	45.7	3	45.7	3	45.7	3
b	nmol/L	179	13	179	13	179	13	179	13
c		368	24	372	25	372	25	372	25
d		1466	98	1476	99	1476	99	1476	99
e		3065	205	3065	205	3065	205	3065	205
Model		BS-800 ¹³		BS-1000M ¹⁴		BS-2000 ¹⁵		BS-2800M ¹⁶	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.85	0.3	4.80	0.3	4.95	0.3	4.80	0.3
b	mg/L	18.8	1.3	18.8	1.3	19.1	1.3	18.8	1.3
c		38.7	2.6	39.1	2.6	39.4	2.6	39.1	2.6
d		154	10	155	10	157	11	155	10
e		322	21	322	21	321	21	322	21
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/

HS-CRP Calibrator



a		46.2	3	45.7	3	47.1	3	45.7	3
b		179	13	179	12	182	12	179	12
c	nmol/L	368	24	372	24	375	25	372	24
d		1466	98	1476	99	1495	100	1476	99
e		3065	205	3065	204	3056	204	3065	204

mindray

HS-CRP Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 043425002

 2026-09-05

