

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



We have updated the control value for new version MYO reagent. Please select the corresponding value and update.

The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-240: BS-230, BS-240;

8.BS-600: BS-600, BS-620;

2.BS-240E: BS240E, BS240Pro;

9.BS-600M: BS-600M, BS-620M;

3.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)(Serial Number starts with "W9-" and software version starts with "35.00");

10.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

4.BS-380: BS-380, BS-390;

11.BS-1000M: BS-1000M, BS-1100M;

5.BS-400: BS-400, BS-420;

12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

6.BS-430: BS-430, BS-450,BS-460;

13.BS-2800M:BS-2600M, BS-2800M;

7.BS-480: BS-480, BS-490;

14.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

LOT : 045824001



: 2024-12-18

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kisaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

	MYO
English	Myoglobin
Русский	Миоглобин
Português	Mioglobina
Español	Mioglobina
Italiano	Mioglobina
Türkçe	Myoglobin

★ The following reference values are applicable to:Reagent lot No.: 045823006 or Previous lot ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a	ng/mL	/	/	/	/	/	/	/	/
b		/	/	/	/	/	/	/	/

c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a	ng/mL	/	/	/	/	/	/	/	/
b		/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
Model		BS-600M ⁹		BS-800 ¹⁰		BS-1000M ¹¹		BS-2000 ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	0.0	/	/	/
a	ng/mL	/	/	/	/	92.0	1.3	/	/
b		/	/	/	/	183	4	/	/
c		/	/	/	/	373	8	/	/
d		/	/	/	/	718	12	/	/
Model		BS-2800M ¹³							
Level	Unit	Calibrator Value	Expanded Uncertainty						
0.9%NaCl		/	/						
a	ng/mL	/	/						
b		/	/						
c		/	/						
d		/	/						

★ The following reference values are applicable to:Reagent lot No.: 045823007 or subsequent lot ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a	ng/mL	/	/	/	/	/	/	/	/
b		/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a	ng/mL	/	/	/	/	/	/	/	/
b		/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
Model		BS-600M ⁹		BS-800 ¹⁰		BS-1000M ¹¹		BS-2000 ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	0	/	/	/
a	ng/mL	/	/	/	/	104	1.5	/	/
b		/	/	/	/	197	4	/	/

c	/	/	/	/	410	9	/	/
d	/	/	/	/	808	14	/	/
Model		BS-2800M ¹³						
Level	Unit	Calibrator Value	Expanded Uncertainty					
0.9%NaCl		/	/					
a	ng/mL	/	/					
b		/	/					
c		/	/					
d		/	/					

mindray

MYO Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 045824001

 2024-12-18

