

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450, BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-

6.BS-300: BS-300, BS-320;

1800plus;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"),

16.BS-1000M: BS-1000M, BS-1100M;

BS-350E(Serial Number starts with "XS-")

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

18.BS-2800M:BS-2600M, BS-2800M.

330E(V35.00)(Serial Number starts with "W8-" and

software version starts with "35.00"), BS-350E(V35.00)

(Serial Number starts with "W9-" and software version

starts with "35.00");

19.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

9.BS-380: BS-380, BS-390;

20.For applicable models of the analyte, please refer to the reagent

10.BS-400: BS-400, BS-420;

parameter sheet and instruction.



: 143225006



: 2026-04-01

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

TBA	
English	Total Bile Acids
Русский	желчные кислоты
Português	Ácido Biliar Total
Español	ácidos biliares totales
Italiano	acidi biliari totali
Türkçe	Total Safra Asitleridir

Abbreviated name	Unit	Calibration Rule			Two-point linear		
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
TBA	μ mol/L	BS-120 ¹	33.4	0.9	BS-400 ¹⁰	34.1	1.0
		BS-200 ²	32.7	0.9	BS-430 ¹¹	33.6	0.9
		BS-200E ³	35.2	1.0	BS-480 ¹²	33.8	0.9
		BS-240 ⁴	33.4	0.9	BS-600 ¹³	33.7	0.9
		BS-240E ⁵	34.1	1.0	BS-600M ¹⁴	34.3	1.0
		BS-300 ⁶	34.1	1.0	BS-800 ¹⁵	34.1	1.0
		BS-330E ⁷	35.2	1.0	BS-1000M ¹⁶	34.8	1.0
		BS-360E ⁸	34.1	1.0	BS-2000 ¹⁷	33.9	1.0
		BS-380 ⁹	33.9	0.9	BS-2800M ¹⁸	34.8	1.0

mindray

TBA Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 143225006

 **2026-04-01**

