

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450, BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-")

16.BS-1000M: BS-1000M, BS-1100M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

9.BS-380: BS-380, BS-390;

18.BS-2800M:BS-2600M, BS-2800M.

10.BS-400: BS-400, BS-420;

19.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;
20.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

: 145323007



: 2024-12-14

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

Fe	
English	Iron
Русский	Железо
Português	Ferro
Español	Hierro
Italiano	Ferro
Türkçe	Demir

Abbreviated name	Unit	Calibration Rule			Two-point linear		
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
Fe	μmol/L	BS-120 ¹	/	/	BS-400 ¹⁰	/	/
		BS-200 ²	/	/	BS-430 ¹¹	/	/
		BS-200E ³	/	/	BS-480 ¹²	/	/
		BS-240 ⁴	/	/	BS-600 ¹³	/	/
		BS-240E ⁵	/	/	BS-600M ¹⁴	/	/
		BS-300 ⁶	/	/	BS-800 ¹⁵	/	/
		BS-330E ⁷	/	/	BS-1000M ¹⁶	43.7	0.5
		BS-360E ⁸	/	/	BS-2000 ¹⁷	/	/
		BS-380 ⁹	/	/	BS-2800M ¹⁸	/	/
	mg/L	BS-120 ¹	/	/	BS-400 ¹⁰	/	/
		BS-200 ²	/	/	BS-430 ¹¹	/	/
		BS-200E ³	/	/	BS-480 ¹²	/	/
		BS-240 ⁴	/	/	BS-600 ¹³	/	/
		BS-240E ⁵	/	/	BS-600M ¹⁴	/	/
		BS-300 ⁶	/	/	BS-800 ¹⁵	/	/
		BS-330E ⁷	/	/	BS-1000M ¹⁶	2.44	0.027
		BS-360E ⁸	/	/	BS-2000 ¹⁷	/	/
		BS-380 ⁹	/	/	BS-2800M ¹⁸	/	/

mindray

Fe Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 145323007

 2024-12-14

