

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

3.BS-200E: BS-200E, BS-220E,;

4.BS-240: BS-230, BS-240;

5.BS-240E: BS240E, BS240Pro;

6.BS-300: BS-300, BS-320;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-")8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-
330E(V35.00)(Serial Number starts with "W8-" and
software version starts with "35.00"), BS-350E(V35.00)
(Serial Number starts with "W9-" and software version
starts with "35.00");

9.BS-380: BS-380, BS-390;

10.BS-400: BS-400, BS-420;

11.BS-430: BS-430, BS-450, BS-460;

12.BS-480: BS-480, BS-490;

13.BS-600: BS-600, BS-620;

14.BS-600M: BS-600M, BS-620M;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-
1800plus;

16.BS-1000M: BS-1000M, BS-1100M;

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

18.BS-2800M:BS-2600M, BS-2800M.

19.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

20.For applicable models of the analyte, please refer to the reagent
parameter sheet and instruction.

: 145423010



: 2025-03-02

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

FUN	
English	Fructosamine
Русский	Фруктозамин
Português	Frutosamina
Español	Fructosamina
Italiano	Fruttosamina
Türkçe	Fruktozamin

Abbreviated name	Unit	Calibration Rule			Two-point linear	
		Model	Calibrator Value	Expanded Uncertainty	Model	Expanded Uncertainty
FUN	μmol/L	BS-120 ¹	/	/	BS-400 ¹⁰	/
		BS-200 ²	/	/	BS-430 ¹¹	/
		BS-200E ³	/	/	BS-480 ¹²	/
		BS-240 ⁴	/	/	BS-600 ¹³	/
		BS-240E ⁵	/	/	BS-600M ¹⁴	/
		BS-300 ⁶	/	/	BS-800 ¹⁵	/
		BS-330E ⁷	/	/	BS-1000M ¹⁶	452 7
		BS-360E ⁸	/	/	BS-2000 ¹⁷	/
		BS-380 ⁹	/	/	BS-2800M ¹⁸	/

mindray

FUN Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 145423010

 **2025-03-02**

