

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;

9.BS-600: BS-600, BS-620;

2.BS-240: BS-230, BS-240;

10.BS-600M: BS-600M, BS-620M;

3.BS-240E: BS240E, BS240Pro;

4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)(Serial Number starts with "W9-" and software version starts with "35.00");

11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12.BS-1000M: BS-1000M, BS-1100M;

5.BS-380: BS-380, BS-390;

13.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

6.BS-400: BS-400, BS-420;

14.BS-2800M:BS-2600M, BS-2800M;

7.BS-430: BS-430, BS-450,BS-460;

15.For applicable models of the analyte, please refer to the

8.BS-480: BS-480, BS-490;

reagent parameter sheet and instruction.

LOT : 147223013

: 2025-05-28

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kisaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

	CysC
English	Cystatin C
Русский	цистатин C
Português	Cistatina C
Español	cistatina C
Italiano	cistatina C
Türkçe	Cystatin C

Abbreviated name		CysC II		Calibration Rule		spline			
Model		BS-200E ¹		BS-240 ²		BS-240E ³		BS-360E ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a	mg/L	/	/	/	/	/	/	/	/
b		/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/

d		/	/	/	/	/	/	/	/
e		/	/	/	/	/	/	/	/
Model		BS-380 ⁵		BS-400 ⁶		BS-430 ⁷		BS-480 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a	mg/L	/	/	/	/	/	/	/	/
b		/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
e		/	/	/	/	/	/	/	/
Model		BS-600 ⁹		BS-600M ¹⁰		BS-800 ¹¹		BS-1000M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	0.000	/
a	mg/L	/	/	/	/	/	/	0.590	0.029
b		/	/	/	/	/	/	1.15	0.029
c		/	/	/	/	/	/	2.21	0.05
d		/	/	/	/	/	/	4.48	0.17
e		/	/	/	/	/	/	8.90	0.20
Model		BS-2000 ¹³		BS-2800M ¹⁴					
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty				
0.9%NaCl		/	/	/	/				
a	mg/L	/	/	/	/				
b		/	/	/	/				
c		/	/	/	/				
d		/	/	/	/				
e		/	/	/	/				

mindray

Cystatin C Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 147223013

 **2025-05-28**

