

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
Português : Planilha de valores do Sistema de Medição BS da Mindray
Español : Hoja de valores del sistema de medición Mindray BS
Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.			Русский: Данные совпадают во всех группах.		
Português : A dados de cada grupo é a mesma.			Español : la datos de cada grupo es la misma.		
Italiano : la dati di ogni gruppo è la stessa.			Türkçe : her grubun veri aynıdır.		
1.BS-200: BS-200, BS-220, BS-330, BS-350;			11.BS-480: BS-480, BS-490;		
2.BS-200E: BS-200E, BS-220E,;			12.BS-600: BS-600, BS-620;		
3.BS-240: BS-230, BS-240;			13.BS-600M: BS-600M, BS-620M;		
4.BS-240E: BS240E, BS240Pro;			14.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;		
5.BS-300: BS-300, BS-320;			15.BS-1000M: BS-1000M, BS-1100M;		
6.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")			16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;		
7.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");			17.BS-2800M:BS-2600M, BS-2800M;		
8.BS-380: BS-380, BS-390;			18.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.		
9.BS-400: BS-400, BS-420;			19.IFCC: HbA1c(mmol/mol)= HbA1c(μmol/L) / Hb(μmol/L)× 1000		
10.BS-430: BS-430, BS-450,BS-460;			NGSP/DCCT: HbA1c(%)=91.5×HbA1c(μmol/L) / Hb(μmol/L)+ 2.15		

LOT: 052525006



: 2027-01-15

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)
HbA1c/Hb					
English	Hemoglobin A1c/ Hemoglobin				
Русский	гемоглобин A1c / гемоглобин				
Português	Hemoglobina A1c /Hemoglobina				
Español	hemoglobina A1c /hemoglobina				
Italiano	Emoglobina A1c/Emoglobina				
Türkçe	Hemoglobin A1c / Hemoglobin				

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Assay Value	Range(Assay Value±3SD)	1 SD
Hb	μmol/L	BS-200 ¹	124	112 — 136	4	BS-430 ¹⁰	121	109 — 133	4
		BS-200E ²	124	112 — 136	4	BS-480 ¹¹	124	112 — 136	4
		BS-240 ³	121	109 — 133	4	BS-600 ¹²	124	112 — 136	4
		BS-240E ⁴	121	109 — 133	4	BS-600M ¹³	124	112 — 136	4
		BS-300 ⁵	121	109 — 133	4	BS-800 ¹⁴	121	109 — 133	4
		BS-330E ⁶	121	109 — 133	4	BS-1000M ¹⁵	124	112 — 136	4
		BS-360E ⁷	121	109 — 133	4	BS-2000 ¹⁶	124	112 — 136	4
		BS-380 ⁸	124	112 — 136	4	BS-2800M ¹⁷	124	112 — 136	4
		BS-400 ⁹	121	109 — 133	4				
		BS-200 ¹	11.4	10.2 — 12.6	0.4	BS-430 ¹⁰	11.1	9.9 — 12.3	0.4
		BS-200E ²	11.4	10.2 — 12.6	0.4	BS-480 ¹¹	11.4	10.2 — 12.6	0.4
		BS-240 ³	11.1	9.9 — 12.3	0.4	BS-600 ¹²	11.4	10.2 — 12.6	0.4

Abbreviated name	Unit	Model	Assay Value	Range(Assay Value±3SD)			1 SD	Model	Assay Value	Range(Assay Value±3SD)			1 SD
HbA1c	μmol/L	BS-240E ⁴	11.1	9.9	—	12.3	0.4	BS-600M ¹³	11.4	10.2	—	12.6	0.4
		BS-300 ⁵	11.1	9.9	—	12.3	0.4	BS-800 ¹⁴	11.1	9.9	—	12.3	0.4
		BS-330E ⁶	11.1	9.9	—	12.3	0.4	BS-1000M ¹⁵	11.3	10.1	—	12.5	0.4
		BS-360E ⁷	11.1	9.9	—	12.3	0.4	BS-2000 ¹⁶	11.3	10.1	—	12.5	0.4
		BS-380 ⁸	11.4	10.2	—	12.6	0.4	BS-2800M ¹⁷	11.4	10.2	—	12.6	0.4
		BS-400 ⁹	11.1	9.9	—	12.3	0.4						
HbA1c(mmol/mol) ¹⁹	mmol/mol	BS-200 ¹	91.9	82.9	—	100.9	3.0	BS-430 ¹⁰	91.7	82.7	—	100.7	3.0
		BS-200E ²	91.9	82.9	—	100.9	3.0	BS-480 ¹¹	91.9	82.9	—	100.9	3.0
		BS-240 ³	91.7	82.7	—	100.7	3.0	BS-600 ¹²	91.9	82.9	—	100.9	3.0
		BS-240E ⁴	91.7	82.7	—	100.7	3.0	BS-600M ¹³	91.9	82.9	—	100.9	3.0
		BS-300 ⁵	91.7	82.7	—	100.7	3.0	BS-800 ¹⁴	91.8	82.8	—	100.8	3.0
		BS-330E ⁶	91.7	82.7	—	100.7	3.0	BS-1000M ¹⁵	91.2	82.2	—	100.2	3.0
		BS-360E ⁷	91.7	82.7	—	100.7	3.0	BS-2000 ¹⁶	91.3	82.3	—	100.3	3.0
		BS-380 ⁸	91.9	82.9	—	100.9	3.0	BS-2800M ¹⁷	92.4	83.4	—	101.4	3.0
		BS-400 ⁹	91.7	82.7	—	100.7	3.0						
HbA1c(%) ¹ ₉	%	BS-200 ¹	10.6	9.7	—	11.5	0.3	BS-430 ¹⁰	10.5	9.6	—	11.4	0.3
		BS-200E ²	10.6	9.7	—	11.5	0.3	BS-480 ¹¹	10.6	9.7	—	11.5	0.3
		BS-240 ³	10.5	9.6	—	11.4	0.3	BS-600 ¹²	10.6	9.7	—	11.5	0.3
		BS-240E ⁴	10.5	9.6	—	11.4	0.3	BS-600M ¹³	10.6	9.7	—	11.5	0.3
		BS-300 ⁵	10.5	9.6	—	11.4	0.3	BS-800 ¹⁴	10.6	9.7	—	11.5	0.3
		BS-330E ⁶	10.5	9.6	—	11.4	0.3	BS-1000M ¹⁵	10.5	9.6	—	11.4	0.3
		BS-360E ⁷	10.5	9.6	—	11.4	0.3	BS-2000 ¹⁶	10.5	9.6	—	11.4	0.3
		BS-380 ⁸	10.6	9.7	—	11.5	0.3	BS-2800M ¹⁷	10.6	9.7	—	11.5	0.3
		BS-400 ⁹	10.5	9.6	—	11.4	0.3						

mindray

HbA1c Control P

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 052525006

2027-01-15

